

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN053019\
 Data File : BN005954.D
 Acq On : 30 May 2019 12:07
 Operator : JU/SJ
 Sample : SST00.188
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST00.188

Manual Integrations
 APPROVED

mohammad
 5/31/2019 8:11:44 AM

Quant Time: May 30 14:07:42 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN053019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 30 13:56:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.74	152	8135	0.40	ng/ul	0.00
2) Naphthalene-d8	10.52	136	29487	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.39	164	15458	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.14	188	36545	0.40	ng/ul	0.00
16) Chrysene-d12	21.35	240	24250	0.40	ng/ul	0.00
20) Perylene-d12	23.63	264	21934	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.12	152	4451	0.10	ng/ul	0.00
14) Fluoranthene-d10	19.18	212	9487	0.08	ng/ul	0.00
Target Compounds				Ovalue		
3) Naphthalene	10.57	128	8125	0.104	ng/ul	96
5) 2-Methylnaphthalene	12.19	142	5548	0.104	ng/ul	99
7) Acenaphthylene	14.10	152	8512	0.109	ng/ul	98
8) Acenaphthene	14.45	153	5938	0.108	ng/ul	100
9) Fluorene	15.44	166	6942	0.108	ng/ul	97
12) Phenanthrene	17.18	178	11799	0.118	ng/ul	98
13) Anthracene	17.27	178	10893	0.110	ng/ul	98
15) Fluoranthene	19.21	202	12786	0.093	ng/ul	94
17) Pyrene	19.57	202	13279	0.148	ng/ul	99
18) Benzo(a)anthracene	21.33	228	10814	0.129	ng/ul	93
19) Chrysene	21.39	228	10295m	0.100	ng/ul	
21) Benzo(b)fluoranthene	22.94	252	9299m	0.141	ng/ul	
22) Benzo(k)fluoranthene	22.99	252	8758	0.106	ng/ul#	62
23) Benzo(a)pyrene	23.53	252	8584	0.116	ng/ul#	55
24) Indeno(1,2,3-cd)pyrene	25.92	276	9514	0.115	ng/ul#	98
25) Dibenzo(a,h)anthracene	25.94	278	7642	0.112	ng/ul#	67
26) Benzo(a,h,i)perylene	26.63	276	7959	0.113	ng/ul#	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN053019\
Data File : BN005954.D
Acq On : 30 May 2019 12:07
Operator : JU/SJ
Sample : SST0.188
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.188

Manual Integrations
APPROVED

mohammad
5/31/2019 8:11:44 AM

Quant Time: May 30 14:07:42 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN053019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 30 13:56:23 2019
Response via : Initial Calibration

