


Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN060220\  
 Data File : BN010740.D  
 Acq On : 02 Jun 2020 11:24  
 Operator : CG/JU  
 Sample : DFTPP62  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 DFTPP62

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\SOM-EPA-SIM-BN060220.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Tue Jun 02 16:19:11 2020



AutoFind: Scans 2486, 2487, 2488; Background Corrected with Scan 2479

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 21.5      | 87613   | PASS             |
| 68          | 69           | 0.00         | 2            | 2.0       | 3114    | PASS             |
| 69          | 198          | 0.00         | 100          | 37.4      | 152357  | PASS             |
| 70          | 69           | 0.00         | 2            | 0.6       | 904     | PASS             |
| 127         | 198          | 10           | 80           | 45.6      | 185682  | PASS             |
| 197         | 198          | 0.00         | 2            | 0.3       | 1369    | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 407296  | PASS             |
| 199         | 198          | 5            | 9            | 6.7       | 27169   | PASS             |
| 275         | 198          | 10           | 60           | 30.2      | 122944  | PASS             |
| 365         | 198          | 1            | 100          | 4.5       | 18435   | PASS             |
| 441         | 198          | 0.01         | 100          | 13.5      | 55064   | PASS             |
| 442         | 198          | 50           | 100          | 87.3      | 355520  | PASS             |
| 443         | 442          | 15           | 24           | 17.3      | 61480   | PASS             |

Scan 2479 (17.369 min): BN010740.D  
DFTPP62

| m/z    | Abundance |
|--------|-----------|
| 39.80  | 605.0     |
| 44.00  | 2073.0    |
| 72.90  | 937.0     |
| 207.00 | 3166.0    |
| 207.90 | 481.0     |
| 281.00 | 1102.0    |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP62

| m/z    | Abundance |
|--------|-----------|
| 37.10  | 760.0     |
| 38.00  | 2784.0    |
| 39.00  | 10366.0   |
| 39.90  | 1110.0    |
| 41.00  | 454.0     |
| 44.00  | 2459.0    |
| 45.10  | 312.0     |
| 49.10  | 1241.0    |
| 50.00  | 31736.0   |
| 51.00  | 120416.0  |
| 52.00  | 5413.0    |
| 52.90  | 352.0     |
| 55.00  | 777.0     |
| 55.90  | 3574.0    |
| 57.00  | 9161.0    |
| 58.00  | 633.0     |
| 60.90  | 2631.0    |
| 62.00  | 3163.0    |
| 63.10  | 6490.0    |
| 64.00  | 1257.0    |
| 65.00  | 2346.0    |
| 65.90  | 486.0     |
| 67.00  | 435.0     |
| 68.10  | 3725.0    |
| 68.90  | 201024.0  |
| 69.90  | 1311.0    |
| 73.00  | 2764.0    |
| 74.00  | 25352.0   |
| 75.00  | 34408.0   |
| 76.00  | 12897.0   |
| 77.00  | 237312.0  |
| 78.00  | 17112.0   |
| 79.00  | 16888.0   |
| 79.90  | 11862.0   |
| 81.00  | 17544.0   |
| 82.00  | 5668.0    |
| 83.00  | 3289.0    |
| 84.10  | 626.0     |
| 84.90  | 2358.0    |
| 86.00  | 5429.0    |
| 87.00  | 2099.0    |
| 88.10  | 881.0     |
| 88.90  | 657.0     |
| 91.00  | 3723.0    |
| 92.00  | 5237.0    |
| 93.00  | 31136.0   |
| 94.00  | 1921.0    |
| 95.10  | 500.0     |
| 96.10  | 1532.0    |
| 96.80  | 588.0     |
| 98.00  | 23296.0   |
| 98.90  | 20912.0   |
| 100.00 | 1661.0    |
| 101.00 | 8366.0    |
| 102.00 | 613.0     |
| 103.00 | 3584.0    |
| 104.00 | 5664.0    |
| 105.00 | 6887.0    |
| 106.00 | 1751.0    |
| 107.00 | 74352.0   |
| 108.00 | 11236.0   |
| 110.00 | 128224.0  |
| 111.00 | 20480.0   |
| 111.90 | 2658.0    |
| 113.10 | 734.0     |
| 114.80 | 393.0     |
| 116.00 | 4394.0    |
| 117.00 | 74136.0   |
| 118.00 | 5127.0    |
| 119.00 | 648.0     |
| 120.00 | 858.0     |
| 121.10 | 392.0     |
| 122.00 | 4354.0    |
| 123.00 | 8311.0    |
| 123.90 | 3719.0    |
| 124.90 | 3440.0    |
| 127.00 | 231680.0  |
| 128.00 | 19328.0   |
| 129.00 | 98872.0   |
| 130.00 | 7291.0    |
| 131.00 | 1670.0    |
| 132.00 | 1398.0    |
| 133.00 | 776.0     |
| 134.00 | 2424.0    |
| 134.90 | 8102.0    |
| 136.00 | 3147.0    |
| 136.90 | 5021.0    |
| 137.90 | 973.0     |
| 140.00 | 1445.0    |
| 141.00 | 13296.0   |
| 142.00 | 4110.0    |
| 143.00 | 2975.0    |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP62

|        |          |
|--------|----------|
| 144.10 | 846.0    |
| 145.10 | 441.0    |
| 146.00 | 3414.0   |
| 147.00 | 7752.0   |
| 148.00 | 19176.0  |
| 149.00 | 3677.0   |
| 149.90 | 1072.0   |
| 151.00 | 1763.0   |
| 151.80 | 1202.0   |
| 153.00 | 4341.0   |
| 154.00 | 3669.0   |
| 155.00 | 8373.0   |
| 156.00 | 10566.0  |
| 157.00 | 1913.0   |
| 157.90 | 2679.0   |
| 159.00 | 2297.0   |
| 160.00 | 4049.0   |
| 161.00 | 5707.0   |
| 162.00 | 1780.0   |
| 162.80 | 382.0    |
| 163.80 | 744.0    |
| 164.90 | 5434.0   |
| 166.00 | 3326.0   |
| 167.00 | 33392.0  |
| 168.00 | 19576.0  |
| 169.00 | 3082.0   |
| 170.00 | 1293.0   |
| 171.00 | 1310.0   |
| 171.90 | 2723.0   |
| 173.00 | 3110.0   |
| 173.90 | 6239.0   |
| 175.00 | 9405.0   |
| 175.90 | 2734.0   |
| 177.00 | 5431.0   |
| 178.00 | 1971.0   |
| 178.90 | 20296.0  |
| 180.00 | 16374.0  |
| 181.00 | 6746.0   |
| 181.90 | 813.0    |
| 182.90 | 850.0    |
| 184.10 | 2029.0   |
| 185.00 | 10350.0  |
| 186.00 | 80384.0  |
| 187.00 | 23192.0  |
| 188.00 | 3000.0   |
| 188.90 | 5209.0   |
| 190.00 | 874.0    |
| 191.00 | 2377.0   |
| 191.90 | 6034.0   |
| 193.00 | 7014.0   |
| 194.10 | 1606.0   |
| 194.90 | 1040.0   |
| 196.00 | 14446.0  |
| 197.10 | 4108.0   |
| 197.90 | 465024.0 |
| 198.90 | 33880.0  |
| 200.00 | 3795.0   |
| 200.80 | 547.0    |
| 201.30 | 2251.0   |
| 202.30 | 548.0    |
| 203.00 | 4123.0   |
| 204.00 | 19896.0  |
| 205.00 | 33640.0  |
| 206.00 | 134528.0 |
| 207.00 | 19592.0  |
| 208.00 | 5791.0   |
| 208.90 | 2224.0   |
| 210.20 | 2506.0   |
| 211.00 | 5676.0   |
| 212.10 | 774.0    |
| 212.90 | 581.0    |
| 214.90 | 2274.0   |
| 216.00 | 3075.0   |
| 217.00 | 38896.0  |
| 218.00 | 4857.0   |
| 219.90 | 854.0    |
| 221.00 | 21432.0  |
| 222.00 | 4246.0   |
| 223.00 | 8909.0   |
| 224.00 | 80576.0  |
| 225.00 | 20640.0  |
| 226.00 | 2663.0   |
| 226.90 | 40904.0  |
| 227.90 | 4291.0   |
| 229.00 | 7319.0   |
| 229.90 | 1135.0   |
| 231.00 | 3321.0   |
| 232.00 | 672.0    |
| 232.90 | 845.0    |
| 233.90 | 2780.0   |
| 235.00 | 3558.0   |
| 235.90 | 1551.0   |
| 237.00 | 5365.0   |
| 237.90 | 568.0    |
| 239.00 | 1824.0   |
| 240.00 | 1131.0   |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP62

|        |          |
|--------|----------|
| 241.00 | 1924.0   |
| 242.00 | 5104.0   |
| 242.90 | 4582.0   |
| 244.00 | 61448.0  |
| 245.00 | 8735.0   |
| 246.00 | 13293.0  |
| 247.10 | 2512.0   |
| 247.90 | 659.0    |
| 248.90 | 1437.0   |
| 250.80 | 490.0    |
| 252.10 | 559.0    |
| 253.00 | 1945.0   |
| 255.00 | 327680.0 |
| 256.00 | 47224.0  |
| 257.00 | 4597.0   |
| 258.00 | 18216.0  |
| 259.00 | 3388.0   |
| 260.00 | 542.0    |
| 261.00 | 851.0    |
| 262.80 | 480.0    |
| 263.90 | 522.0    |
| 265.00 | 8285.0   |
| 266.00 | 1418.0   |
| 266.90 | 406.0    |
| 269.90 | 390.0    |
| 271.10 | 926.0    |
| 272.20 | 1096.0   |
| 273.00 | 10186.0  |
| 274.00 | 24984.0  |
| 275.00 | 137856.0 |
| 276.00 | 19816.0  |
| 277.00 | 12940.0  |
| 278.00 | 2427.0   |
| 279.00 | 451.0    |
| 281.00 | 1128.0   |
| 281.90 | 466.0    |
| 283.00 | 1556.0   |
| 283.80 | 1001.0   |
| 285.10 | 1920.0   |
| 286.10 | 335.0    |
| 288.70 | 483.0    |
| 289.90 | 764.0    |
| 290.90 | 579.0    |
| 291.90 | 583.0    |
| 293.00 | 2963.0   |
| 294.00 | 1074.0   |
| 295.10 | 1663.0   |
| 296.00 | 39960.0  |
| 296.90 | 6207.0   |
| 297.90 | 560.0    |
| 300.90 | 724.0    |
| 302.00 | 696.0    |
| 303.00 | 5695.0   |
| 304.00 | 1495.0   |
| 305.10 | 667.0    |
| 308.00 | 424.0    |
| 309.90 | 648.0    |
| 313.00 | 558.0    |
| 314.00 | 2332.0   |
| 315.00 | 4684.0   |
| 316.00 | 3002.0   |
| 316.90 | 435.0    |
| 321.00 | 1359.0   |
| 321.90 | 596.0    |
| 322.20 | 624.0    |
| 323.00 | 10792.0  |
| 324.10 | 2350.0   |
| 326.90 | 2807.0   |
| 328.00 | 1320.0   |
| 332.10 | 886.0    |
| 333.00 | 1231.0   |
| 334.00 | 8794.0   |
| 335.00 | 2047.0   |
| 338.80 | 360.0    |
| 341.00 | 1707.0   |
| 341.90 | 344.0    |
| 346.00 | 2381.0   |
| 346.90 | 589.0    |
| 351.10 | 409.0    |
| 352.00 | 4523.0   |
| 353.10 | 2776.0   |
| 354.00 | 2210.0   |
| 355.00 | 756.0    |
| 364.90 | 18680.0  |
| 366.00 | 3355.0   |
| 369.90 | 452.0    |
| 370.90 | 700.0    |
| 372.00 | 6257.0   |
| 373.00 | 1774.0   |
| 383.00 | 1824.0   |
| 383.90 | 586.0    |
| 390.00 | 424.0    |
| 391.10 | 548.0    |
| 401.90 | 2750.0   |
| 403.00 | 3819.0   |
| 403.90 | 1556.0   |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP62

|        |          |
|--------|----------|
| 404.80 | 581.0    |
| 421.00 | 2147.0   |
| 422.00 | 2797.0   |
| 423.00 | 20000.0  |
| 424.00 | 7299.0   |
| 425.10 | 1306.0   |
| 441.00 | 51824.0  |
| 442.00 | 328192.0 |
| 443.00 | 62992.0  |
| 444.00 | 5794.0   |
| 444.90 | 575.0    |

**Instrument :**  
BNA\_N  
**ClientSampleId :**  
DFTPP62

| m/z    | Abundance |
|--------|-----------|
| 36.90  | 933.0     |
| 38.00  | 1837.0    |
| 39.10  | 8859.0    |
| 40.00  | 832.0     |
| 40.90  | 338.0     |
| 43.90  | 1700.0    |
| 49.00  | 1037.0    |
| 50.00  | 27560.0   |
| 51.00  | 100424.0  |
| 52.00  | 5667.0    |
| 52.90  | 418.0     |
| 54.90  | 950.0     |
| 56.00  | 3109.0    |
| 57.00  | 8303.0    |
| 61.00  | 2705.0    |
| 62.00  | 1947.0    |
| 63.00  | 6335.0    |
| 63.90  | 1254.0    |
| 65.00  | 2709.0    |
| 68.00  | 3953.0    |
| 69.00  | 177152.0  |
| 70.00  | 901.0     |
| 73.00  | 3028.0    |
| 74.00  | 21288.0   |
| 75.00  | 30696.0   |
| 76.00  | 10739.0   |
| 77.00  | 208896.0  |
| 78.00  | 14046.0   |
| 79.00  | 15496.0   |
| 80.00  | 11737.0   |
| 81.00  | 14890.0   |
| 81.90  | 2913.0    |
| 83.00  | 2961.0    |
| 85.00  | 2548.0    |
| 86.00  | 4404.0    |
| 86.90  | 2396.0    |
| 87.90  | 1025.0    |
| 91.00  | 3786.0    |
| 92.00  | 4901.0    |
| 93.00  | 24192.0   |
| 94.00  | 2152.0    |
| 94.90  | 806.0     |
| 95.90  | 897.0     |
| 96.10  | 895.0     |
| 97.10  | 694.0     |
| 98.00  | 21520.0   |
| 99.00  | 17160.0   |
| 99.90  | 1327.0    |
| 100.90 | 8169.0    |
| 101.80 | 338.0     |
| 102.20 | 375.0     |
| 102.90 | 3117.0    |
| 104.00 | 5578.0    |
| 105.00 | 4323.0    |
| 105.90 | 1803.0    |
| 107.00 | 64120.0   |
| 108.00 | 9295.0    |
| 108.90 | 1947.0    |
| 110.00 | 106184.0  |
| 111.00 | 18288.0   |
| 112.00 | 3235.0    |
| 112.90 | 902.0     |
| 114.90 | 438.0     |
| 116.00 | 4587.0    |
| 117.00 | 65208.0   |
| 118.00 | 4510.0    |
| 119.10 | 920.0     |
| 120.00 | 1085.0    |
| 120.90 | 458.0     |
| 122.00 | 4729.0    |
| 122.90 | 6544.0    |
| 123.90 | 3700.0    |
| 125.00 | 2483.0    |
| 127.00 | 223168.0  |
| 128.00 | 15213.0   |
| 129.00 | 87944.0   |
| 130.00 | 7842.0    |
| 130.90 | 1667.0    |
| 132.10 | 559.0     |
| 133.00 | 1545.0    |
| 134.00 | 2623.0    |
| 134.90 | 8364.0    |
| 136.00 | 3289.0    |
| 137.00 | 4600.0    |
| 138.00 | 690.0     |
| 138.90 | 664.0     |
| 140.00 | 1413.0    |
| 141.00 | 13313.0   |
| 142.10 | 3495.0    |
| 143.00 | 2776.0    |
| 144.00 | 853.0     |
| 144.80 | 382.0     |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP62

|        |          |
|--------|----------|
| 145.90 | 1951.0   |
| 146.90 | 5845.0   |
| 147.90 | 15664.0  |
| 149.00 | 4020.0   |
| 149.90 | 1221.0   |
| 151.00 | 1903.0   |
| 151.80 | 1404.0   |
| 153.00 | 3701.0   |
| 153.90 | 2639.0   |
| 155.00 | 6472.0   |
| 156.00 | 10100.0  |
| 157.00 | 2065.0   |
| 157.90 | 2688.0   |
| 159.00 | 1763.0   |
| 159.90 | 4387.0   |
| 160.90 | 6120.0   |
| 162.00 | 1559.0   |
| 162.90 | 529.0    |
| 164.10 | 740.0    |
| 165.00 | 4604.0   |
| 166.00 | 4064.0   |
| 167.00 | 28384.0  |
| 168.00 | 18744.0  |
| 169.00 | 1999.0   |
| 169.90 | 928.0    |
| 171.00 | 1461.0   |
| 171.90 | 2188.0   |
| 173.00 | 3453.0   |
| 174.00 | 5567.0   |
| 175.00 | 10681.0  |
| 175.90 | 2683.0   |
| 177.00 | 5461.0   |
| 178.00 | 1807.0   |
| 179.00 | 20944.0  |
| 180.00 | 14658.0  |
| 181.00 | 6709.0   |
| 182.00 | 835.0    |
| 182.70 | 378.0    |
| 183.90 | 1352.0   |
| 185.00 | 10092.0  |
| 186.00 | 76672.0  |
| 187.00 | 23184.0  |
| 188.00 | 2396.0   |
| 188.90 | 5589.0   |
| 190.20 | 847.0    |
| 190.90 | 3072.0   |
| 191.90 | 6495.0   |
| 193.00 | 8829.0   |
| 194.00 | 1263.0   |
| 195.10 | 1283.0   |
| 196.00 | 16230.0  |
| 198.00 | 497984.0 |
| 199.00 | 32144.0  |
| 200.00 | 2582.0   |
| 201.10 | 1267.0   |
| 201.80 | 1223.0   |
| 203.00 | 3982.0   |
| 204.00 | 19112.0  |
| 205.00 | 34704.0  |
| 206.10 | 148096.0 |
| 207.00 | 20840.0  |
| 208.00 | 5517.0   |
| 208.90 | 1866.0   |
| 210.10 | 2496.0   |
| 211.00 | 6587.0   |
| 211.60 | 2178.0   |
| 213.20 | 312.0    |
| 214.90 | 1690.0   |
| 216.00 | 2830.0   |
| 216.90 | 39024.0  |
| 218.00 | 4750.0   |
| 219.00 | 475.0    |
| 221.00 | 16045.0  |
| 222.00 | 5535.0   |
| 223.00 | 8111.0   |
| 224.00 | 82200.0  |
| 225.00 | 19192.0  |
| 225.90 | 2117.0   |
| 227.00 | 37272.0  |
| 228.00 | 6789.0   |
| 229.00 | 7202.0   |
| 229.90 | 1167.0   |
| 231.00 | 2682.0   |
| 232.00 | 312.0    |
| 233.10 | 595.0    |
| 234.10 | 2111.0   |
| 235.00 | 3459.0   |
| 235.90 | 1974.0   |
| 237.00 | 4785.0   |
| 238.10 | 701.0    |
| 238.90 | 1666.0   |
| 240.00 | 1196.0   |
| 240.90 | 1753.0   |
| 242.00 | 4082.0   |
| 243.10 | 4223.0   |
| 244.00 | 57944.0  |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP62

|        |          |
|--------|----------|
| 245.00 | 7493.0   |
| 246.00 | 13586.0  |
| 247.00 | 3030.0   |
| 247.90 | 962.0    |
| 249.00 | 2221.0   |
| 251.00 | 803.0    |
| 252.00 | 822.0    |
| 252.90 | 2217.0   |
| 255.00 | 334784.0 |
| 256.00 | 47840.0  |
| 257.00 | 4515.0   |
| 258.00 | 20240.0  |
| 259.00 | 3424.0   |
| 260.10 | 425.0    |
| 260.90 | 934.0    |
| 262.80 | 505.0    |
| 264.00 | 834.0    |
| 265.00 | 9639.0   |
| 265.90 | 1234.0   |
| 270.90 | 868.0    |
| 272.10 | 1845.0   |
| 272.90 | 9055.0   |
| 274.00 | 24632.0  |
| 275.00 | 148224.0 |
| 276.00 | 19320.0  |
| 277.00 | 12629.0  |
| 277.90 | 1875.0   |
| 279.10 | 792.0    |
| 280.90 | 954.0    |
| 281.90 | 686.0    |
| 282.90 | 1590.0   |
| 283.90 | 1276.0   |
| 285.00 | 1439.0   |
| 285.80 | 381.0    |
| 289.00 | 456.0    |
| 290.00 | 429.0    |
| 291.90 | 520.0    |
| 292.90 | 3391.0   |
| 294.00 | 1223.0   |
| 296.00 | 43304.0  |
| 296.90 | 6876.0   |
| 298.00 | 688.0    |
| 301.10 | 570.0    |
| 302.00 | 932.0    |
| 302.90 | 5267.0   |
| 304.10 | 1379.0   |
| 307.90 | 544.0    |
| 308.90 | 507.0    |
| 310.00 | 395.0    |
| 313.90 | 3193.0   |
| 315.00 | 5192.0   |
| 316.00 | 3012.0   |
| 316.90 | 562.0    |
| 321.10 | 1531.0   |
| 321.90 | 633.0    |
| 323.00 | 11679.0  |
| 324.00 | 2879.0   |
| 326.00 | 313.0    |
| 326.90 | 3016.0   |
| 328.00 | 2036.0   |
| 328.90 | 400.0    |
| 332.00 | 1359.0   |
| 332.80 | 983.0    |
| 334.00 | 10534.0  |
| 335.10 | 2334.0   |
| 336.00 | 481.0    |
| 340.90 | 1814.0   |
| 341.90 | 488.0    |
| 345.90 | 3126.0   |
| 347.00 | 730.0    |
| 351.90 | 3361.0   |
| 352.90 | 2648.0   |
| 354.00 | 5085.0   |
| 354.90 | 1345.0   |
| 365.00 | 23040.0  |
| 366.00 | 3006.0   |
| 369.80 | 382.0    |
| 371.00 | 1381.0   |
| 372.00 | 7583.0   |
| 372.90 | 1731.0   |
| 382.90 | 2379.0   |
| 383.90 | 1085.0   |
| 389.90 | 1000.0   |
| 390.80 | 401.0    |
| 391.00 | 346.0    |
| 392.10 | 803.0    |
| 401.00 | 481.0    |
| 402.00 | 3241.0   |
| 403.00 | 4620.0   |
| 404.00 | 2160.0   |
| 405.00 | 390.0    |
| 415.10 | 332.0    |
| 421.00 | 4374.0   |
| 422.00 | 4175.0   |
| 423.00 | 25680.0  |
| 424.00 | 7176.0   |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP62

|        |          |
|--------|----------|
| 425.00 | 1301.0   |
| 441.00 | 67928.0  |
| 442.00 | 448960.0 |
| 443.00 | 73312.0  |
| 444.10 | 7145.0   |
| 444.90 | 488.0    |

|                  |
|------------------|
| Instrument :     |
| BNA_N            |
| ClientSampleId : |
| DFTPP62          |

| m/z    | Abundance |
|--------|-----------|
| 37.00  | 384.0     |
| 38.00  | 671.0     |
| 39.00  | 4632.0    |
| 40.00  | 1073.0    |
| 44.00  | 2255.0    |
| 44.80  | 301.0     |
| 49.00  | 592.0     |
| 50.00  | 11970.0   |
| 51.00  | 42000.0   |
| 52.00  | 2097.0    |
| 54.90  | 381.0     |
| 56.00  | 1520.0    |
| 56.90  | 3600.0    |
| 60.70  | 587.0     |
| 61.10  | 594.0     |
| 61.90  | 1233.0    |
| 63.00  | 2661.0    |
| 63.90  | 417.0     |
| 65.00  | 1537.0    |
| 65.80  | 488.0     |
| 67.10  | 433.0     |
| 67.90  | 1666.0    |
| 69.00  | 78896.0   |
| 70.00  | 501.0     |
| 73.00  | 1787.0    |
| 74.00  | 8057.0    |
| 75.00  | 12976.0   |
| 76.00  | 5255.0    |
| 77.00  | 88160.0   |
| 78.00  | 6161.0    |
| 78.90  | 6128.0    |
| 79.90  | 4668.0    |
| 81.00  | 7175.0    |
| 82.00  | 1935.0    |
| 82.90  | 1183.0    |
| 84.00  | 552.0     |
| 85.00  | 1231.0    |
| 85.90  | 1632.0    |
| 87.00  | 1276.0    |
| 91.10  | 1589.0    |
| 91.90  | 1932.0    |
| 93.00  | 12131.0   |
| 94.00  | 1051.0    |
| 96.00  | 590.0     |
| 97.20  | 428.0     |
| 98.00  | 10095.0   |
| 99.00  | 8040.0    |
| 99.90  | 477.0     |
| 101.00 | 4065.0    |
| 102.90 | 1592.0    |
| 104.00 | 2301.0    |
| 105.00 | 2258.0    |
| 105.90 | 1324.0    |
| 107.00 | 29536.0   |
| 108.00 | 4804.0    |
| 109.10 | 1088.0    |
| 110.00 | 52544.0   |
| 111.00 | 8283.0    |
| 112.00 | 976.0     |
| 113.10 | 372.0     |
| 115.00 | 322.0     |
| 116.00 | 1579.0    |
| 117.00 | 29864.0   |
| 118.00 | 2683.0    |
| 118.90 | 341.0     |
| 119.90 | 450.0     |
| 122.00 | 2621.0    |
| 122.90 | 3446.0    |
| 123.90 | 1964.0    |
| 125.00 | 1040.0    |
| 126.00 | 684.0     |
| 127.00 | 102200.0  |
| 127.90 | 8500.0    |
| 129.00 | 41648.0   |
| 130.00 | 3595.0    |
| 131.10 | 1029.0    |
| 131.90 | 326.0     |
| 132.70 | 512.0     |
| 134.00 | 1269.0    |
| 135.00 | 3729.0    |
| 136.00 | 1393.0    |
| 136.90 | 2149.0    |
| 137.90 | 442.0     |
| 139.80 | 652.0     |
| 140.90 | 5090.0    |
| 141.90 | 1668.0    |
| 142.90 | 1594.0    |
| 144.00 | 390.0     |
| 144.80 | 552.0     |
| 145.90 | 1336.0    |
| 147.00 | 3639.0    |
| 147.90 | 8015.0    |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP62

|        |          |
|--------|----------|
| 149.00 | 1775.0   |
| 149.80 | 321.0    |
| 151.10 | 776.0    |
| 152.00 | 352.0    |
| 153.00 | 1871.0   |
| 153.90 | 1473.0   |
| 155.00 | 3604.0   |
| 156.00 | 5228.0   |
| 157.00 | 1184.0   |
| 157.80 | 1295.0   |
| 158.90 | 873.0    |
| 160.00 | 1830.0   |
| 160.90 | 3437.0   |
| 161.80 | 745.0    |
| 163.00 | 312.0    |
| 164.90 | 2460.0   |
| 165.90 | 1799.0   |
| 167.00 | 14423.0  |
| 168.00 | 10156.0  |
| 169.00 | 1278.0   |
| 170.00 | 637.0    |
| 171.10 | 904.0    |
| 171.90 | 1235.0   |
| 173.00 | 1425.0   |
| 173.90 | 2758.0   |
| 175.10 | 5140.0   |
| 176.10 | 1230.0   |
| 177.10 | 2733.0   |
| 177.90 | 1012.0   |
| 178.90 | 9927.0   |
| 180.00 | 7283.0   |
| 180.90 | 2819.0   |
| 182.00 | 452.0    |
| 183.10 | 585.0    |
| 184.00 | 1407.0   |
| 185.00 | 5285.0   |
| 186.00 | 38376.0  |
| 187.00 | 11790.0  |
| 188.10 | 1715.0   |
| 189.00 | 3288.0   |
| 189.90 | 628.0    |
| 191.00 | 1610.0   |
| 191.90 | 3933.0   |
| 193.00 | 4179.0   |
| 194.10 | 963.0    |
| 194.70 | 326.0    |
| 196.00 | 7884.0   |
| 197.90 | 258880.0 |
| 199.00 | 15484.0  |
| 200.00 | 1739.0   |
| 201.40 | 1468.0   |
| 202.20 | 397.0    |
| 202.90 | 2178.0   |
| 204.00 | 12056.0  |
| 205.00 | 18352.0  |
| 206.00 | 72768.0  |
| 207.00 | 12191.0  |
| 208.00 | 2731.0   |
| 208.90 | 1589.0   |
| 210.20 | 1343.0   |
| 211.00 | 3560.0   |
| 212.90 | 325.0    |
| 215.00 | 1020.0   |
| 216.00 | 1767.0   |
| 217.00 | 20576.0  |
| 217.90 | 3013.0   |
| 221.00 | 12180.0  |
| 222.00 | 3224.0   |
| 223.00 | 5261.0   |
| 224.00 | 41720.0  |
| 225.00 | 11365.0  |
| 226.00 | 1260.0   |
| 227.00 | 21160.0  |
| 228.00 | 2986.0   |
| 228.90 | 3494.0   |
| 229.90 | 602.0    |
| 231.00 | 1273.0   |
| 232.00 | 360.0    |
| 232.80 | 364.0    |
| 234.00 | 1326.0   |
| 234.90 | 934.0    |
| 236.00 | 1254.0   |
| 237.00 | 3258.0   |
| 237.90 | 311.0    |
| 238.80 | 745.0    |
| 239.10 | 663.0    |
| 240.00 | 668.0    |
| 241.00 | 878.0    |
| 242.00 | 2349.0   |
| 243.00 | 2381.0   |
| 244.00 | 32976.0  |
| 245.00 | 4788.0   |
| 245.90 | 6589.0   |
| 247.00 | 1720.0   |
| 247.80 | 320.0    |
| 249.00 | 1583.0   |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP62

|        |          |
|--------|----------|
| 250.80 | 320.0    |
| 252.90 | 856.0    |
| 255.00 | 176128.0 |
| 255.90 | 25632.0  |
| 256.90 | 2177.0   |
| 258.00 | 13126.0  |
| 258.90 | 1730.0   |
| 259.80 | 437.0    |
| 261.20 | 382.0    |
| 263.90 | 367.0    |
| 264.90 | 4372.0   |
| 266.00 | 468.0    |
| 270.90 | 646.0    |
| 271.90 | 623.0    |
| 273.00 | 5588.0   |
| 274.00 | 13659.0  |
| 275.00 | 82752.0  |
| 276.00 | 9336.0   |
| 277.00 | 8435.0   |
| 278.00 | 1029.0   |
| 281.10 | 1353.0   |
| 282.90 | 716.0    |
| 284.10 | 548.0    |
| 285.00 | 1123.0   |
| 288.90 | 444.0    |
| 290.10 | 349.0    |
| 291.90 | 380.0    |
| 293.10 | 1489.0   |
| 293.90 | 791.0    |
| 295.10 | 933.0    |
| 295.90 | 25376.0  |
| 297.00 | 4291.0   |
| 302.00 | 482.0    |
| 303.00 | 2713.0   |
| 304.10 | 582.0    |
| 310.10 | 424.0    |
| 314.00 | 1197.0   |
| 315.00 | 3344.0   |
| 316.00 | 1827.0   |
| 320.90 | 846.0    |
| 323.00 | 8106.0   |
| 324.00 | 1587.0   |
| 326.00 | 326.0    |
| 327.00 | 1677.0   |
| 327.70 | 698.0    |
| 328.00 | 575.0    |
| 332.10 | 902.0    |
| 332.90 | 735.0    |
| 334.00 | 4441.0   |
| 335.00 | 1516.0   |
| 341.00 | 1207.0   |
| 341.90 | 530.0    |
| 345.80 | 2042.0   |
| 346.70 | 332.0    |
| 352.00 | 2941.0   |
| 353.00 | 2361.0   |
| 354.00 | 3034.0   |
| 354.90 | 704.0    |
| 364.90 | 13587.0  |
| 365.90 | 1428.0   |
| 369.90 | 456.0    |
| 371.00 | 631.0    |
| 372.00 | 4225.0   |
| 373.10 | 873.0    |
| 383.00 | 1106.0   |
| 389.90 | 919.0    |
| 391.00 | 631.0    |
| 401.90 | 2676.0   |
| 403.00 | 3029.0   |
| 404.10 | 1387.0   |
| 420.90 | 2846.0   |
| 422.00 | 3055.0   |
| 422.90 | 15972.0  |
| 424.00 | 5321.0   |
| 424.80 | 442.0    |
| 439.80 | 586.0    |
| 441.00 | 45440.0  |
| 442.00 | 289408.0 |
| 443.00 | 48136.0  |
| 444.00 | 5266.0   |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP62