

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060225\
 Data File : BN037134.D
 Acq On : 02 Jun 2025 16:21
 Operator : RC/JU
 Sample : SSTDICC0.2
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2

Quant Time: Jun 03 05:32:15 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 03 02:42:35 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.604	152	1644	0.400	ng	0.00
7) Naphthalene-d8	10.383	136	4276	0.400	ng	0.00
13) Acenaphthene-d10	14.245	164	2554	0.400	ng	0.00
19) Phenanthrene-d10	16.996	188	5606	0.400	ng	0.00
29) Chrysene-d12	21.189	240	3943	0.400	ng	0.00
35) Perylene-d12	23.380	264	3415	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.199	112	848	0.207	ng	0.00
5) Phenol-d6	6.781	99	1015	0.199	ng	0.00
8) Nitrobenzene-d5	8.760	82	846	0.186	ng	0.01
11) 2-Methylnaphthalene-d10	11.981	152	1174	0.192	ng	0.00
14) 2,4,6-Tribromophenol	15.743	330	239	0.199	ng	0.00
15) 2-Fluorobiphenyl	12.873	172	2190	0.207	ng	0.00
27) Fluoranthene-d10	19.031	212	2796	0.184	ng	0.00
31) Terphenyl-d14	19.640	244	1884	0.203	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.133	88	492	0.222	ng	98
3) n-Nitrosodimethylamine	3.451	42	847	0.197	ng	# 96
6) bis(2-Chloroethyl)ether	7.041	93	841	0.183	ng	91
9) Naphthalene	10.426	128	2430	0.198	ng	97
10) Hexachlorobutadiene	10.725	225	542	0.205	ng	# 99
12) 2-Methylnaphthalene	12.057	142	1581	0.194	ng	96
16) Acenaphthylene	13.967	152	2408	0.194	ng	99
17) Acenaphthene	14.309	154	1574	0.195	ng	100
18) Fluorene	15.304	166	2173	0.196	ng	98
20) 4,6-Dinitro-2-methylph...	15.400	198	188	0.274	ng	# 88
21) 4-Bromophenyl-phenylether	16.202	248	676	0.190	ng	96
22) Hexachlorobenzene	16.301	284	790	0.204	ng	98
23) Atrazine	16.475	200	591	0.184	ng	# 88
24) Pentachlorophenol	16.661	266	357	0.169	ng	95
25) Phenanthrene	17.034	178	3508	0.193	ng	100
26) Anthracene	17.133	178	3044	0.183	ng	99
28) Fluoranthene	19.063	202	3896	0.182	ng	100
30) Pyrene	19.426	202	3907	0.209	ng	99
32) Benzo(a)anthracene	21.171	228	2732	0.191	ng	96
33) Chrysene	21.225	228	3278	0.206	ng	99
34) Bis(2-ethylhexyl)phtha...	21.117	149	2079	0.204	ng	100
36) Indeno(1,2,3-cd)pyrene	25.585	276	2336	0.186	ng	100
37) Benzo(b)fluoranthene	22.725	252	2729	0.195	ng	# 89
38) Benzo(k)fluoranthene	22.769	252	2636	0.188	ng	# 86
39) Benzo(a)pyrene	23.287	252	2238	0.194	ng	# 84
40) Dibenzo(a,h)anthracene	25.602	278	1652	0.176	ng	# 83
41) Benzo(g,h,i)perylene	26.257	276	2187	0.196	ng	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

BNA N

ClientSampleId :
SSTDICC0.2


Quant Time: Jun 03 05:32:15 2025


Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060225.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Tue Jun 03 02:42:35 2025

Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.604 min Scan# 61
 Delta R.T. -0.000 min
 Lab File: BN037134.D
 Acq: 02 Jun 2025 16:21




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2


Tgt Ion:152 Resp: 1644
 Ion Ratio Lower Upper
 152 100
 150 154.5 120.6 180.8
 115 71.7 56.6 85.0



#2
 1,4-Dioxane
 Concen: 0.222 ng
 RT: 3.133 min Scan# 35
 Delta R.T. 0.007 min
 Lab File: BN037134.D
 Acq: 02 Jun 2025 16:21

Tgt Ion: 88 Resp: 492
 Ion Ratio Lower Upper
 88 100
 43 48.8 41.6 62.4
 58 82.5 65.8 98.8









#3
n-Nitrosodimethylamine
Concen: 0.197 ng
RT: 3.451 min Scan# 79
Delta R.T. 0.007 min
Lab File: BN037134.D
Acq: 02 Jun 2025 16:21




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion: 42 Resp: 847
Ion Ratio Lower Upper
42 100
74 67.9 53.0 79.4
44 16.1 7.2 10.8#



#4
2-Fluorophenol
Concen: 0.207 ng
RT: 5.199 min Scan# 321
Delta R.T. 0.000 min
Lab File: BN037134.D
Acq: 02 Jun 2025 16:21

Tgt Ion: 112 Resp: 848
Ion Ratio Lower Upper
112 100
64 71.3 56.0 84.0
63 44.5 37.0 55.4









#5
Phenol-d6
Concen: 0.199 ng
RT: 6.781 min Scan# 540
Delta R.T. 0.000 min
Lab File: BN037134.D
Acq: 02 Jun 2025 16:21




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion: 99 Resp: 1015
Ion Ratio Lower Upper
99 100
42 39.8 30.5 45.7
71 44.7 39.1 58.7



#6
bis(2-Chloroethyl)ether
Concen: 0.183 ng
RT: 7.041 min Scan# 576
Delta R.T. 0.007 min
Lab File: BN037134.D
Acq: 02 Jun 2025 16:21

Tgt Ion: 93 Resp: 841
Ion Ratio Lower Upper
93 100
63 93.8 68.4 102.6
95 34.7 24.2 36.4





#7

Naphthalene-d8

Concen: 0.400 ng

RT: 10.383 min Scan# 946

Delta R.T. 0.000 min

Lab File: BN037134.D

Acq: 02 Jun 2025 16:21

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2

Tgt Ion:136 Resp: 4276





Ion Ratio Lower Upper

136 100

137 12.8 10.3 15.5

54 12.3 9.9 14.9

68 7.2 5.9 8.9



#8

Nitrobenzene-d5

Concen: 0.186 ng

RT: 8.760 min Scan# 794

Delta R.T. 0.011 min

Lab File: BN037134.D

Acq: 02 Jun 2025 16:21




Tgt Ion: 82 Resp: 846


Ion Ratio Lower Upper

82 100

128 44.5 28.2 42.2#

54 71.1 60.2 90.2









#9
Naphthalene
Concen: 0.198 ng
RT: 10.426 min Scan# 950
Delta R.T. 0.000 min
Lab File: BN037134.D
Acq: 02 Jun 2025 16:21




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:128 Resp: 2430
Ion Ratio Lower Upper
128 100
129 14.0 10.2 15.4
127 16.7 12.6 18.8



#10
Hexachlorobutadiene
Concen: 0.205 ng
RT: 10.725 min Scan# 978
Delta R.T. 0.000 min
Lab File: BN037134.D
Acq: 02 Jun 2025 16:21

Tgt Ion:225 Resp: 542
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.2 51.8 77.8









#11
2-Methylnaphthalene-d10
Concen: 0.192 ng
RT: 11.981 min Scan# 1145
Delta R.T. 0.000 min
Lab File: BN037134.D
Acq: 02 Jun 2025 16:21




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:152 Resp: 1174
Ion Ratio Lower Upper
152 100
151 21.0 17.8 26.6



#12
2-Methylnaphthalene
Concen: 0.194 ng
RT: 12.057 min Scan# 1160
Delta R.T. 0.000 min
Lab File: BN037134.D
Acq: 02 Jun 2025 16:21

Tgt Ion:142 Resp: 1581
Ion Ratio Lower Upper
142 100
141 92.7 70.6 106.0
115 50.9 39.7 59.5









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.245 min Scan# 1463
Delta R.T. 0.000 min
Lab File: BN037134.D
Acq: 02 Jun 2025 16:21




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:164 Resp: 2554
Ion Ratio Lower Upper
164 100
162 107.0 85.1 127.7
160 56.8 44.2 66.4



#14
2,4,6-Tribromophenol
Concen: 0.199 ng
RT: 15.743 min Scan# 1600
Delta R.T. 0.000 min
Lab File: BN037134.D
Acq: 02 Jun 2025 16:21

Tgt Ion:330 Resp: 239
Ion Ratio Lower Upper
330 100
332 89.5 76.6 115.0
141 59.4 47.3 70.9









#15
2-Fluorobiphenyl
Concen: 0.207 ng
RT: 12.873 min Scan# 11
Delta R.T. 0.005 min
Lab File: BN037134.D
Acq: 02 Jun 2025 16:21




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:172 Resp: 2190
Ion Ratio Lower Upper
172 100
171 39.2 29.9 44.9
170 26.6 20.5 30.7



#16
Acenaphthylene
Concen: 0.194 ng
RT: 13.967 min Scan# 1437
Delta R.T. 0.000 min
Lab File: BN037134.D
Acq: 02 Jun 2025 16:21

Tgt Ion:152 Resp: 2408
Ion Ratio Lower Upper
152 100
151 20.2 16.4 24.6
153 13.2 10.5 15.7





#17

Acenaphthene

Concen: 0.195 ng

RT: 14.309 min Scan# 1469

Delta R.T. 0.000 min

Lab File: BN037134.D

Acq: 02 Jun 2025 16:21

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:154 Resp: 1574

Ion Ratio Lower Upper

154 100

153 119.4 95.4 143.2

152 63.3 50.6 75.8



#18

Fluorene

Concen: 0.196 ng

RT: 15.304 min Scan# 1562

Delta R.T. 0.000 min

Lab File: BN037134.D

Acq: 02 Jun 2025 16:21




Tgt Ion:166 Resp: 2173


Ion Ratio Lower Upper

166 100

165 99.5 81.4 122.0

167 13.0 10.9 16.3





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 16.996 min Scan# 11

Delta R.T. 0.000 min

Lab File: BN037134.D

Acq: 02 Jun 2025 16:21

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:188 Resp: 5606

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 10.9 8.4 12.6



#20

4,6-Dinitro-2-methylphenol

Concen: 0.274 ng

RT: 15.400 min Scan# 1571

Delta R.T. 0.011 min

Lab File: BN037134.D

Acq: 02 Jun 2025 16:21




Tgt Ion:198 Resp: 188


Ion Ratio Lower Upper

198 100

51 147.1 109.0 163.6

105 68.6 43.0 64.4#









#21
4-Bromophenyl-phenylether
Concen: 0.190 ng
RT: 16.202 min Scan# 1637
Delta R.T. 0.000 min
Lab File: BN037134.D
Acq: 02 Jun 2025 16:21




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:248 Resp: 676
Ion Ratio Lower Upper
248 100
250 104.4 83.8 125.6
141 60.0 41.9 62.9



#22
Hexachlorobenzene
Concen: 0.204 ng
RT: 16.301 min Scan# 1645
Delta R.T. 0.000 min
Lab File: BN037134.D
Acq: 02 Jun 2025 16:21

Tgt Ion:284 Resp: 790
Ion Ratio Lower Upper
284 100
142 51.4 43.0 64.6
249 36.5 29.4 44.0









#23
Atrazine
Concen: 0.184 ng
RT: 16.475 min Scan# 1659
Delta R.T. 0.000 min
Lab File: BN037134.D
Acq: 02 Jun 2025 16:21




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:200 Resp: 591
Ion Ratio Lower Upper
200 100
173 35.7 22.3 33.5#
215 54.9 38.3 57.5



#24
Pentachlorophenol
Concen: 0.169 ng
RT: 16.661 min Scan# 1674
Delta R.T. 0.013 min
Lab File: BN037134.D
Acq: 02 Jun 2025 16:21

Tgt Ion:266 Resp: 357
Ion Ratio Lower Upper
266 100
264 65.8 49.0 73.4
268 61.1 51.6 77.4





#25

Phenanthrene

Concen: 0.193 ng

RT: 17.034 min Scan# 1704

Delta R.T. 0.000 min

Lab File: BN037134.D

Acq: 02 Jun 2025 16:21

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:178 Resp: 3508

Ion Ratio Lower Upper

178 100

176 20.1 16.1 24.1

179 15.8 12.4 18.6



#26

Anthracene

Concen: 0.183 ng

RT: 17.133 min Scan# 1712

Delta R.T. 0.000 min

Lab File: BN037134.D

Acq: 02 Jun 2025 16:21




Tgt Ion:178 Resp: 3044


Ion Ratio Lower Upper

178 100

176 19.3 15.0 22.4

179 15.7 12.3 18.5









#27
Fluoranthene-d10
Concen: 0.184 ng
RT: 19.031 min Scan# 1935
Delta R.T. 0.000 min
Lab File: BN037134.D
Acq: 02 Jun 2025 16:21




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:212 Resp: 2796
Ion Ratio Lower Upper
212 100
106 13.2 10.8 16.2
104 8.6 6.8 10.2



#28
Fluoranthene
Concen: 0.182 ng
RT: 19.063 min Scan# 1942
Delta R.T. 0.000 min
Lab File: BN037134.D
Acq: 02 Jun 2025 16:21

Tgt Ion:202 Resp: 3896
Ion Ratio Lower Upper
202 100
101 10.7 8.4 12.6
203 16.8 13.6 20.4





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.189 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN037134.D

Acq: 02 Jun 2025 16:21

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:240 Resp: 3943

Ion Ratio Lower Upper

240 100

120 14.0 10.5 15.7

236 29.8 23.0 34.6



#30

Pyrene

Concen: 0.209 ng

RT: 19.426 min Scan# 2020

Delta R.T. 0.000 min

Lab File: BN037134.D

Acq: 02 Jun 2025 16:21




Tgt Ion:202 Resp: 3907


Ion Ratio Lower Upper

202 100

200 21.8 17.0 25.6

203 18.3 14.2 21.4





#31

Terphenyl-d14

Concen: 0.203 ng

RT: 19.640 min Scan# 2066

Delta R.T. 0.000 min

Lab File: BN037134.D

Acq: 02 Jun 2025 16:21

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:244 Resp: 1884

Ion Ratio Lower Upper

244 100

212 13.0 9.2 13.8

122 16.5 13.2 19.8



#32

Benzo(a)anthracene

Concen: 0.191 ng

RT: 21.171 min Scan# 2275

Delta R.T. 0.000 min

Lab File: BN037134.D

Acq: 02 Jun 2025 16:21




Tgt Ion:228 Resp: 2732


Ion Ratio Lower Upper

228 100

226 30.4 23.0 34.4

229 21.7 15.8 23.8





#33

Chrysene

Concen: 0.206 ng

RT: 21.225 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN037134.D


Acq: 02 Jun 2025 16:21

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:228 Resp: 3278

Ion Ratio Lower Upper

228 100

226 32.2 25.7 38.5

229 21.4 16.3 24.5



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.204 ng

RT: 21.117 min Scan# 2269

Delta R.T. 0.000 min

Lab File: BN037134.D

Acq: 02 Jun 2025 16:21




Tgt Ion:149 Resp: 2079


Ion Ratio Lower Upper

149 100

167 25.8 20.7 31.1

279 4.0 2.8 4.2





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.380 min Scan# 21

Delta R.T. -0.003 min

Lab File: BN037134.D

Acq: 02 Jun 2025 16:21

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:264 Resp: 3415

Ion Ratio Lower Upper

264 100

260 27.5 21.4 32.2

265 70.2 51.2 76.8



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.186 ng

RT: 25.585 min Scan# 3500

Delta R.T. 0.000 min

Lab File: BN037134.D

Acq: 02 Jun 2025 16:21




Tgt Ion:276 Resp: 2336


Ion Ratio Lower Upper

276 100

138 24.4 19.5 29.3

277 24.5 19.5 29.3









#37
Benzo(b)fluoranthene
Concen: 0.195 ng
RT: 22.725 min Scan# 2523
Delta R.T. -0.003 min
Lab File: BN037134.D
Acq: 02 Jun 2025 16:21




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:252 Resp: 2729
Ion Ratio Lower Upper
252 100
253 33.8 22.6 34.0
125 21.8 13.0 19.6#



#38
Benzo(k)fluoranthene
Concen: 0.188 ng
RT: 22.769 min Scan# 2537
Delta R.T. 0.000 min
Lab File: BN037134.D
Acq: 02 Jun 2025 16:21

Tgt Ion:252 Resp: 2636
Ion Ratio Lower Upper
252 100
253 35.0 21.8 32.8#
125 22.4 13.2 19.8#





#39

Benzo(a)pyrene

Concen: 0.194 ng

RT: 23.287 min Scan# 21

Delta R.T. -0.003 min

Lab File: BN037134.D

Acq: 02 Jun 2025 16:21

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:252 Resp: 2238

Ion Ratio Lower Upper

252 100

253 40.7 24.8 37.2#

125 30.1 18.7 28.1#



#40

Dibenzo(a,h)anthracene

Concen: 0.176 ng

RT: 25.602 min Scan# 3506

Delta R.T. 0.000 min

Lab File: BN037134.D

Acq: 02 Jun 2025 16:21




Tgt Ion:278 Resp: 1652


Ion Ratio Lower Upper

278 100

139 26.7 17.2 25.8#

279 44.0 26.0 39.0#





#41
 Benzo(g,h,i)perylene
 Concen: 0.196 ng
 RT: 26.257 min Scan# 31
 Delta R.T. 0.003 min
 Lab File: BN037134.D
 Acq: 02 Jun 2025 16:21

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2

Tgt Ion:	276	Resp:	2187
Ion	Ratio	Lower	Upper
276	100		
277	30.4	21.9	32.9
138	30.4	20.7	31.1

