


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060225\
 Data File : BN037137.D
 Acq On : 02 Jun 2025 18:10
 Operator : RC/JU
 Sample : SSTDICC1.6
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Quant Time: Jun 03 05:33:21 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 03 02:42:35 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.604	152	1808	0.400	ng	0.00
7) Naphthalene-d8	10.383	136	4689	0.400	ng	0.00
13) Acenaphthene-d10	14.245	164	2878	0.400	ng	0.00
19) Phenanthrene-d10	16.996	188	6035	0.400	ng	0.00
29) Chrysene-d12	21.189	240	5242	0.400	ng	0.00
35) Perylene-d12	23.380	264	4718	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.199	112	7450	1.655	ng	0.00
5) Phenol-d6	6.781	99	9446	1.685	ng	0.00
8) Nitrobenzene-d5	8.749	82	8568	1.718	ng	0.00
11) 2-Methylnaphthalene-d10	11.976	152	11454	1.709	ng	0.00
14) 2,4,6-Tribromophenol	15.743	330	2302	1.700	ng	0.00
15) 2-Fluorobiphenyl	12.868	172	20620	1.729	ng	0.00
27) Fluoranthene-d10	19.031	212	28205	1.727	ng	0.00
31) Terphenyl-d14	19.635	244	20650	1.674	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.126	88	3689	1.514	ng	94
3) n-Nitrosodimethylamine	3.437	42	8209	1.732	ng	# 98
6) bis(2-Chloroethyl)ether	7.033	93	8857	1.748	ng	97
9) Naphthalene	10.426	128	22731	1.687	ng	96
10) Hexachlorobutadiene	10.725	225	4806	1.661	ng	# 98
12) 2-Methylnaphthalene	12.052	142	15454	1.728	ng	98
16) Acenaphthylene	13.967	152	24069	1.720	ng	99
17) Acenaphthene	14.309	154	15563	1.710	ng	97
18) Fluorene	15.304	166	21401	1.709	ng	99
20) 4,6-Dinitro-2-methylph...	15.389	198	2750	1.580	ng	# 35
21) 4-Bromophenyl-phenylether	16.190	248	6677	1.747	ng	# 70
22) Hexachlorobenzene	16.301	284	6981	1.671	ng	100
23) Atrazine	16.463	200	6176	1.786	ng	91
24) Pentachlorophenol	16.649	266	4045	1.780	ng	98
25) Phenanthrene	17.034	178	33869	1.734	ng	99
26) Anthracene	17.120	178	31530	1.759	ng	99
28) Fluoranthene	19.059	202	40811	1.772	ng	99
30) Pyrene	19.421	202	41151	1.658	ng	100
32) Benzo(a)anthracene	21.171	228	33633	1.767	ng	99
33) Chrysene	21.225	228	35461	1.678	ng	98
34) Bis(2-ethylhexyl)phtha...	21.117	149	21681	1.598	ng	99
36) Indeno(1,2,3-cd)pyrene	25.582	276	30852	1.782	ng	95
37) Benzo(b)fluoranthene	22.723	252	32961	1.707	ng	# 90
38) Benzo(k)fluoranthene	22.766	252	33274	1.718	ng	# 91
39) Benzo(a)pyrene	23.284	252	27436	1.718	ng	# 85
40) Dibenzo(a,h)anthracene	25.594	278	23852	1.840	ng	# 90
41) Benzo(g,h,i)perylene	26.251	276	27100	1.754	ng	96





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.604 min Scan# 61
 Delta R.T. -0.000 min
 Lab File: BN037137.D
 Acq: 02 Jun 2025 18:10




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion:152 Resp: 1808
 Ion Ratio Lower Upper
 152 100
 150 148.0 120.6 180.8
 115 68.2 56.6 85.0



#2
 1,4-Dioxane
 Concen: 1.514 ng
 RT: 3.126 min Scan# 34
 Delta R.T. 0.000 min
 Lab File: BN037137.D
 Acq: 02 Jun 2025 18:10

Tgt Ion: 88 Resp: 3689
 Ion Ratio Lower Upper
 88 100
 43 54.0 41.6 62.4
 58 90.1 65.8 98.8









#3
n-Nitrosodimethylamine
Concen: 1.732 ng
RT: 3.437 min Scan# 71
Delta R.T. -0.007 min
Lab File: BN037137.D
Acq: 02 Jun 2025 18:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion: 42 Resp: 8209
Ion Ratio Lower Upper
42 100
74 67.0 53.0 79.4
44 5.3 7.2 10.8#



#4
2-Fluorophenol
Concen: 1.655 ng
RT: 5.199 min Scan# 321
Delta R.T. 0.000 min
Lab File: BN037137.D
Acq: 02 Jun 2025 18:10

Tgt Ion: 112 Resp: 7450
Ion Ratio Lower Upper
112 100
64 69.9 56.0 84.0
63 44.2 37.0 55.4









#5
Phenol-d6
Concen: 1.685 ng
RT: 6.781 min Scan# 540
Delta R.T. 0.000 min
Lab File: BN037137.D
Acq: 02 Jun 2025 18:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion: 99 Resp: 9446
Ion Ratio Lower Upper
99 100
42 38.5 30.5 45.7
71 47.2 39.1 58.7



#6
bis(2-Chloroethyl)ether
Concen: 1.748 ng
RT: 7.033 min Scan# 575
Delta R.T. 0.000 min
Lab File: BN037137.D
Acq: 02 Jun 2025 18:10

Tgt Ion: 93 Resp: 8857
Ion Ratio Lower Upper
93 100
63 88.5 68.4 102.6
95 31.4 24.2 36.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.383 min Scan# 946
Delta R.T. 0.000 min
Lab File: BN037137.D
Acq: 02 Jun 2025 18:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:136	Resp:	4689
Ion Ratio	Lower	Upper
136	100	
137	12.4	10.3 15.5
54	10.5	9.9 14.9
68	6.1	5.9 8.9



#8
Nitrobenzene-d5
Concen: 1.718 ng
RT: 8.749 min Scan# 793
Delta R.T. 0.000 min
Lab File: BN037137.D
Acq: 02 Jun 2025 18:10

Tgt Ion: 82	Resp:	8568
Ion Ratio	Lower	Upper
82	100	
128	34.9	28.2 42.2
54	68.1	60.2 90.2









#9
Naphthalene
Concen: 1.687 ng
RT: 10.426 min Scan# 950
Delta R.T. 0.000 min
Lab File: BN037137.D
Acq: 02 Jun 2025 18:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	128	Resp:	22731
Ion Ratio	Lower	Upper	
128	100		
129	11.3	10.2	15.4
127	14.1	12.6	18.8



#10
Hexachlorobutadiene
Concen: 1.661 ng
RT: 10.725 min Scan# 978
Delta R.T. 0.000 min
Lab File: BN037137.D
Acq: 02 Jun 2025 18:10

Tgt Ion:	225	Resp:	4806
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	63.5	51.8	77.8









#11
2-Methylnaphthalene-d10
Concen: 1.709 ng
RT: 11.976 min Scan# 1144
Delta R.T. -0.005 min
Lab File: BN037137.D
Acq: 02 Jun 2025 18:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:152 Resp: 11454
Ion Ratio Lower Upper
152 100
151 21.7 17.8 26.6



#12
2-Methylnaphthalene
Concen: 1.728 ng
RT: 12.052 min Scan# 1159
Delta R.T. -0.005 min
Lab File: BN037137.D
Acq: 02 Jun 2025 18:10

Tgt Ion:142 Resp: 15454
Ion Ratio Lower Upper
142 100
141 89.6 70.6 106.0
115 46.7 39.7 59.5









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.245 min Scan# 1463
Delta R.T. 0.000 min
Lab File: BN037137.D
Acq: 02 Jun 2025 18:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:164 Resp: 2878
Ion Ratio Lower Upper
164 100
162 108.8 85.1 127.7
160 55.5 44.2 66.4



#14
2,4,6-Tribromophenol
Concen: 1.700 ng
RT: 15.743 min Scan# 1600
Delta R.T. 0.000 min
Lab File: BN037137.D
Acq: 02 Jun 2025 18:10

Tgt Ion:330 Resp: 2302
Ion Ratio Lower Upper
330 100
332 95.5 76.6 115.0
141 63.2 47.3 70.9









#15
2-Fluorobiphenyl
Concen: 1.729 ng
RT: 12.868 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN037137.D
Acq: 02 Jun 2025 18:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:172 Resp: 20620
Ion Ratio Lower Upper
172 100
171 36.4 29.9 44.9
170 24.4 20.5 30.7



#16
Acenaphthylene
Concen: 1.720 ng
RT: 13.967 min Scan# 1437
Delta R.T. 0.000 min
Lab File: BN037137.D
Acq: 02 Jun 2025 18:10

Tgt Ion:152 Resp: 24069
Ion Ratio Lower Upper
152 100
151 20.2 16.4 24.6
153 12.9 10.5 15.7









#17
Acenaphthene
Concen: 1.710 ng
RT: 14.309 min Scan# 1469
Delta R.T. 0.000 min
Lab File: BN037137.D
Acq: 02 Jun 2025 18:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:154 Resp: 15563
Ion Ratio Lower Upper
154 100
153 116.7 95.4 143.2
152 61.0 50.6 75.8



#18
Fluorene
Concen: 1.709 ng
RT: 15.304 min Scan# 1562
Delta R.T. 0.000 min
Lab File: BN037137.D
Acq: 02 Jun 2025 18:10

Tgt Ion:166 Resp: 21401
Ion Ratio Lower Upper
166 100
165 100.4 81.4 122.0
167 13.3 10.9 16.3





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 16.996 min Scan# 11

Delta R.T. 0.000 min

Lab File: BN037137.D

Acq: 02 Jun 2025 18:10

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:188 Resp: 6035

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 9.2 8.4 12.6



#20

4,6-Dinitro-2-methylphenol

Concen: 1.580 ng

RT: 15.389 min Scan# 1570

Delta R.T. 0.000 min

Lab File: BN037137.D

Acq: 02 Jun 2025 18:10




Tgt Ion:198 Resp: 2750


Ion Ratio Lower Upper

198 100

51 43.6 109.0 163.6#

105 30.2 43.0 64.4#









#21
4-Bromophenyl-phenylether
Concen: 1.747 ng
RT: 16.190 min Scan# 1636
Delta R.T. -0.012 min
Lab File: BN037137.D
Acq: 02 Jun 2025 18:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:248 Resp: 6677
Ion Ratio Lower Upper
248 100
250 91.8 83.8 125.6
141 98.9 41.9 62.9#



#22
Hexachlorobenzene
Concen: 1.671 ng
RT: 16.301 min Scan# 1645
Delta R.T. 0.000 min
Lab File: BN037137.D
Acq: 02 Jun 2025 18:10

Tgt Ion:284 Resp: 6981
Ion Ratio Lower Upper
284 100
142 53.3 43.0 64.6
249 36.7 29.4 44.0









#23
Atrazine
Concen: 1.786 ng
RT: 16.463 min Scan# 1658
Delta R.T. -0.012 min
Lab File: BN037137.D
Acq: 02 Jun 2025 18:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:200 Resp: 6176
Ion Ratio Lower Upper
200 100
173 32.8 22.3 33.5
215 42.2 38.3 57.5



#24
Pentachlorophenol
Concen: 1.780 ng
RT: 16.649 min Scan# 1673
Delta R.T. 0.000 min
Lab File: BN037137.D
Acq: 02 Jun 2025 18:10

Tgt Ion:266 Resp: 4045
Ion Ratio Lower Upper
266 100
264 63.6 49.0 73.4
268 64.4 51.6 77.4





#25

Phenanthrene

Concen: 1.734 ng

RT: 17.034 min Scan# 1710

Delta R.T. 0.000 min

Lab File: BN037137.D

Acq: 02 Jun 2025 18:10

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:178 Resp: 33869

Ion Ratio Lower Upper

178 100

176 19.7 16.1 24.1

179 15.2 12.4 18.6



#26

Anthracene

Concen: 1.759 ng

RT: 17.120 min Scan# 1711

Delta R.T. -0.012 min

Lab File: BN037137.D

Acq: 02 Jun 2025 18:10




Tgt Ion:178 Resp: 31530


Ion Ratio Lower Upper

178 100

176 19.0 15.0 22.4

179 15.0 12.3 18.5









#27
Fluoranthene-d10
Concen: 1.727 ng
RT: 19.031 min Scan# 1935
Delta R.T. 0.000 min
Lab File: BN037137.D
Acq: 02 Jun 2025 18:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:212 Resp: 28205
Ion Ratio Lower Upper
212 100
106 13.9 10.8 16.2
104 8.3 6.8 10.2



#28
Fluoranthene
Concen: 1.772 ng
RT: 19.059 min Scan# 1941
Delta R.T. -0.004 min
Lab File: BN037137.D
Acq: 02 Jun 2025 18:10

Tgt Ion:202 Resp: 40811
Ion Ratio Lower Upper
202 100
101 10.8 8.4 12.6
203 17.3 13.6 20.4





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.189 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN037137.D

Acq: 02 Jun 2025 18:10

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:240 Resp: 5242

Ion Ratio Lower Upper

240 100

120 12.0 10.5 15.7

236 28.6 23.0 34.6



#30

Pyrene

Concen: 1.658 ng

RT: 19.421 min Scan# 2019

Delta R.T. -0.004 min

Lab File: BN037137.D

Acq: 02 Jun 2025 18:10




Tgt Ion:202 Resp: 41151


Ion Ratio Lower Upper

202 100

200 21.4 17.0 25.6

203 17.7 14.2 21.4





#31

Terphenyl-d14

Concen: 1.674 ng

RT: 19.635 min Scan# 2065

Delta R.T. -0.004 min

Lab File: BN037137.D


Acq: 02 Jun 2025 18:10

Instrument :



BNA_N

ClientSampleId :

SSTDICC1.6



Tgt Ion:244	Resp:	20650
Ion Ratio	Lower	Upper
244	100	
212	10.9	9.2 13.8
122	16.2	13.2 19.8



#32

Benzo(a)anthracene

Concen: 1.767 ng




RT: 21.171 min Scan# 2275


Delta R.T. 0.000 min

Lab File: BN037137.D

Acq: 02 Jun 2025 18:10

Tgt Ion:228	Resp:	33633
Ion Ratio	Lower	Upper
228	100	
226	28.1	23.0 34.4
229	19.0	15.8 23.8





#33

Chrysene

Concen: 1.678 ng

RT: 21.225 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN037137.D

Acq: 02 Jun 2025 18:10

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:228 Resp: 35461

Ion Ratio Lower Upper

228 100

226 30.8 25.7 38.5

229 19.8 16.3 24.5



#34

Bis(2-ethylhexyl)phthalate

Concen: 1.598 ng

RT: 21.117 min Scan# 2269

Delta R.T. 0.000 min

Lab File: BN037137.D

Acq: 02 Jun 2025 18:10




Tgt Ion:149 Resp: 21681


Ion Ratio Lower Upper

149 100

167 26.5 20.7 31.1

279 2.9 2.8 4.2









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.380 min Scan# 21
Delta R.T. -0.003 min
Lab File: BN037137.D
Acq: 02 Jun 2025 18:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:264 Resp: 4718
Ion Ratio Lower Upper
264 100
260 27.3 21.4 32.2
265 60.0 51.2 76.8



#36
Indeno(1,2,3-cd)pyrene
Concen: 1.782 ng
RT: 25.582 min Scan# 3499
Delta R.T. -0.003 min
Lab File: BN037137.D
Acq: 02 Jun 2025 18:10

Tgt Ion:276 Resp: 30852
Ion Ratio Lower Upper
276 100
138 29.0 19.5 29.3
277 24.8 19.5 29.3









#37
Benzo(b)fluoranthene
Concen: 1.707 ng
RT: 22.723 min Scan# 2521
Delta R.T. -0.006 min
Lab File: BN037137.D
Acq: 02 Jun 2025 18:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:252 Resp: 32961
Ion Ratio Lower Upper
252 100
253 22.7 22.6 34.0
125 12.2 13.0 19.6#



#38
Benzo(k)fluoranthene
Concen: 1.718 ng
RT: 22.766 min Scan# 2536
Delta R.T. -0.003 min
Lab File: BN037137.D
Acq: 02 Jun 2025 18:10


Tgt Ion:252 Resp: 33274
Ion Ratio Lower Upper
252 100
253 22.7 21.8 32.8
125 12.2 13.2 19.8#








#39
Benzo(a)pyrene
Concen: 1.718 ng
RT: 23.284 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN037137.D
Acq: 02 Jun 2025 18:10

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6







Tgt Ion:252 Resp: 27436
Ion Ratio Lower Upper
252 100
253 23.4 24.8 37.2#
125 14.7 18.7 28.1#



#40
Dibenzo(a,h)anthracene
Concen: 1.840 ng
RT: 25.594 min Scan# 3503
Delta R.T. -0.008 min
Lab File: BN037137.D
Acq: 02 Jun 2025 18:10


Tgt Ion:278 Resp: 23852
Ion Ratio Lower Upper
278 100
139 19.8 17.2 25.8
279 24.4 26.0 39.0#





#41
 Benzo(g,h,i)perylene
 Concen: 1.754 ng
 RT: 26.251 min Scan# 31
 Delta R.T. -0.003 min
 Lab File: BN037137.D
 Acq: 02 Jun 2025 18:10

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6



Tgt Ion:	276	Resp:	27100
Ion Ratio	Lower	Upper	
276	100		
277	23.8	21.9	32.9
138	26.0	20.7	31.1

