

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060225\
 Data File : BN037138.D
 Acq On : 02 Jun 2025 18:46
 Operator : RC/JU
 Sample : SSTDICC3.2
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Quant Time: Jun 03 05:33:44 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 03 02:42:35 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.604	152	2004	0.400	ng	0.00
7) Naphthalene-d8	10.383	136	5237	0.400	ng	0.00
13) Acenaphthene-d10	14.245	164	3291	0.400	ng	0.00
19) Phenanthrene-d10	16.996	188	6845	0.400	ng	0.00
29) Chrysene-d12	21.189	240	6282	0.400	ng	0.00
35) Perylene-d12	23.383	264	5329	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.199	112	15601	3.126	ng	0.00
5) Phenol-d6	6.781	99	20508	3.301	ng	0.00
8) Nitrobenzene-d5	8.749	82	18726	3.362	ng	0.00
11) 2-Methylnaphthalene-d10	11.976	152	24881	3.323	ng	0.00
14) 2,4,6-Tribromophenol	15.743	330	5230	3.378	ng	0.00
15) 2-Fluorobiphenyl	12.868	172	43288	3.175	ng	0.00
27) Fluoranthene-d10	19.031	212	63337	3.420	ng	0.00
31) Terphenyl-d14	19.635	244	46528	3.148	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.126	88	7657	2.835	ng	96
3) n-Nitrosodimethylamine	3.437	42	16662	3.171	ng	# 96
6) bis(2-Chloroethyl)ether	7.033	93	18839	3.355	ng	97
9) Naphthalene	10.426	128	48767	3.240	ng	96
10) Hexachlorobutadiene	10.725	225	10004	3.095	ng	# 98
12) 2-Methylnaphthalene	12.052	142	33758	3.380	ng	97
16) Acenaphthylene	13.967	152	53841	3.365	ng	99
17) Acenaphthene	14.309	154	34206	3.286	ng	97
18) Fluorene	15.304	166	47767	3.337	ng	98
20) 4,6-Dinitro-2-methylph...	15.389	198	6585	3.147	ng	# 30
21) 4-Bromophenyl-phenylether	16.190	248	14564	3.360	ng	# 67
22) Hexachlorobenzene	16.301	284	15148	3.197	ng	99
23) Atrazine	16.463	200	13883	3.539	ng	# 90
24) Pentachlorophenol	16.649	266	9416	3.654	ng	98
25) Phenanthrene	17.034	178	74411	3.360	ng	99
26) Anthracene	17.120	178	70837	3.484	ng	99
28) Fluoranthene	19.059	202	91411	3.500	ng	99
30) Pyrene	19.421	202	91956	3.091	ng	100
32) Benzo(a)anthracene	21.171	228	76902	3.372	ng	99
33) Chrysene	21.225	228	80251	3.168	ng	97
34) Bis(2-ethylhexyl)phtha...	21.117	149	51947	3.194	ng	# 100
36) Indeno(1,2,3-cd)pyrene	25.582	276	67201	3.436	ng	# 94
37) Benzo(b)fluoranthene	22.725	252	73509	3.370	ng	# 89
38) Benzo(k)fluoranthene	22.769	252	73922	3.378	ng	# 90
39) Benzo(a)pyrene	23.284	252	61029	3.384	ng	# 82
40) Dibenzo(a,h)anthracene	25.596	278	52284	3.572	ng	# 89
41) Benzo(g,h,i)perylene	26.251	276	57357	3.286	ng	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060225\
Data File : BN037138.D
Acq On : 02 Jun 2025 18:46
Operator : RC/JU
Sample : SSTDICC3.2
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Quant Time: Jun 03 05:33:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 03 02:42:35 2025
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.604 min Scan# 61
 Delta R.T. -0.000 min
 Lab File: BN037138.D
 Acq: 02 Jun 2025 18:46




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2


Tgt Ion:152 Resp: 2004
 Ion Ratio Lower Upper
 152 100
 150 148.9 120.6 180.8
 115 65.3 56.6 85.0



#2
 1,4-Dioxane
 Concen: 2.835 ng
 RT: 3.126 min Scan# 34
 Delta R.T. 0.000 min
 Lab File: BN037138.D
 Acq: 02 Jun 2025 18:46

Tgt Ion: 88 Resp: 7657
 Ion Ratio Lower Upper
 88 100
 43 51.2 41.6 62.4
 58 87.9 65.8 98.8









#3
n-Nitrosodimethylamine
Concen: 3.171 ng
RT: 3.437 min Scan# 71
Delta R.T. -0.007 min
Lab File: BN037138.D
Acq: 02 Jun 2025 18:46




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 42 Resp: 16662
Ion Ratio Lower Upper
42 100
74 69.1 53.0 79.4
44 5.4 7.2 10.8#



#4
2-Fluorophenol
Concen: 3.126 ng
RT: 5.199 min Scan# 321
Delta R.T. 0.000 min
Lab File: BN037138.D
Acq: 02 Jun 2025 18:46

Tgt Ion: 112 Resp: 15601
Ion Ratio Lower Upper
112 100
64 70.7 56.0 84.0
63 44.0 37.0 55.4









#5
Phenol-d6
Concen: 3.301 ng
RT: 6.781 min Scan# 540
Delta R.T. 0.000 min
Lab File: BN037138.D
Acq: 02 Jun 2025 18:46




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 99 Resp: 20508
Ion Ratio Lower Upper
99 100
42 37.8 30.5 45.7
71 46.8 39.1 58.7



#6
bis(2-Chloroethyl)ether
Concen: 3.355 ng
RT: 7.033 min Scan# 575
Delta R.T. 0.000 min
Lab File: BN037138.D
Acq: 02 Jun 2025 18:46

Tgt Ion: 93 Resp: 18839
Ion Ratio Lower Upper
93 100
63 88.6 68.4 102.6
95 31.9 24.2 36.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.383 min Scan# 946
Delta R.T. 0.000 min
Lab File: BN037138.D
Acq: 02 Jun 2025 18:46




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:	136	Resp:	5237
Ion	Ratio	Lower	Upper
136	100		
137	11.6	10.3	15.5
54	10.6	9.9	14.9
68	6.2	5.9	8.9



#8
Nitrobenzene-d5
Concen: 3.362 ng
RT: 8.749 min Scan# 793
Delta R.T. 0.000 min
Lab File: BN037138.D
Acq: 02 Jun 2025 18:46

Tgt Ion:	82	Resp:	18726
Ion	Ratio	Lower	Upper
82	100		
128	35.0	28.2	42.2
54	65.9	60.2	90.2









#9
Naphthalene
Concen: 3.240 ng
RT: 10.426 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN037138.D
Acq: 02 Jun 2025 18:46




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:128 Resp: 48767
Ion Ratio Lower Upper
128 100
129 11.1 10.2 15.4
127 13.9 12.6 18.8



#10
Hexachlorobutadiene
Concen: 3.095 ng
RT: 10.725 min Scan# 978
Delta R.T. 0.000 min
Lab File: BN037138.D
Acq: 02 Jun 2025 18:46

Tgt Ion:225 Resp: 10004
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.4 51.8 77.8









#11
2-Methylnaphthalene-d10
Concen: 3.323 ng
RT: 11.976 min Scan# 1144
Delta R.T. -0.005 min
Lab File: BN037138.D
Acq: 02 Jun 2025 18:46




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:152 Resp: 24881
Ion Ratio Lower Upper
152 100
151 21.7 17.8 26.6



#12
2-Methylnaphthalene
Concen: 3.380 ng
RT: 12.052 min Scan# 1159
Delta R.T. -0.005 min
Lab File: BN037138.D
Acq: 02 Jun 2025 18:46

Tgt Ion:142 Resp: 33758
Ion Ratio Lower Upper
142 100
141 90.7 70.6 106.0
115 46.8 39.7 59.5









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.245 min Scan# 1463
Delta R.T. 0.000 min
Lab File: BN037138.D
Acq: 02 Jun 2025 18:46




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:164 Resp: 3291
Ion Ratio Lower Upper
164 100
162 106.6 85.1 127.7
160 55.0 44.2 66.4



#14
2,4,6-Tribromophenol
Concen: 3.378 ng
RT: 15.743 min Scan# 1600
Delta R.T. 0.000 min
Lab File: BN037138.D
Acq: 02 Jun 2025 18:46

Tgt Ion:330 Resp: 5230
Ion Ratio Lower Upper
330 100
332 97.1 76.6 115.0
141 64.7 47.3 70.9









#15
2-Fluorobiphenyl
Concen: 3.175 ng
RT: 12.868 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN037138.D
Acq: 02 Jun 2025 18:46




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:172 Resp: 43288
Ion Ratio Lower Upper
172 100
171 35.7 29.9 44.9
170 23.8 20.5 30.7



#16
Acenaphthylene
Concen: 3.365 ng
RT: 13.967 min Scan# 1437
Delta R.T. 0.000 min
Lab File: BN037138.D
Acq: 02 Jun 2025 18:46

Tgt Ion:152 Resp: 53841
Ion Ratio Lower Upper
152 100
151 20.0 16.4 24.6
153 12.9 10.5 15.7





#17

Acenaphthene

Concen: 3.286 ng

RT: 14.309 min Scan# 1469

Delta R.T. 0.000 min

Lab File: BN037138.D

Acq: 02 Jun 2025 18:46

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:154 Resp: 34206

Ion Ratio Lower Upper

154 100

153 116.8 95.4 143.2

152 60.1 50.6 75.8



#18

Fluorene

Concen: 3.337 ng

RT: 15.304 min Scan# 1562

Delta R.T. 0.000 min

Lab File: BN037138.D

Acq: 02 Jun 2025 18:46




Tgt Ion:166 Resp: 47767


Ion Ratio Lower Upper

166 100

165 99.6 81.4 122.0

167 13.3 10.9 16.3





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 16.996 min Scan# 11

Delta R.T. 0.000 min

Lab File: BN037138.D

Acq: 02 Jun 2025 18:46

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:188 Resp: 6845

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 9.6 8.4 12.6



#20

4,6-Dinitro-2-methylphenol

Concen: 3.147 ng

RT: 15.389 min Scan# 1570

Delta R.T. 0.000 min

Lab File: BN037138.D

Acq: 02 Jun 2025 18:46




Tgt Ion:198 Resp: 6585


Ion Ratio Lower Upper

198 100

51 36.6 109.0 163.6#

105 26.6 43.0 64.4#









#21
4-Bromophenyl-phenylether
Concen: 3.360 ng
RT: 16.190 min Scan# 1636
Delta R.T. -0.012 min
Lab File: BN037138.D
Acq: 02 Jun 2025 18:46




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:248 Resp: 14564
Ion Ratio Lower Upper
248 100
250 91.4 83.8 125.6
141 104.3 41.9 62.9#



#22
Hexachlorobenzene
Concen: 3.197 ng
RT: 16.301 min Scan# 1645
Delta R.T. 0.000 min
Lab File: BN037138.D
Acq: 02 Jun 2025 18:46

Tgt Ion:284 Resp: 15148
Ion Ratio Lower Upper
284 100
142 53.2 43.0 64.6
249 36.3 29.4 44.0









#23
Atrazine
Concen: 3.539 ng
RT: 16.463 min Scan# 1658
Delta R.T. -0.012 min
Lab File: BN037138.D
Acq: 02 Jun 2025 18:46




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:200 Resp: 13883
Ion Ratio Lower Upper
200 100
173 34.4 22.3 33.5#
215 41.7 38.3 57.5



#24
Pentachlorophenol
Concen: 3.654 ng
RT: 16.649 min Scan# 1673
Delta R.T. 0.000 min
Lab File: BN037138.D
Acq: 02 Jun 2025 18:46

Tgt Ion:266 Resp: 9416
Ion Ratio Lower Upper
266 100
264 63.2 49.0 73.4
268 63.5 51.6 77.4









#25
Phenanthrene
Concen: 3.360 ng
RT: 17.034 min Scan# 1710
Delta R.T. 0.000 min
Lab File: BN037138.D
Acq: 02 Jun 2025 18:46




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:178 Resp: 74411
Ion Ratio Lower Upper
178 100
176 19.7 16.1 24.1
179 15.1 12.4 18.6



#26
Anthracene
Concen: 3.484 ng
RT: 17.120 min Scan# 1711
Delta R.T. -0.012 min
Lab File: BN037138.D
Acq: 02 Jun 2025 18:46

Tgt Ion:178 Resp: 70837
Ion Ratio Lower Upper
178 100
176 19.1 15.0 22.4
179 15.2 12.3 18.5









#27
Fluoranthene-d10
Concen: 3.420 ng
RT: 19.031 min Scan# 1935
Delta R.T. 0.000 min
Lab File: BN037138.D
Acq: 02 Jun 2025 18:46




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:212 Resp: 63337
Ion Ratio Lower Upper
212 100
106 14.2 10.8 16.2
104 8.5 6.8 10.2



#28
Fluoranthene
Concen: 3.500 ng
RT: 19.059 min Scan# 1941
Delta R.T. -0.004 min
Lab File: BN037138.D
Acq: 02 Jun 2025 18:46

Tgt Ion:202 Resp: 91411
Ion Ratio Lower Upper
202 100
101 11.3 8.4 12.6
203 17.1 13.6 20.4





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.189 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN037138.D

Acq: 02 Jun 2025 18:46

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:240 Resp: 6282

Ion Ratio Lower Upper

240 100

120 11.9 10.5 15.7

236 28.4 23.0 34.6



#30

Pyrene

Concen: 3.091 ng

RT: 19.421 min Scan# 2019

Delta R.T. -0.004 min

Lab File: BN037138.D

Acq: 02 Jun 2025 18:46




Tgt Ion:202 Resp: 91956


Ion Ratio Lower Upper

202 100

200 21.6 17.0 25.6

203 17.7 14.2 21.4









#31
 Terphenyl-d14
 Concen: 3.148 ng
 RT: 19.635 min Scan# 2066
 Delta R.T. -0.004 min
 Lab File: BN037138.D
 Acq: 02 Jun 2025 18:46




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2


Tgt Ion:244 Resp: 46528
 Ion Ratio Lower Upper
 244 100
 212 10.7 9.2 13.8
 122 16.4 13.2 19.8



#32
 Benzo(a)anthracene
 Concen: 3.372 ng
 RT: 21.171 min Scan# 2275
 Delta R.T. 0.000 min
 Lab File: BN037138.D
 Acq: 02 Jun 2025 18:46

Tgt Ion:228 Resp: 76902
 Ion Ratio Lower Upper
 228 100
 226 28.2 23.0 34.4
 229 19.2 15.8 23.8





#33

Chrysene

Concen: 3.168 ng

RT: 21.225 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN037138.D

Acq: 02 Jun 2025 18:46

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:228 Resp: 80251

Ion Ratio Lower Upper

228 100

226 30.4 25.7 38.5

229 19.4 16.3 24.5



#34

Bis(2-ethylhexyl)phthalate

Concen: 3.194 ng

RT: 21.117 min Scan# 2269

Delta R.T. 0.000 min

Lab File: BN037138.D

Acq: 02 Jun 2025 18:46




Tgt Ion:149 Resp: 51947


Ion Ratio Lower Upper

149 100

167 25.9 20.7 31.1

279 2.6 2.8 4.2#





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.383 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN037138.D

Acq: 02 Jun 2025 18:46

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:264 Resp: 5329

Ion Ratio Lower Upper

264 100

260 26.5 21.4 32.2

265 59.8 51.2 76.8



#36

Indeno(1,2,3-cd)pyrene

Concen: 3.436 ng

RT: 25.582 min Scan# 3499

Delta R.T. -0.003 min

Lab File: BN037138.D

Acq: 02 Jun 2025 18:46




Tgt Ion:276 Resp: 67201


Ion Ratio Lower Upper

276 100

138 30.1 19.5 29.3#


277 24.8 19.5 29.3







#37
Benzo(b)fluoranthene
Concen: 3.370 ng
RT: 22.725 min Scan# 2522
Delta R.T. -0.003 min
Lab File: BN037138.D
Acq: 02 Jun 2025 18:46

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2







Tgt Ion:252 Resp: 73509
Ion Ratio Lower Upper
252 100
253 22.0 22.6 34.0#
125 12.2 13.0 19.6#



#38
Benzo(k)fluoranthene
Concen: 3.378 ng
RT: 22.769 min Scan# 2537
Delta R.T. 0.000 min
Lab File: BN037138.D
Acq: 02 Jun 2025 18:46

Tgt Ion:252 Resp: 73922
Ion Ratio Lower Upper
252 100
253 22.1 21.8 32.8
125 12.1 13.2 19.8#









#39
Benzo(a)pyrene
Concen: 3.384 ng
RT: 23.284 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN037138.D
Acq: 02 Jun 2025 18:46




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:252 Resp: 61029
Ion Ratio Lower Upper
252 100
253 22.2 24.8 37.2#
125 13.9 18.7 28.1#



#40
Dibenzo(a,h)anthracene
Concen: 3.572 ng
RT: 25.596 min Scan# 3504
Delta R.T. -0.006 min
Lab File: BN037138.D
Acq: 02 Jun 2025 18:46


Tgt Ion:278 Resp: 52284
Ion Ratio Lower Upper
278 100
139 19.9 17.2 25.8
279 23.6 26.0 39.0#





#41
Benzo(g,h,i)perylene
Concen: 3.286 ng
RT: 26.251 min Scan# 31
Delta R.T. -0.003 min
Lab File: BN037138.D
Acq: 02 Jun 2025 18:46

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2



Tgt Ion:276 Resp: 57357
Ion Ratio Lower Upper
276 100
277 23.8 21.9 32.9
138 25.8 20.7 31.1

