

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060524\
 Data File : BN031811.D
 Acq On : 05 Jun 2024 22:23
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020275

Quant Time: Jun 06 00:02:13 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN052924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed May 29 15:13:07 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.675	152	367845	20.000	ng/u1	-0.01
20) Naphthalene-d8	10.457	136	1757902	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.316	164	1165465	20.000	ng/u1	-0.01
64) Phenanthrene-d10	17.063	188	2481456	20.000	ng/u1	0.00
79) Chrysene-d12	21.298	240	2488416	20.000	ng/u1	0.00
88) Perylene-d12	24.239	264	2889128	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.181	96	57958	6.518	ng/uL	-0.01
4) Pyridine-d5	3.587	84	456318	18.231	ng/u1	-0.01
7) Phenol-d5	6.846	99	625707	19.842	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.011	67	341923	18.537	ng/u1	-0.02
11) 2-Chlorophenol-d4	7.205	132	461501	19.180	ng/u1	-0.01
15) 4-Methylphenol-d8	8.381	113	507561	19.745	ng/u1	0.00
21) Nitrobenzene-d5	8.828	128	251367	18.702	ng/u1	-0.01
24) 2-Nitrophenol-d4	9.546	143	270724	19.499	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.081	165	479793	18.510	ng/u1	0.00
31) 4-Chloroaniline-d4	10.599	131	735940	19.123	ng/u1	-0.01
46) Dimethylphthalate-d6	13.734	166	1461924	18.005	ng/u1	-0.01
49) Acenaphthylene-d8	14.010	160	1729057	18.520	ng/u1	0.00
54) 4-Nitrophenol-d4	14.528	143	325533	20.025	ng/u1	0.00
60) Fluorene-d10	15.310	176	1269666	18.482	ng/u1	-0.01
65) 4,6-Dinitro-2-methylph...	15.440	200	244524	19.604	ng/u1	0.00
73) Anthracene-d10	17.163	188	1981928	18.656	ng/u1	0.00
81) Pyrene-d10	19.463	212	2385737	17.916	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.039	264	2491802	18.062	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.217	88	60581	6.368	ng/uL	99
5) Pyridine	3.605	79	457981	18.300	ng/u1	98
6) Benzaldehyde	6.822	77	254093	17.179	ng/u1	99
8) Phenol	6.869	94	639996	19.942	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.105	93	502434	18.933	ng/u1	95
12) 2-Chlorophenol	7.240	128	482603	19.166	ng/u1	99
13) 2-Methylphenol	8.116	108	483116	19.490	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.199	45	555260	18.177	ng/u1	98
16) Acetophenone	8.493	105	789707	20.271	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.481	70	389320	19.286	ng/u1	97
18) 4-Methylphenol	8.440	108	537036	19.987	ng/u1	99
19) Hexachloroethane	8.740	117	192507	18.054	ng/u1	95
22) Nitrobenzene	8.869	77	586078	18.302	ng/u1	98
23) Isophorone	9.393	82	1168764	17.792	ng/u1	100
25) 2-Nitrophenol	9.575	139	291138	19.193	ng/u1	98
26) 2,4-Dimethylphenol	9.640	107	526714	17.849	ng/u1	98
27) Bis(2-Chloroethoxy)met...	9.875	93	715184	17.469	ng/u1	99
29) 2,4-Dichlorophenol	10.105	162	482926	18.869	ng/u1	97
30) Naphthalene	10.505	128	1674283	18.403	ng/u1	99
32) 4-Chloroaniline	10.622	127	715336	19.145	ng/u1	97
33) Hexachlorobutadiene	10.781	225	280027	18.217	ng/u1	93
34) Caprolactam	11.399	113	155969	15.613	ng/u1	89
35) 4-Chloro-3-methylphenol	11.752	107	568873	19.016	ng/u1	98
36) 2-Methylnaphthalene	12.122	142	1154579	18.566	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060524\
 Data File : BN031811.D
 Acq On : 05 Jun 2024 22:23
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 17 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :
 SSTD020275


Quant Time: Jun 06 00:02:13 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN052924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed May 29 15:13:07 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.346	142	1149880	18.429	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.493	216	571997	18.558	ng/ul	99
40) Hexachlorocyclopentadiene	12.469	237	440132	24.776	ng/ul	94
41) 2,4,6-Trichlorophenol	12.740	196	397443	18.998	ng/ul	97
42) 2,4,5-Trichlorophenol	12.810	196	434071	19.093	ng/ul	95
43) 1,1'-Biphenyl	13.146	154	1509511	18.414	ng/ul	98
44) 2-Chloronaphthalene	13.187	162	1183844	18.438	ng/ul	99
45) 2-Nitroaniline	13.398	65	349901	18.800	ng/ul	91
47) Dimethylphthalate	13.781	163	1513619	18.228	ng/ul	100
48) 2,6-Dinitrotoluene	13.904	165	316245	19.160	ng/ul	99
50) Acenaphthylene	14.040	152	1976959	18.737	ng/ul	99
51) 3-Nitroaniline	14.228	138	328852	18.635	ng/ul	99
52) Acenaphthene	14.381	153	1294631	18.545	ng/ul	98
53) 2,4-Dinitrophenol	14.440	184	232148	27.860	ng/ul	97
55) 4-Nitrophenol	14.540	109	289423	24.000	ng/ul	96
56) Dibenzofuran	14.716	168	1762429	18.581	ng/ul	97
57) 2,4-Dinitrotoluene	14.693	165	476217	19.663	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	14.945	232	364463	19.166	ng/ul	99
59) Diethylphthalate	15.151	149	1543120	18.009	ng/ul	100
61) Fluorene	15.369	166	1438729	18.837	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.369	204	696934	18.839	ng/ul	99
63) 4-Nitroaniline	15.393	138	336907	20.054	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.457	198	266532	20.101	ng/ul	98
67) N-Nitrosodiphenylamine	15.581	169	1254129	18.336	ng/ul	99
68) 4-Bromophenyl-phenylether	16.257	248	434608	18.076	ng/ul	97
69) Hexachlorobenzene	16.369	284	503748	18.486	ng/ul	97
70) Atrazine	16.540	200	383178	17.743	ng/ul	99
71) Pentachlorophenol	16.716	266	334027	19.159	ng/ul	92
72) Phenanthrene	17.104	178	2338392	18.499	ng/ul	98
74) Anthracene	17.198	178	2444572	19.085	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.104	216	572793	17.934	ng/ul	97
76) Pentachlorobenzene	14.634	250	574170	18.456	ng/ul	96
77) Carbazole	17.469	167	2263345	20.723	ng/ul	98
78) Di-n-butylphthalate	18.039	149	2740452	18.238	ng/ul	100
80) Fluoranthene	19.128	202	2828189	17.772	ng/ul	99
82) Pyrene	19.492	202	2947753	18.054	ng/ul	98
83) Butylbenzylphthalate	20.398	149	1342473	17.408	ng/ul	94
84) 3,3'-Dichlorobenzidine	21.210	252	951372	19.937	ng/ul	94
85) Benzo(a)anthracene	21.280	228	3094648	18.580	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.210	149	1894917	16.856	ng/ul	97
87) Chrysene	21.345	228	2867796	18.433	ng/ul	98
89) Di-n-octyl phthalate	22.322	149	3495244	18.071	ng/ul	100
90) Benzo(b)fluoranthene	23.310	252	3188182	18.462	ng/ul	100
91) Benzo(k)fluoranthene	23.369	252	3021376	17.830	ng/ul	99
93) Benzo(a)pyrene	24.098	252	2865710	17.718	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	27.539	276	1653133	8.870	ng/ul	97
95) Dibenzo(a,h)anthracene	27.598	278	2690224	17.680	ng/ul	97
96) Benzo(g,h,i)perylene	28.568	276	2441191	16.477	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jun 06 00:02:13 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN052924.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed May 29 15:13:07 2024
Response via : Initial Calibration





#2

1,4-Dioxane

Concen: 6.368 ng/uL

RT: 3.217 min Scan# 39

Delta R.T. -0.012 min

Lab File: BN031811.D

Acq: 05 Jun 2024 22:23

Instrument :

BNA_N

ClientSampleId :

SSTD020275





Tgt Ion: 88 Resp: 60581

Ion Ratio Lower Upper

88 100

43 23.7 18.5 27.7

58 64.6 52.6 78.8



#5

Pyridine

Concen: 18.300 ng/uL

RT: 3.605 min Scan# 105

Delta R.T. -0.012 min

Lab File: BN031811.D

Acq: 05 Jun 2024 22:23




Tgt Ion: 79 Resp: 457981


Ion Ratio Lower Upper

79 100

52 64.2 53.0 79.6

51 28.9 23.3 34.9









#6
Benzaldehyde
Concen: 17.179 ng/ul
RT: 6.822 min Scan# 611
Delta R.T. -0.012 min
Lab File: BN031811.D
Acq: 05 Jun 2024 22:23




Instrument :
BNA_N
ClientSampleId :
SSTD020275


Tgt Ion: 77 Resp: 254093
Ion Ratio Lower Upper
77 100
105 102.1 80.2 120.4
106 95.7 76.8 115.2



#8
Phenol
Concen: 19.942 ng/ul
RT: 6.869 min Scan# 660
Delta R.T. -0.012 min
Lab File: BN031811.D
Acq: 05 Jun 2024 22:23

Tgt Ion: 94 Resp: 639996
Ion Ratio Lower Upper
94 100
65 25.1 20.3 30.5
66 32.6 25.8 38.8









#10
 Bis(2-Chloroethyl)ether
 Concen: 18.933 ng/ul
 RT: 7.105 min Scan# 700
 Delta R.T. -0.012 min
 Lab File: BN031811.D
 Acq: 05 Jun 2024 22:23




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020275


Tgt Ion: 93 Resp: 502434
 Ion Ratio Lower Upper
 93 100
 63 66.0 56.8 85.2
 95 31.7 26.0 39.0



#12
 2-Chlorophenol
 Concen: 19.166 ng/ul
 RT: 7.240 min Scan# 723
 Delta R.T. -0.012 min
 Lab File: BN031811.D
 Acq: 05 Jun 2024 22:23


Tgt Ion:128 Resp: 482603
 Ion Ratio Lower Upper
 128 100
 64 41.4 33.9 50.9
 130 33.5 26.3 39.5








#13
2-Methylphenol
Concen: 19.490 ng/ul
RT: 8.116 min Scan# 81
Delta R.T. -0.006 min
Lab File: BN031811.D
Acq: 05 Jun 2024 22:23

Instrument :
BNA_N
ClientSampleId :
SSTD020275







Tgt Ion:108 Resp: 483116
Ion Ratio Lower Upper
108 100
107 91.0 71.6 107.4



#14
2,2'-oxybis(1-Chloropropane)
Concen: 18.177 ng/ul
RT: 8.199 min Scan# 886
Delta R.T. -0.018 min
Lab File: BN031811.D
Acq: 05 Jun 2024 22:23

Tgt Ion: 45 Resp: 555260
Ion Ratio Lower Upper
45 100
77 18.0 13.6 20.4
79 13.8 10.6 15.8









#16
 Acetophenone
 Concen: 20.271 ng/ul
 RT: 8.493 min Scan# 91
 Delta R.T. -0.012 min
 Lab File: BN031811.D
 Acq: 05 Jun 2024 22:23




Instrument :
 BNA_N
 ClientSampleId :
 SST020275


Tgt Ion:105 Resp: 789707
 Ion Ratio Lower Upper
 105 100
 77 73.2 58.6 87.8
 51 21.4 17.2 25.8



#17
 N-Nitroso-di-n-propylamine
 Concen: 19.286 ng/ul
 RT: 8.481 min Scan# 934
 Delta R.T. -0.012 min
 Lab File: BN031811.D
 Acq: 05 Jun 2024 22:23

Tgt Ion: 70 Resp: 389320
 Ion Ratio Lower Upper
 70 100
 42 45.0 37.3 55.9
 101 10.7 7.8 11.6
 130 23.2 16.4 24.6









#18
4-Methylphenol
Concen: 19.987 ng/ul
RT: 8.440 min Scan# 911
Delta R.T. -0.012 min
Lab File: BN031811.D
Acq: 05 Jun 2024 22:23




Instrument :
BNA_N
ClientSampleId :
SSTD020275


Tgt Ion:108 Resp: 537036
Ion Ratio Lower Upper
108 100
107 111.1 89.6 134.4



#19
Hexachloroethane
Concen: 18.054 ng/ul
RT: 8.740 min Scan# 978
Delta R.T. -0.018 min
Lab File: BN031811.D
Acq: 05 Jun 2024 22:23

Tgt Ion:117 Resp: 192507
Ion Ratio Lower Upper
117 100
201 94.1 71.6 107.4
199 60.6 45.6 68.4









#22
Nitrobenzene
Concen: 18.302 ng/ul
RT: 8.869 min Scan# 1000
Delta R.T. -0.012 min
Lab File: BN031811.D
Acq: 05 Jun 2024 22:23




Instrument :
BNA_N
ClientSampleId :
SSTD020275


Tgt Ion: 77 Resp: 586078
Ion Ratio Lower Upper
77 100
123 47.6 36.6 55.0
65 12.5 10.2 15.4



#23
Isophorone
Concen: 17.792 ng/ul
RT: 9.393 min Scan# 1089
Delta R.T. -0.012 min
Lab File: BN031811.D
Acq: 05 Jun 2024 22:23

Tgt Ion: 82 Resp: 1168764
Ion Ratio Lower Upper
82 100
95 6.8 5.5 8.3
138 17.7 14.1 21.1









#25
2-Nitrophenol
Concen: 19.193 ng/ul
RT: 9.575 min Scan# 1121
Delta R.T. -0.012 min
Lab File: BN031811.D
Acq: 05 Jun 2024 22:23




Instrument :
BNA_N
ClientSampleId :
SSTD020275


Tgt Ion:139 Resp: 291138
Ion Ratio Lower Upper
139 100
65 46.3 38.3 57.5
109 28.7 22.1 33.1



#26
2,4-Dimethylphenol
Concen: 17.849 ng/ul
RT: 9.640 min Scan# 1131
Delta R.T. -0.006 min
Lab File: BN031811.D
Acq: 05 Jun 2024 22:23

Tgt Ion:107 Resp: 526714
Ion Ratio Lower Upper
107 100
121 49.7 40.8 61.2
122 86.4 70.4 105.6





#27

Bis(2-Chloroethoxy)methane

Concen: 17.469 ng/ul

RT: 9.875 min Scan# 1171

Delta R.T. -0.018 min

Lab File: BN031811.D

Acq: 05 Jun 2024 22:23

Instrument :

BNA_N

ClientSampleId :

SSTD020275





Tgt Ion: 93 Resp: 715184

Ion Ratio Lower Upper

93 100

95 32.2 26.1 39.1

123 11.5 9.0 13.6



#29

2,4-Dichlorophenol

Concen: 18.869 ng/ul

RT: 10.105 min Scan# 1210

Delta R.T. -0.012 min

Lab File: BN031811.D

Acq: 05 Jun 2024 22:23




Tgt Ion:162 Resp: 482926


Ion Ratio Lower Upper

162 100

164 63.5 52.7 79.1

98 37.0 30.5 45.7









#30
Naphthalene
Concen: 18.403 ng/ul
RT: 10.505 min Scan# 1278
Delta R.T. -0.012 min
Lab File: BN031811.D
Acq: 05 Jun 2024 22:23




Instrument :
BNA_N
ClientSampleId :
SSTD020275


Tgt Ion:128 Resp: 1674283
Ion Ratio Lower Upper
128 100
129 10.9 8.5 12.7
127 12.5 10.3 15.5



#32
4-Chloroaniline
Concen: 19.145 ng/ul
RT: 10.622 min Scan# 1298
Delta R.T. -0.006 min
Lab File: BN031811.D
Acq: 05 Jun 2024 22:23

Tgt Ion:127 Resp: 715336
Ion Ratio Lower Upper
127 100
129 32.7 25.0 37.4









#33
Hexachlorobutadiene
Concen: 18.217 ng/ul
RT: 10.781 min Scan# 11
Delta R.T. -0.018 min
Lab File: BN031811.D
Acq: 05 Jun 2024 22:23




Instrument :
BNA_N
ClientSampleId :
SSTD020275


Tgt Ion:225 Resp: 280027
Ion Ratio Lower Upper
225 100
223 61.0 52.3 78.5
227 59.5 52.5 78.7



#34
Caprolactam
Concen: 15.613 ng/ul
RT: 11.399 min Scan# 1430
Delta R.T. -0.000 min
Lab File: BN031811.D
Acq: 05 Jun 2024 22:23

Tgt Ion:113 Resp: 155969
Ion Ratio Lower Upper
113 100
55 133.0 117.2 175.8
56 100.8 89.4 134.0









#35
 4-Chloro-3-methylphenol
 Concen: 19.016 ng/ul
 RT: 11.752 min Scan# 1490
 Delta R.T. 0.000 min
 Lab File: BN031811.D
 Acq: 05 Jun 2024 22:23

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020275

Tgt Ion:107 Resp: 568873




Ion	Ratio	Lower	Upper
107	100		
144	25.1	20.5	30.7
142	78.2	64.0	96.0




#36
 2-Methylnaphthalene
 Concen: 18.566 ng/ul
 RT: 12.122 min Scan# 1553
 Delta R.T. -0.012 min
 Lab File: BN031811.D
 Acq: 05 Jun 2024 22:23

Tgt Ion:142 Resp: 1154579

Ion	Ratio	Lower	Upper
142	100		
141	88.3	70.1	105.1





#37

1-Methylnaphthalene

Concen: 18.429 ng/ul

RT: 12.346 min Scan# 1591

Delta R.T. -0.012 min

Lab File: BN031811.D

Acq: 05 Jun 2024 22:23

Instrument :

BNA_N

ClientSampleId :

SSTD020275





Tgt Ion:142 Resp: 1149880

Ion Ratio Lower Upper

142 100

141 91.9 71.7 107.5

116 3.7 3.0 4.4



#39

1,2,4,5-Tetrachlorobenzene

Concen: 18.558 ng/ul

RT: 12.493 min Scan# 1616

Delta R.T. -0.012 min

Lab File: BN031811.D

Acq: 05 Jun 2024 22:23

Tgt Ion:216 Resp: 571997




Ion Ratio Lower Upper


216 100

214 77.5 61.7 92.5

179 20.1 16.7 25.1

108 17.1 14.5 21.7









#40
Hexachlorocyclopentadiene
Concen: 24.776 ng/ul
RT: 12.469 min Scan# 1612
Delta R.T. -0.012 min
Lab File: BN031811.D
Acq: 05 Jun 2024 22:23




Instrument :
BNA_N
ClientSampleId :
SSTD020275


Tgt Ion:237 Resp: 440132
Ion Ratio Lower Upper
237 100
235 65.2 47.8 71.6
272 13.2 10.0 15.0



#41
2,4,6-Trichlorophenol
Concen: 18.998 ng/ul
RT: 12.740 min Scan# 1658
Delta R.T. -0.006 min
Lab File: BN031811.D
Acq: 05 Jun 2024 22:23

Tgt Ion:196 Resp: 397443
Ion Ratio Lower Upper
196 100
198 91.7 76.4 114.6
200 31.2 24.2 36.4









#42
2,4,5-Trichlorophenol
Concen: 19.093 ng/ul
RT: 12.810 min Scan# 1670
Delta R.T. -0.006 min
Lab File: BN031811.D
Acq: 05 Jun 2024 22:23




Instrument :
BNA_N
ClientSampleId :
SSTD020275


Tgt Ion:196 Resp: 434071
Ion Ratio Lower Upper
196 100
198 97.2 73.8 110.6
200 29.8 25.6 38.4



#43
1,1'-Biphenyl
Concen: 18.414 ng/ul
RT: 13.146 min Scan# 1727
Delta R.T. -0.012 min
Lab File: BN031811.D
Acq: 05 Jun 2024 22:23

Tgt Ion:154 Resp: 1509511
Ion Ratio Lower Upper
154 100
153 39.6 30.6 46.0
76 13.9 11.1 16.7





#44

2-Chloronaphthalene

Concen: 18.438 ng/ul

RT: 13.187 min Scan# 1

Delta R.T. -0.012 min

Lab File: BN031811.D

Acq: 05 Jun 2024 22:23

Instrument :

BNA_N

ClientSampleId :

SSTD020275





Tgt Ion:162 Resp: 1183844

Ion Ratio Lower Upper

162 100

164 33.3 27.0 40.4

127 38.5 31.8 47.6



#45

2-Nitroaniline

Concen: 18.800 ng/ul

RT: 13.398 min Scan# 1770

Delta R.T. -0.006 min

Lab File: BN031811.D

Acq: 05 Jun 2024 22:23




Tgt Ion: 65 Resp: 349901


Ion Ratio Lower Upper

65 100

92 77.9 58.8 88.2

138 120.5 86.2 129.4





#47

Dimethylphthalate

Concen: 18.228 ng/ul

RT: 13.781 min Scan# 1835

Delta R.T. -0.012 min

Lab File: BN031811.D

Acq: 05 Jun 2024 22:23

Instrument :

BNA_N

ClientSampleId :

SSTD020275





Tgt Ion:163 Resp: 1513619

Ion Ratio Lower Upper

163 100

194 4.1 3.2 4.8

164 10.1 8.2 12.2



#48

2,6-Dinitrotoluene

Concen: 19.160 ng/ul

RT: 13.904 min Scan# 1856

Delta R.T. -0.006 min

Lab File: BN031811.D

Acq: 05 Jun 2024 22:23




Tgt Ion:165 Resp: 316245


Ion Ratio Lower Upper

165 100

89 56.8 44.6 67.0

121 22.0 17.3 25.9





#50

Acenaphthylene

Concen: 18.737 ng/ul

RT: 14.040 min Scan# 1879

Delta R.T. -0.006 min

Lab File: BN031811.D

Acq: 05 Jun 2024 22:23

Instrument :

BNA_N

ClientSampleId :

SSTD020275





Tgt Ion:152 Resp: 1976959

Ion Ratio Lower Upper

152 100

151 19.5 15.9 23.9

153 13.3 10.4 15.6



#51

3-Nitroaniline

Concen: 18.635 ng/ul

RT: 14.228 min Scan# 1911

Delta R.T. -0.006 min

Lab File: BN031811.D

Acq: 05 Jun 2024 22:23




Tgt Ion:138 Resp: 328852


Ion Ratio Lower Upper

138 100

108 11.3 9.2 13.8

92 118.6 94.0 141.0





#52

Acenaphthene

Concen: 18.545 ng/ul

RT: 14.381 min Scan# 1937

Delta R.T. -0.006 min

Lab File: BN031811.D

Acq: 05 Jun 2024 22:23

Instrument :

BNA_N

ClientSampleId :

SSTD020275





Tgt Ion:153 Resp: 1294631

Ion Ratio Lower Upper

153 100

152 46.0 37.8 56.6

154 88.8 69.8 104.8



#53

2,4-Dinitrophenol

Concen: 27.860 ng/ul

RT: 14.440 min Scan# 1947

Delta R.T. -0.000 min

Lab File: BN031811.D

Acq: 05 Jun 2024 22:23




Tgt Ion:184 Resp: 232148


Ion Ratio Lower Upper

184 100

63 61.6 50.2 75.4

154 64.3 49.1 73.7









#55
4-Nitrophenol
Concen: 24.000 ng/ul
RT: 14.540 min Scan# 1965
Delta R.T. 0.000 min
Lab File: BN031811.D
Acq: 05 Jun 2024 22:23




Instrument :
BNA_N
ClientSampleId :
SSTD020275


Tgt Ion:109 Resp: 289423
Ion Ratio Lower Upper
109 100
139 143.6 113.4 170.2
65 125.1 106.2 159.2



#56
Dibenzofuran
Concen: 18.581 ng/ul
RT: 14.716 min Scan# 1994
Delta R.T. -0.012 min
Lab File: BN031811.D
Acq: 05 Jun 2024 22:23

Tgt Ion:168 Resp: 1762429
Ion Ratio Lower Upper
168 100
139 38.6 29.4 44.0









#57
 2,4-Dinitrotoluene
 Concen: 19.663 ng/ul
 RT: 14.693 min Scan# 1990
 Delta R.T. 0.000 min
 Lab File: BN031811.D
 Acq: 05 Jun 2024 22:23




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020275


Tgt Ion:165 Resp: 476217
 Ion Ratio Lower Upper
 165 100
 63 39.3 32.2 48.2
 182 3.0 2.3 3.5



#58
 2,3,4,6-Tetrachlorophenol
 Concen: 19.166 ng/ul
 RT: 14.945 min Scan# 2033
 Delta R.T. -0.006 min
 Lab File: BN031811.D
 Acq: 05 Jun 2024 22:23

Tgt Ion:232 Resp: 364463
 Ion Ratio Lower Upper
 232 100
 131 46.5 36.6 55.0
 130 2.2 1.8 2.8
 166 30.7 25.0 37.4





#59

Diethylphthalate

Concen: 18.009 ng/ul

RT: 15.151 min Scan# 2068

Delta R.T. -0.006 min

Lab File: BN031811.D

Acq: 05 Jun 2024 22:23

Instrument :

BNA_N

ClientSampleId :

SSTD020275





Tgt Ion:149 Resp: 1543120

Ion Ratio Lower Upper

149 100

177 21.5 17.1 25.7

150 12.2 10.0 15.0



#61

Fluorene

Concen: 18.837 ng/ul

RT: 15.369 min Scan# 2105

Delta R.T. -0.006 min

Lab File: BN031811.D

Acq: 05 Jun 2024 22:23




Tgt Ion:166 Resp: 1438729


Ion Ratio Lower Upper

166 100

165 99.0 80.3 120.5

167 13.3 11.3 16.9









#62
4-Chlorophenyl-phenylether
Concen: 18.839 ng/ul
RT: 15.369 min Scan# 2105
Delta R.T. -0.006 min
Lab File: BN031811.D
Acq: 05 Jun 2024 22:23




Instrument :
BNA_N
ClientSampleId :
SSTD020275


Tgt Ion:204 Resp: 696934
Ion Ratio Lower Upper
204 100
206 32.0 25.7 38.5
141 59.1 47.7 71.5



#63
4-Nitroaniline
Concen: 20.054 ng/ul
RT: 15.393 min Scan# 2109
Delta R.T. -0.006 min
Lab File: BN031811.D
Acq: 05 Jun 2024 22:23

Tgt Ion:138 Resp: 336907
Ion Ratio Lower Upper
138 100
92 50.8 41.0 61.4
108 74.0 60.1 90.1









#66
4,6-Dinitro-2-methylphenol
Concen: 20.101 ng/ul
RT: 15.457 min Scan# 2120
Delta R.T. 0.000 min
Lab File: BN031811.D
Acq: 05 Jun 2024 22:23




Instrument :
BNA_N
ClientSampleId :
SSTD020275


Tgt Ion:198 Resp: 266532
Ion Ratio Lower Upper
198 100
182 3.9 3.0 4.6
77 23.4 19.4 29.2



#67
N-Nitrosodiphenylamine
Concen: 18.336 ng/ul
RT: 15.581 min Scan# 2141
Delta R.T. -0.006 min
Lab File: BN031811.D
Acq: 05 Jun 2024 22:23

Tgt Ion:169 Resp: 1254129
Ion Ratio Lower Upper
169 100
168 66.9 52.7 79.1
167 36.6 29.6 44.4









#68
4-Bromophenyl-phenylether
Concen: 18.076 ng/ul
RT: 16.257 min Scan# 21
Delta R.T. -0.012 min
Lab File: BN031811.D
Acq: 05 Jun 2024 22:23




Instrument :
BNA_N
ClientSampleId :
SSTD020275


Tgt Ion:248 Resp: 434608
Ion Ratio Lower Upper
248 100
250 99.5 76.5 114.7
141 81.5 63.9 95.9



#69
Hexachlorobenzene
Concen: 18.486 ng/ul
RT: 16.369 min Scan# 2275
Delta R.T. 0.000 min
Lab File: BN031811.D
Acq: 05 Jun 2024 22:23

Tgt Ion:284 Resp: 503748
Ion Ratio Lower Upper
284 100
142 35.7 31.1 46.7
249 31.9 25.7 38.5





#70

Atrazine

Concen: 17.743 ng/ul

RT: 16.540 min Scan# 21

Delta R.T. -0.006 min

Lab File: BN031811.D

Acq: 05 Jun 2024 22:23

Instrument :

BNA_N

ClientSampleId :

SSTD020275





Tgt Ion:200 Resp: 383178

Ion Ratio Lower Upper

200 100

173 26.8 21.2 31.8

215 48.2 38.9 58.3



#71

Pentachlorophenol

Concen: 19.159 ng/ul

RT: 16.716 min Scan# 2334

Delta R.T. -0.000 min

Lab File: BN031811.D

Acq: 05 Jun 2024 22:23




Tgt Ion:266 Resp: 334027


Ion Ratio Lower Upper

266 100

264 59.3 54.7 82.1

268 62.2 53.4 80.2









#72
Phenanthrene
Concen: 18.499 ng/ul
RT: 17.104 min Scan# 2400
Delta R.T. -0.012 min
Lab File: BN031811.D
Acq: 05 Jun 2024 22:23




Instrument :
BNA_N
ClientSampleId :
SSTD020275


Tgt Ion:178 Resp: 2338392
Ion Ratio Lower Upper
178 100
179 14.8 12.6 19.0
176 19.0 15.8 23.6



#74
Anthracene
Concen: 19.085 ng/ul
RT: 17.198 min Scan# 2416
Delta R.T. -0.006 min
Lab File: BN031811.D
Acq: 05 Jun 2024 22:23

Tgt Ion:178 Resp: 2444572
Ion Ratio Lower Upper
178 100
179 15.3 11.8 17.6
176 18.2 14.8 22.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 17.934 ng/uL
RT: 13.104 min Scan# 1721
Delta R.T. -0.012 min
Lab File: BN031811.D
Acq: 05 Jun 2024 22:23




Instrument :
BNA_N
ClientSampleId :
SSTD020275


Tgt Ion:216 Resp: 572793
Ion Ratio Lower Upper
216 100
214 75.4 62.1 93.1
218 47.9 40.2 60.4



#76
Pentachlorobenzene
Concen: 18.456 ng/uL
RT: 14.634 min Scan# 1980
Delta R.T. -0.006 min
Lab File: BN031811.D
Acq: 05 Jun 2024 22:23

Tgt Ion:250 Resp: 574170
Ion Ratio Lower Upper
250 100
248 60.0 49.8 74.6
252 61.9 52.3 78.5









#77
 Carbazole
 Concen: 20.723 ng/ul
 RT: 17.469 min Scan# 2462
 Delta R.T. -0.006 min
 Lab File: BN031811.D
 Acq: 05 Jun 2024 22:23

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020275

Tgt Ion:167 Resp: 2263345




Ion	Ratio	Lower	Upper
167	100		
166	20.7	17.9	26.9
139	13.0	10.5	15.7




#78
 Di-n-butylphthalate
 Concen: 18.238 ng/ul
 RT: 18.039 min Scan# 2559
 Delta R.T. -0.006 min
 Lab File: BN031811.D
 Acq: 05 Jun 2024 22:23

Tgt Ion:149 Resp: 2740452

Ion	Ratio	Lower	Upper
149	100		
150	9.1	7.4	11.0
104	4.9	4.1	6.1









#80
Fluoranthene
Concen: 17.772 ng/ul
RT: 19.128 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN031811.D
Acq: 05 Jun 2024 22:23




Instrument :
BNA_N
ClientSampleId :
SSTD020275


Tgt Ion:202 Resp: 2828189
Ion Ratio Lower Upper
202 100
101 12.7 10.3 15.5
100 9.7 8.0 12.0



#82
Pyrene
Concen: 18.054 ng/ul
RT: 19.492 min Scan# 2806
Delta R.T. 0.000 min
Lab File: BN031811.D
Acq: 05 Jun 2024 22:23

Tgt Ion:202 Resp: 2947753
Ion Ratio Lower Upper
202 100
101 14.7 12.6 18.8
100 12.1 10.2 15.2





#83

Butylbenzylphthalate

Concen: 17.408 ng/ul

RT: 20.398 min Scan# 2960

Delta R.T. -0.006 min

Lab File: BN031811.D

Acq: 05 Jun 2024 22:23

Instrument :

BNA_N

ClientSampleId :

SSTD020275





Tgt Ion:149 Resp: 1342473

Ion Ratio Lower Upper

149 100

91 70.0 61.0 91.4

206 21.5 16.5 24.7



#84

3,3'-Dichlorobenzidine

Concen: 19.937 ng/ul

RT: 21.210 min Scan# 3098

Delta R.T. -0.006 min

Lab File: BN031811.D

Acq: 05 Jun 2024 22:23




Tgt Ion:252 Resp: 951372


Ion Ratio Lower Upper

252 100

254 61.4 53.6 80.4

126 12.7 10.2 15.4





#85

Benzo(a)anthracene

Concen: 18.580 ng/ul

RT: 21.280 min Scan# 3111

Delta R.T. -0.006 min

Lab File: BN031811.D

Acq: 05 Jun 2024 22:23

Instrument :

BNA_N

ClientSampleId :

SSTD020275





Tgt Ion:228 Resp: 3094648

Ion Ratio Lower Upper

228 100

229 18.9 16.7 25.1

226 27.4 22.8 34.2



#86

Bis(2-ethylhexyl)phthalate

Concen: 16.856 ng/ul

RT: 21.210 min Scan# 3098

Delta R.T. -0.012 min

Lab File: BN031811.D

Acq: 05 Jun 2024 22:23




Tgt Ion:149 Resp: 1894917


Ion Ratio Lower Upper

149 100

167 28.5 21.4 32.0

279 4.3 3.3 4.9





#87

Chrysene

Concen: 18.433 ng/ul

RT: 21.345 min Scan# 3121

Delta R.T. -0.006 min

Lab File: BN031811.D

Acq: 05 Jun 2024 22:23

Instrument :

BNA_N

ClientSampleId :

SSTD020275





Tgt Ion:228 Resp: 2867796

Ion Ratio Lower Upper

228 100

226 30.9 23.7 35.5

229 18.6 15.1 22.7



#89

Di-n-octyl phthalate

Concen: 18.071 ng/ul




RT: 22.322 min Scan# 3287


Delta R.T. -0.018 min

Lab File: BN031811.D

Acq: 05 Jun 2024 22:23

Tgt Ion:149 Resp: 3495244





#90

Benzo(b)fluoranthene

Concen: 18.462 ng/ul

RT: 23.310 min Scan# 3456

Delta R.T. -0.000 min

Lab File: BN031811.D

Acq: 05 Jun 2024 22:23

Instrument :

BNA_N

ClientSampleId :

SSTD020275





Tgt Ion:252 Resp: 3188182

Ion Ratio Lower Upper

252 100

253 21.2 17.0 25.6

125 11.3 9.2 13.8



#91

Benzo(k)fluoranthene

Concen: 17.830 ng/ul

RT: 23.369 min Scan# 3465

Delta R.T. -0.006 min

Lab File: BN031811.D

Acq: 05 Jun 2024 22:23




Tgt Ion:252 Resp: 3021376


Ion Ratio Lower Upper

252 100

253 21.1 17.6 26.4

125 12.1 9.4 14.2





#93

Benzo(a)pyrene

Concen: 17.718 ng/ul

RT: 24.098 min Scan# 31

Delta R.T. -0.006 min

Lab File: BN031811.D

Acq: 05 Jun 2024 22:23

Instrument :

BNA_N

ClientSampleId :

SSTD020275





Tgt Ion:252 Resp: 2865710

Ion Ratio Lower Upper

252 100

253 22.2 16.6 24.8

125 12.9 9.8 14.8



#94

Indeno(1,2,3-cd)pyrene

Concen: 8.870 ng/ul

RT: 27.539 min Scan# 4174

Delta R.T. 0.006 min

Lab File: BN031811.D

Acq: 05 Jun 2024 22:23




Tgt Ion:276 Resp: 1653133


Ion Ratio Lower Upper

276 100

138 24.8 18.1 27.1

277 23.4 18.2 27.2









#95
 Dibenzo(a,h)anthracene
 Concen: 17.680 ng/ul
 RT: 27.598 min Scan# 4184
 Delta R.T. -0.006 min
 Lab File: BN031811.D
 Acq: 05 Jun 2024 22:23

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020275

Tgt Ion:278 Resp: 2690224
 Ion Ratio Lower Upper
 278 100
 139 18.2 15.0 22.6
 279 24.1 17.4 26.0



#96
 Benzo(g,h,i)perylene
 Concen: 16.477 ng/ul
 RT: 28.568 min Scan# 4349
 Delta R.T. -0.000 min
 Lab File: BN031811.D
 Acq: 05 Jun 2024 22:23

Tgt Ion:276 Resp: 2441191
 Ion Ratio Lower Upper
 276 100
 138 24.7 20.9 31.3
 277 22.0 18.0 27.0

