

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN060619\
 Data File : BN006140.D
 Acq On : 06 Jun 2019 17:42
 Operator : JU/SJ
 Sample : K3202-01
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0AD2

Manual Integrations
 APPROVED

mohammad
 6/11/2019 5:07:13 PM

Quant Time: Jun 07 00:36:56 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN053019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 07 00:25:28 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.65	152	11892	0.40	ng/ul	0.00
2) Naphthalene-d8	10.43	136	45811	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.30	164	22661	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.05	188	45780	0.40	ng/ul	0.00
16) Chrysene-d12	21.27	240	33366	0.40	ng/ul	-0.03
20) Perylene-d12	23.50	264	41589	0.40	ng/ul	-0.06
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.03	152	2572	0.04	ng/ul	0.00
14) Fluoranthene-d10	19.09	212	5291	0.04	ng/ul	0.00
Target Compounds					Ovalue	
3) Naphthalene	10.48	128	75825	0.600	ng/ul	99
5) 2-Methylnaphthalene	12.10	142	56368	0.640	ng/ul	99
7) Acenaphthylene	14.02	152	90213	0.716	ng/ul	99
8) Acenaphthene	14.36	153	108198	1.238	ng/ul	99
9) Fluorene	15.36	166	169150	1.655	ng/ul	99
12) Phenanthrene	17.09	178	1982904	13.493	ng/ul	99
13) Anthracene	17.19	178	530305	3.779	ng/ul	99
15) Fluoranthene	19.12	202	2085676	12.853	ng/ul	95
17) Pyrene	19.49	202	1624228	9.242	ng/ul	99
18) Benzo(a)anthracene	21.25	228	989114	6.703	ng/ul	100
19) Chrysene	21.31	228	868767	6.271	ng/ul	97
21) Benzo(b)fluoranthene	22.84	252	951295m	5.486	ng/ul	
22) Benzo(k)fluoranthene	22.87	252	428267m	2.535	ng/ul	
23) Benzo(a)pyrene	23.40	252	640061	3.963	ng/ul#	85
24) Indeno(1,2,3-cd)pyrene	25.73	276	376526	2.154	ng/ul#	90
25) Dibenzo(a,h)anthracene	25.74	278	124146	0.888	ng/ul	96
26) Benzo(a,h,i)perylene	26.41	276	159288	1.088	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN060619\
Data File : BN006140.D
Acq On : 06 Jun 2019 17:42
Operator : JU/SJ
Sample : K3202-01
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0AD2

Manual Integrations
APPROVED

mohammad
6/11/2019 5:07:13 PM

Quant Time: Jun 07 00:36:56 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN053019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 07 00:25:28 2019
Response via : Initial Calibration

