

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN060619\
 Data File : BN006154.D
 Acq On : 07 Jun 2019 02:56
 Operator : JU/SJ
 Sample : K3016-08
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 A51C7

Manual Integrations
 APPROVED

mohammad
 6/11/2019 5:07:53 PM

Quant Time: Jun 07 04:00:16 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN053019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 07 01:31:14 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.65	152	5066	0.40	ng/ul	0.00
2) Naphthalene-d8	10.43	136	19091	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.30	164	9837	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.05	188	20030m	0.40	ng/ul	0.00
16) Chrysene-d12	21.27	240	17444m	0.40	ng/ul	0.00
20) Perylene-d12	23.49	264	15452m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.03	152	7287	0.25	ng/ul	0.00
14) Fluoranthene-d10	19.09	212	15618m	0.29	ng/ul	0.00
Target Compounds				Ovalue		
7) Acenaphthylene	14.01	152	2466	0.045	ng/ul#	84
12) Phenanthrene	17.09	178	1406m	0.022	ng/ul	
15) Fluoranthene	19.12	202	3642m	0.051	ng/ul	
17) Pyrene	19.49	202	3331	0.036	ng/ul#	69
18) Benzo(a)anthracene	21.25	228	2583	0.033	ng/ul#	42
19) Chrysene	21.30	228	2336	0.032	ng/ul#	51

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN060619\
Data File : BN006154.D
Acq On : 07 Jun 2019 02:56
Operator : JU/SJ
Sample : K3016-08
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
A51C7

Manual Integrations
APPROVED

mohammad
6/11/2019 5:07:53 PM

Quant Time: Jun 07 04:00:16 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN053019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 07 01:31:14 2019
Response via : Initial Calibration

