

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060623\
 Data File : BN025758.D
 Acq On : 06 Jun 2023 17:24
 Operator : CG/JU
 Sample : 03021-03
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SP-303-20230601

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/07/2023
 Supervised By :mohammad ahmed 06/08/2023

Quant Time: Jun 07 05:28:57 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN053123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 06 15:34:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.012	152	6892	0.400	ng	0.00
7) Naphthalene-d8	10.835	136	18008	0.400	ng	0.01
13) Acenaphthene-d10	14.651	164	10720	0.400	ng	0.01
19) Phenanthrene-d10	17.393	188	20454	0.400	ng	0.01
29) Chrysene-d12	21.578	240	10006	0.400	ng	# 0.00
35) Perylene-d12	24.057	264	9245	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.564	112	1940	0.100	ng	0.01
5) Phenol-d6	7.218	99	925m	0.039	ng	0.06
8) Nitrobenzene-d5	9.202	82	3965	0.237	ng	0.02
11) 2-Methylnaphthalene-d10	12.423	152	8434	0.325	ng	0.02
14) 2,4,6-Tribromophenol	16.152	330	772	0.234	ng	0.01
15) 2-Fluorobiphenyl	13.283	172	14760	0.326	ng	0.01
27) Fluoranthene-d10	19.421	212	17858	0.424	ng	0.00
31) Terphenyl-d14	20.006	244	14312	0.590	ng	0.00
Target Compounds						
34) Bis(2-ethylhexyl)phtha...	21.471	149	1479	0.064	ng	# 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060623\
Data File : BN025758.D
Acq On : 06 Jun 2023 17:24
Operator : CG/JU
Sample : 03021-03
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SP-303-20230601

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/07/2023
Supervised By :mohammad ahmed 06/08/2023

Quant Time: Jun 07 05:28:57 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN053123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 06 15:34:02 2023
Response via : Initial Calibration

