

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060723\
 Data File : BN025769.D
 Acq On : 07 Jun 2023 11:51
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 08 02:56:28 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN053123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 06 15:34:02 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	116	0.00
2	1,4-Dioxane	0.400	0.400	0.0	106	0.00
3	n-Nitrosodimethylamine	0.400	0.232	42.0#	63	0.00
4 S	2-Fluorophenol	0.400	0.345	13.8	92	0.00
5 S	Phenol-d6	0.400	0.350	12.5	95	0.00
6	bis(2-Chloroethyl)ether	0.400	0.396	1.0	104	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	113	0.00
8 S	Nitrobenzene-d5	0.400	0.359	10.3	99	0.00
9	Naphthalene	0.400	0.394	1.5	105	0.00
10	Hexachlorobutadiene	0.400	0.496	-24.0	130	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.425	-6.2	111	0.00
12	2-Methylnaphthalene	0.400	0.425	-6.2	111	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	112	0.00
14 S	2,4,6-Tribromophenol	0.400	0.327	18.3	92	0.00
15 S	2-Fluorobiphenyl	0.400	0.424	-6.0	114	0.00
16	Acenaphthylene	0.400	0.398	0.5	107	0.00
17	Acenaphthene	0.400	0.421	-5.2	112	0.00
18	Fluorene	0.400	0.421	-5.2	108	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	102	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.327	18.3	83	0.00
21	4-Bromophenyl-phenylether	0.400	0.449	-12.2	109	0.00
22	Hexachlorobenzene	0.400	0.519	-29.8#	126	0.00
23	Atrazine	0.400	0.351	12.3	82	0.00
24	Pentachlorophenol	0.400	0.225	43.8#	57	0.00
25	Phenanthrene	0.400	0.424	-6.0	101	0.00
26	Anthracene	0.400	0.394	1.5	96	0.00
27 SURR	Fluoranthene-d10	0.400	0.402	-0.5	91	0.00
28	Fluoranthene	0.400	0.406	-1.5	92	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	85	0.00
30	Pyrene	0.400	0.455	-13.7	91	0.00
31 S	Terphenyl-d14	0.400	0.418	-4.5	85	0.00
32	Benzo(a)anthracene	0.400	0.385	3.8	79	0.00
33	Chrysene	0.400	0.436	-9.0	85	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.385	3.8	69	0.00
35 I	Perylene-d12	0.400	0.400	0.0	96	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.456	-14.0	110	0.00
37	Benzo(b)fluoranthene	0.400	0.414	-3.5	91	0.00
38	Benzo(k)fluoranthene	0.400	0.421	-5.2	95	0.00
39 C	Benzo(a)pyrene	0.400	0.435	-8.7	99	0.00
40	Dibenzo(a,h)anthracene	0.400	0.444	-11.0	107	0.01
41	Benzo(g,h,i)perylene	0.400	0.497	-24.2	119	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4