

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN060819\
 Data File : BN006190.D
 Acq On : 08 Jun 2019 14:29
 Operator : JU/SJ
 Sample : K3016-16
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 A51D3

Manual Integrations
 APPROVED

mohammad
 6/11/2019 4:30:38 PM

Quant Time: Jun 08 15:49:47 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN053019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 07 01:31:14 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.65	152	5120	0.40	ng/ul	0.00
2) Naphthalene-d8	10.42	136	19800	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.30	164	9954	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.05	188	19058m	0.40	ng/ul	0.00
16) Chrysene-d12	21.27	240	13804m	0.40	ng/ul	0.00
20) Perylene-d12	23.51	264	21325	0.40	ng/ul	0.01
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.02	152	5271	0.17	ng/ul	0.00
14) Fluoranthene-d10	19.09	212	10102m	0.20	ng/ul	0.00
Target Compounds					Ovalue	
3) Naphthalene	10.48	128	5112	0.094	ng/ul#	95
5) 2-Methylnaphthalene	12.10	142	2171	0.057	ng/ul	96
7) Acenaphthylene	14.01	152	62083	1.122	ng/ul	100
8) Acenaphthene	14.36	153	1651	0.043	ng/ul#	94
9) Fluorene	15.35	166	6059	0.135	ng/ul#	95
12) Phenanthrene	17.09	178	145618m	2.380	ng/ul	
13) Anthracene	17.18	178	41562m	0.711	ng/ul	
15) Fluoranthene	19.12	202	280326m	4.150	ng/ul	
17) Pyrene	19.49	202	241939	3.327	ng/ul	98
18) Benzo(a)anthracene	21.26	228	78675	1.289	ng/ul	94
19) Chrysene	21.31	228	133076	2.322	ng/ul	99
21) Benzo(b)fluoranthene	22.85	252	179212m	2.015	ng/ul	
22) Benzo(k)fluoranthene	22.89	252	78833m	0.910	ng/ul	
23) Benzo(a)pyrene	23.42	252	107765	1.301	ng/ul#	86
24) Indeno(1,2,3-cd)pyrene	25.75	276	92694	1.034	ng/ul#	87
25) Dibenzo(a,h)anthracene	25.75	278	21000	0.293	ng/ul	100
26) Benzo(a,h,i)perylene	26.43	276	78526	1.046	ng/ul	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN060819\
Data File : BN006190.D
Acq On : 08 Jun 2019 14:29
Operator : JU/SJ
Sample : K3016-16
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
A51D3

Manual Integrations
APPROVED

mohammad
6/11/2019 4:30:38 PM

Quant Time: Jun 08 15:49:47 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN053019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 07 01:31:14 2019
Response via : Initial Calibration

