

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060824\
 Data File : BN031870.D
 Acq On : 08 Jun 2024 23:04
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020282

Quant Time: Jun 10 01:49:13 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN052924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed May 29 15:13:07 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.669	152	395912	20.000	ng/u1	-0.02
20) Naphthalene-d8	10.451	136	1761328	20.000	ng/u1	-0.02
38) Acenaphthene-d10	14.316	164	1096860	20.000	ng/u1	-0.01
64) Phenanthrene-d10	17.063	188	2113359	20.000	ng/u1	0.00
79) Chrysene-d12	21.298	240	1400139	20.000	ng/u1	0.00
88) Perylene-d12	24.233	264	1508969	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.181	96	62454	6.526	ng/uL	-0.01
4) Pyridine-d5	3.587	84	483464	17.947	ng/u1	-0.01
7) Phenol-d5	6.846	99	654103	19.272	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.010	67	354467	17.855	ng/u1	-0.02
11) 2-Chlorophenol-d4	7.204	132	516229	19.934	ng/u1	-0.01
15) 4-Methylphenol-d8	8.381	113	531359	19.206	ng/u1	0.00
21) Nitrobenzene-d5	8.828	128	270414	20.080	ng/u1	-0.01
24) 2-Nitrophenol-d4	9.546	143	291292	20.939	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.081	165	528048	20.332	ng/u1	0.00
31) 4-Chloroaniline-d4	10.593	131	729296	18.913	ng/u1	-0.02
46) Dimethylphthalate-d6	13.728	166	1438652	18.827	ng/u1	-0.02
49) Acenaphthylene-d8	14.004	160	1759146	20.021	ng/u1	-0.01
54) 4-Nitrophenol-d4	14.534	143	280140	18.310	ng/u1	0.00
60) Fluorene-d10	15.310	176	1259664	19.483	ng/u1	-0.01
65) 4,6-Dinitro-2-methylph...	15.439	200	184486	17.367	ng/u1	0.00
73) Anthracene-d10	17.163	188	1830282	20.229	ng/u1	0.00
81) Pyrene-d10	19.463	212	1720572	22.964	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.033	264	1413327	19.615	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.216	88	69280	6.766	ng/uL	98
5) Pyridine	3.605	79	474930	17.632	ng/u1	96
6) Benzaldehyde	6.822	77	294911	18.525	ng/u1	98
8) Phenol	6.875	94	666772	19.303	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.104	93	522247	18.285	ng/u1	96
12) 2-Chlorophenol	7.234	128	535008	19.741	ng/u1	99
13) 2-Methylphenol	8.116	108	515336	19.316	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.193	45	608817	18.517	ng/u1	99
16) Acetophenone	8.493	105	798020	19.033	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.481	70	384448	17.694	ng/u1	96
18) 4-Methylphenol	8.446	108	558849	19.324	ng/u1	100
19) Hexachloroethane	8.734	117	210170	18.313	ng/u1	97
22) Nitrobenzene	8.869	77	610662	19.032	ng/u1	98
23) Isophorone	9.393	82	1135476	17.251	ng/u1	98
25) 2-Nitrophenol	9.575	139	310969	20.460	ng/u1	97
26) 2,4-Dimethylphenol	9.640	107	564624	19.096	ng/u1	97
27) Bis(2-Chloroethoxy)met...	9.875	93	708438	17.270	ng/u1	99
29) 2,4-Dichlorophenol	10.104	162	509271	19.860	ng/u1	97
30) Naphthalene	10.504	128	1758746	19.294	ng/u1	99
32) 4-Chloroaniline	10.622	127	715488	19.112	ng/u1	100
33) Hexachlorobutadiene	10.781	225	305532	19.838	ng/u1	97
34) Caprolactam	11.404	113	155729	15.558	ng/u1	96
35) 4-Chloro-3-methylphenol	11.751	107	574720	19.174	ng/u1	98
36) 2-Methylnaphthalene	12.122	142	1207313	19.377	ng/u1	100


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060824\
 Data File : BN031870.D
 Acq On : 08 Jun 2024 23:04
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020282

Quant Time: Jun 10 01:49:13 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN052924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed May 29 15:13:07 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.340	142	1208394	19.329	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.492	216	605464	20.872	ng/ul	99
40) Hexachlorocyclopentadiene	12.463	237	144668	8.653	ng/ul	91
41) 2,4,6-Trichlorophenol	12.740	196	411364	20.893	ng/ul	99
42) 2,4,5-Trichlorophenol	12.816	196	451504	21.102	ng/ul	97
43) 1,1'-Biphenyl	13.145	154	1520058	19.702	ng/ul	96
44) 2-Chloronaphthalene	13.187	162	1225454	20.280	ng/ul	99
45) 2-Nitroaniline	13.398	65	348639	19.904	ng/ul	90
47) Dimethylphthalate	13.781	163	1461265	18.698	ng/ul	99
48) 2,6-Dinitrotoluene	13.904	165	318701	20.516	ng/ul	96
50) Acenaphthylene	14.034	152	2011810	20.260	ng/ul	99
51) 3-Nitroaniline	14.228	138	348832	21.003	ng/ul	98
52) Acenaphthene	14.381	153	1312210	19.972	ng/ul	99
53) 2,4-Dinitrophenol	14.445	184	112441	14.338	ng/ul	97
55) 4-Nitrophenol	14.545	109	211430	18.629	ng/ul	96
56) Dibenzofuran	14.716	168	1774591	19.880	ng/ul	100
57) 2,4-Dinitrotoluene	14.692	165	453663	19.903	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	14.945	232	356999	19.947	ng/ul	100
59) Diethylphthalate	15.145	149	1454783	18.040	ng/ul	99
61) Fluorene	15.369	166	1423731	19.806	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.363	204	684109	19.649	ng/ul	96
63) 4-Nitroaniline	15.398	138	337824	21.367	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.457	198	201782	17.868	ng/ul	100
67) N-Nitrosodiphenylamine	15.581	169	1248846	21.438	ng/ul	98
68) 4-Bromophenyl-phenylether	16.257	248	438506	21.415	ng/ul	95
69) Hexachlorobenzene	16.363	284	484760	20.888	ng/ul	97
70) Atrazine	16.533	200	345366	18.777	ng/ul	98
71) Pentachlorophenol	16.716	266	281818	18.980	ng/ul#	84
72) Phenanthrene	17.104	178	2173424	20.188	ng/ul	99
74) Anthracene	17.198	178	2211498	20.273	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.104	216	607111	22.319	ng/ul	96
76) Pentachlorobenzene	14.634	250	569236	21.485	ng/ul	96
77) Carbazole	17.469	167	1935848	20.812	ng/ul	98
78) Di-n-butylphthalate	18.033	149	2547761	19.909	ng/ul	100
80) Fluoranthene	19.127	202	2133679	23.829	ng/ul	97
82) Pyrene	19.492	202	2101129	22.871	ng/ul	97
83) Butylbenzylphthalate	20.398	149	953442	21.973	ng/ul	93
84) 3,3'-Dichlorobenzidine	21.210	252	503863	18.766	ng/ul	93
85) Benzo(a)anthracene	21.280	228	1842659	19.662	ng/ul	96
86) Bis(2-ethylhexyl)phtha...	21.204	149	1383550	21.874	ng/ul	96
87) Chrysene	21.339	228	1687858	19.281	ng/ul	97
89) Di-n-octyl phthalate	22.315	149	2154288	21.325	ng/ul	100
90) Benzo(b)fluoranthene	23.304	252	1706068	18.915	ng/ul	98
91) Benzo(k)fluoranthene	23.368	252	1752784	19.804	ng/ul	99
93) Benzo(a)pyrene	24.098	252	1675796	19.838	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	27.545	276	1980360	20.344	ng/ul	99
95) Dibenzo(a,h)anthracene	27.598	278	984294	12.385	ng/ul	98
96) Benzo(g,h,i)perylene	28.580	276	1590082	20.549	ng/ul	97





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
 1,4-Dioxane
 Concen: 6.766 ng/uL
 RT: 3.216 min Scan# 39
 Delta R.T. -0.012 min
 Lab File: BN031870.D
 Acq: 08 Jun 2024 23:04




Instrument :
 BNA_N
 ClientSampleId :
 SST020282


Tgt Ion: 88 Resp: 69280
 Ion Ratio Lower Upper
 88 100
 43 23.5 18.5 27.7
 58 63.4 52.6 78.8



#5
 Pyridine
 Concen: 17.632 ng/uL
 RT: 3.605 min Scan# 105
 Delta R.T. -0.012 min
 Lab File: BN031870.D
 Acq: 08 Jun 2024 23:04

Tgt Ion: 79 Resp: 474930
 Ion Ratio Lower Upper
 79 100
 52 70.1 53.0 79.6
 51 31.0 23.3 34.9









#6
Benzaldehyde
Concen: 18.525 ng/ul
RT: 6.822 min Scan# 611
Delta R.T. -0.012 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04




Instrument :
BNA_N
ClientSampleId :
SSTD020282


Tgt Ion: 77 Resp: 294911
Ion Ratio Lower Upper
77 100
105 102.6 80.2 120.4
106 97.0 76.8 115.2



#8
Phenol
Concen: 19.303 ng/ul
RT: 6.875 min Scan# 661
Delta R.T. -0.006 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04

Tgt Ion: 94 Resp: 666772
Ion Ratio Lower Upper
94 100
65 25.7 20.3 30.5
66 35.1 25.8 38.8









#10
 Bis(2-Chloroethyl)ether
 Concen: 18.285 ng/ul
 RT: 7.104 min Scan# 700
 Delta R.T. -0.012 min
 Lab File: BN031870.D
 Acq: 08 Jun 2024 23:04

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020282

Tgt Ion: 93 Resp: 522247




Ion	Ratio	Lower	Upper
93	100		
63	66.0	56.8	85.2
95	32.6	26.0	39.0




#12
 2-Chlorophenol
 Concen: 19.741 ng/ul
 RT: 7.234 min Scan# 722
 Delta R.T. -0.018 min
 Lab File: BN031870.D
 Acq: 08 Jun 2024 23:04

Tgt Ion: 128 Resp: 535008

Ion	Ratio	Lower	Upper
128	100		
64	43.2	33.9	50.9
130	32.0	26.3	39.5





#13

2-Methylphenol

Concen: 19.316 ng/ul

RT: 8.116 min Scan# 81

Delta R.T. -0.006 min

Lab File: BN031870.D

Acq: 08 Jun 2024 23:04

Instrument :

BNA_N

ClientSampleId :





SSTD020282

Tgt Ion:108 Resp: 515336

Ion Ratio Lower Upper

108 100

107 89.1 71.6 107.4



#14

2,2'-oxybis(1-Chloropropane)

Concen: 18.517 ng/ul

RT: 8.193 min Scan# 885

Delta R.T. -0.024 min

Lab File: BN031870.D

Acq: 08 Jun 2024 23:04




Tgt Ion: 45 Resp: 608817


Ion Ratio Lower Upper

45 100

77 17.5 13.6 20.4

79 12.9 10.6 15.8









#16
Acetophenone
Concen: 19.033 ng/ul
RT: 8.493 min Scan# 91
Delta R.T. -0.012 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04




Instrument :
BNA_N
ClientSampleId :
SSTD020282


Tgt Ion:105 Resp: 798020
Ion Ratio Lower Upper
105 100
77 74.6 58.6 87.8
51 22.6 17.2 25.8



#17
N-Nitroso-di-n-propylamine
Concen: 17.694 ng/ul
RT: 8.481 min Scan# 934
Delta R.T. -0.012 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04

Tgt Ion: 70 Resp: 384448
Ion Ratio Lower Upper
70 100
42 48.5 37.3 55.9
101 11.1 7.8 11.6
130 23.7 16.4 24.6









#18
4-Methylphenol
Concen: 19.324 ng/ul
RT: 8.446 min Scan# 911
Delta R.T. -0.006 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04




Instrument :
BNA_N
ClientSampleId :
SSTD020282


Tgt Ion:108 Resp: 558849
Ion Ratio Lower Upper
108 100
107 112.1 89.6 134.4



#19
Hexachloroethane
Concen: 18.313 ng/ul
RT: 8.734 min Scan# 977
Delta R.T. -0.024 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04

Tgt Ion:117 Resp: 210170
Ion Ratio Lower Upper
117 100
201 91.6 71.6 107.4
199 60.3 45.6 68.4









#22
Nitrobenzene
Concen: 19.032 ng/ul
RT: 8.869 min Scan# 1000
Delta R.T. -0.012 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04




Instrument :
BNA_N
ClientSampleId :
SSTD020282


Tgt Ion: 77 Resp: 610662
Ion Ratio Lower Upper
77 100
123 47.8 36.6 55.0
65 12.7 10.2 15.4



#23
Isophorone
Concen: 17.251 ng/ul
RT: 9.393 min Scan# 1089
Delta R.T. -0.012 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04

Tgt Ion: 82 Resp: 1135476
Ion Ratio Lower Upper
82 100
95 7.2 5.5 8.3
138 18.5 14.1 21.1









#25
2-Nitrophenol
Concen: 20.460 ng/ul
RT: 9.575 min Scan# 1120
Delta R.T. -0.012 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04




Instrument :
BNA_N
ClientSampleId :
SSTD020282


Tgt Ion:139 Resp: 310969
Ion Ratio Lower Upper
139 100
65 47.2 38.3 57.5
109 30.4 22.1 33.1



#26
2,4-Dimethylphenol
Concen: 19.096 ng/ul
RT: 9.640 min Scan# 1131
Delta R.T. -0.006 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04

Tgt Ion:107 Resp: 564624
Ion Ratio Lower Upper
107 100
121 51.9 40.8 61.2
122 84.7 70.4 105.6





#27

Bis(2-Chloroethoxy)methane

Concen: 17.270 ng/ul

RT: 9.875 min Scan# 1171

Delta R.T. -0.018 min

Lab File: BN031870.D

Acq: 08 Jun 2024 23:04

Instrument :

BNA_N

ClientSampleId :

SSTD020282





Tgt Ion: 93 Resp: 708438

Ion Ratio Lower Upper

93 100

95 32.3 26.1 39.1

123 11.8 9.0 13.6



#29

2,4-Dichlorophenol

Concen: 19.860 ng/ul

RT: 10.104 min Scan# 1210

Delta R.T. -0.012 min

Lab File: BN031870.D

Acq: 08 Jun 2024 23:04




Tgt Ion: 162 Resp: 509271


Ion Ratio Lower Upper

162 100

164 63.4 52.7 79.1

98 35.7 30.5 45.7









#30
Naphthalene
Concen: 19.294 ng/ul
RT: 10.504 min Scan# 1278
Delta R.T. -0.012 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04

Instrument :
BNA_N
ClientSampleId :
SSTD020282

Tgt Ion:128 Resp: 1758746




Ion	Ratio	Lower	Upper
128	100		
129	10.9	8.5	12.7
127	12.8	10.3	15.5




#32
4-Chloroaniline
Concen: 19.112 ng/ul
RT: 10.622 min Scan# 1298
Delta R.T. -0.006 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04

Tgt Ion:127 Resp: 715488

Ion	Ratio	Lower	Upper
127	100		
129	31.1	25.0	37.4





#33

Hexachlorobutadiene

Concen: 19.838 ng/ul

RT: 10.781 min Scan# 1325

Delta R.T. -0.018 min

Lab File: BN031870.D

Acq: 08 Jun 2024 23:04

Instrument :

BNA_N

ClientSampleId :

SSTD020282





Tgt Ion:225 Resp: 305532

Ion Ratio Lower Upper

225 100

223 60.4 52.3 78.5

227 65.9 52.5 78.7



#34

Caprolactam

Concen: 15.558 ng/ul

RT: 11.404 min Scan# 1431

Delta R.T. 0.006 min

Lab File: BN031870.D

Acq: 08 Jun 2024 23:04




Tgt Ion:113 Resp: 155729


Ion Ratio Lower Upper

113 100

55 140.9 117.2 175.8

56 109.2 89.4 134.0









#35
4-Chloro-3-methylphenol
Concen: 19.174 ng/ul
RT: 11.751 min Scan# 1490
Delta R.T. -0.000 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04




Instrument :
BNA_N
ClientSampleId :
SSTD020282


Tgt Ion:107 Resp: 574720
Ion Ratio Lower Upper
107 100
144 25.9 20.5 30.7
142 77.7 64.0 96.0



#36
2-Methylnaphthalene
Concen: 19.377 ng/ul
RT: 12.122 min Scan# 1553
Delta R.T. -0.012 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04

Tgt Ion:142 Resp: 1207313
Ion Ratio Lower Upper
142 100
141 87.5 70.1 105.1









#37
1-Methylnaphthalene
Concen: 19.329 ng/ul
RT: 12.340 min Scan# 1590
Delta R.T. -0.018 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04




Instrument :
BNA_N
ClientSampleId :
SSTD020282


Tgt Ion:142 Resp: 1208394
Ion Ratio Lower Upper
142 100
141 92.0 71.7 107.5
116 3.7 3.0 4.4



#39
1,2,4,5-Tetrachlorobenzene
Concen: 20.872 ng/ul
RT: 12.492 min Scan# 1616
Delta R.T. -0.012 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04

Tgt Ion:216 Resp: 605464
Ion Ratio Lower Upper
216 100
214 77.1 61.7 92.5
179 20.3 16.7 25.1
108 16.9 14.5 21.7









#40
Hexachlorocyclopentadiene
Concen: 8.653 ng/ul
RT: 12.463 min Scan# 1611
Delta R.T. -0.018 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04




Instrument :
BNA_N
ClientSampleId :
SSTD020282


Tgt Ion:237 Resp: 144668
Ion Ratio Lower Upper
237 100
235 67.5 47.8 71.6
272 14.0 10.0 15.0



#41
2,4,6-Trichlorophenol
Concen: 20.893 ng/ul
RT: 12.740 min Scan# 1658
Delta R.T. -0.006 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04

Tgt Ion:196 Resp: 411364
Ion Ratio Lower Upper
196 100
198 94.6 76.4 114.6
200 30.8 24.2 36.4









#42
2,4,5-Trichlorophenol
Concen: 21.102 ng/ul
RT: 12.816 min Scan# 1671
Delta R.T. -0.000 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04




Instrument :
BNA_N
ClientSampleId :
SSTD020282


Tgt Ion:196 Resp: 451504
Ion Ratio Lower Upper
196 100
198 94.5 73.8 110.6
200 30.5 25.6 38.4



#43
1,1'-Biphenyl
Concen: 19.702 ng/ul
RT: 13.145 min Scan# 1727
Delta R.T. -0.012 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04

Tgt Ion:154 Resp: 1520058
Ion Ratio Lower Upper
154 100
153 40.9 30.6 46.0
76 13.0 11.1 16.7









#44
2-Chloronaphthalene
Concen: 20.280 ng/ul
RT: 13.187 min Scan# 1734
Delta R.T. -0.012 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04




Instrument :
BNA_N
ClientSampleId :
SSTD020282


Tgt Ion:162 Resp: 1225454
Ion Ratio Lower Upper
162 100
164 32.9 27.0 40.4
127 39.4 31.8 47.6



#45
2-Nitroaniline
Concen: 19.904 ng/ul
RT: 13.398 min Scan# 1770
Delta R.T. -0.006 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04

Tgt Ion: 65 Resp: 348639
Ion Ratio Lower Upper
65 100
92 78.1 58.8 88.2
138 122.2 86.2 129.4









#47
Dimethylphthalate
Concen: 18.698 ng/ul
RT: 13.781 min Scan# 1835
Delta R.T. -0.012 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04




Instrument :
BNA_N
ClientSampleId :
SSTD020282


Tgt Ion:163 Resp: 1461265
Ion Ratio Lower Upper
163 100
194 4.1 3.2 4.8
164 9.9 8.2 12.2



#48
2,6-Dinitrotoluene
Concen: 20.516 ng/ul
RT: 13.904 min Scan# 1856
Delta R.T. -0.006 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04

Tgt Ion:165 Resp: 318701
Ion Ratio Lower Upper
165 100
89 52.5 44.6 67.0
121 20.4 17.3 25.9









#50
Acenaphthylene
Concen: 20.260 ng/ul
RT: 14.034 min Scan# 1878
Delta R.T. -0.012 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04

Instrument : BNA_N
ClientSampleId : SSTD020282

Tgt Ion:152 Resp: 2011810




Ion	Ratio	Lower	Upper
152	100		
151	20.3	15.9	23.9
153	13.3	10.4	15.6




#51
3-Nitroaniline
Concen: 21.003 ng/ul
RT: 14.228 min Scan# 1911
Delta R.T. -0.006 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04

Tgt Ion:138 Resp: 348832

Ion	Ratio	Lower	Upper
138	100		
108	11.4	9.2	13.8
92	115.6	94.0	141.0









#52
Acenaphthene
Concen: 19.972 ng/ul
RT: 14.381 min Scan# 1937
Delta R.T. -0.006 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04




Instrument :
BNA_N
ClientSampleId :
SSTD020282


Tgt Ion:153 Resp: 1312210
Ion Ratio Lower Upper
153 100
152 48.3 37.8 56.6
154 86.3 69.8 104.8



#53
2,4-Dinitrophenol
Concen: 14.338 ng/ul
RT: 14.445 min Scan# 1948
Delta R.T. 0.006 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04

Tgt Ion:184 Resp: 112441
Ion Ratio Lower Upper
184 100
63 59.7 50.2 75.4
154 62.7 49.1 73.7









#55
4-Nitrophenol
Concen: 18.629 ng/ul
RT: 14.545 min Scan# 1965
Delta R.T. 0.006 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04




Instrument :
BNA_N
ClientSampleId :
SSTD020282


Tgt Ion:109 Resp: 211430
Ion Ratio Lower Upper
109 100
139 137.7 113.4 170.2
65 126.8 106.2 159.2



#56
Dibenzofuran
Concen: 19.880 ng/ul
RT: 14.716 min Scan# 1994
Delta R.T. -0.012 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04

Tgt Ion:168 Resp: 1774591
Ion Ratio Lower Upper
168 100
139 37.0 29.4 44.0









#57
2,4-Dinitrotoluene
Concen: 19.903 ng/ul
RT: 14.692 min Scan# 1990
Delta R.T. -0.000 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04




Instrument :
BNA_N
ClientSampleId :
SSTD020282


Tgt Ion:165 Resp: 453663
Ion Ratio Lower Upper
165 100
63 40.5 32.2 48.2
182 2.9 2.3 3.5



#58
2,3,4,6-Tetrachlorophenol
Concen: 19.947 ng/ul
RT: 14.945 min Scan# 2033
Delta R.T. -0.006 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04

Tgt Ion:232 Resp: 356999
Ion Ratio Lower Upper
232 100
131 45.8 36.6 55.0
130 2.4 1.8 2.8
166 31.6 25.0 37.4









#59
Diethylphthalate
Concen: 18.040 ng/ul
RT: 15.145 min Scan# 2067
Delta R.T. -0.012 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04




Instrument :
BNA_N
ClientSampleId :
SSTD020282


Tgt Ion:149 Resp: 1454783
Ion Ratio Lower Upper
149 100
177 22.2 17.1 25.7
150 12.2 10.0 15.0



#61
Fluorene
Concen: 19.806 ng/ul
RT: 15.369 min Scan# 2105
Delta R.T. -0.006 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04

Tgt Ion:166 Resp: 1423731
Ion Ratio Lower Upper
166 100
165 97.8 80.3 120.5
167 13.5 11.3 16.9









#62
4-Chlorophenyl-phenylether
Concen: 19.649 ng/ul
RT: 15.363 min Scan# 2104
Delta R.T. -0.012 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04




Instrument :
BNA_N
ClientSampleId :
SSTD020282


Tgt Ion:204 Resp: 684109
Ion Ratio Lower Upper
204 100
206 33.7 25.7 38.5
141 62.8 47.7 71.5



#63
4-Nitroaniline
Concen: 21.367 ng/ul
RT: 15.398 min Scan# 2110
Delta R.T. -0.000 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04

Tgt Ion:138 Resp: 337824
Ion Ratio Lower Upper
138 100
92 48.5 41.0 61.4
108 76.4 60.1 90.1









#66
 4,6-Dinitro-2-methylphenol
 Concen: 17.868 ng/ul
 RT: 15.457 min Scan# 2120
 Delta R.T. -0.000 min
 Lab File: BN031870.D
 Acq: 08 Jun 2024 23:04




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020282


Tgt Ion:198 Resp: 201782
 Ion Ratio Lower Upper
 198 100
 182 4.1 3.0 4.6
 77 24.1 19.4 29.2



#67
 N-Nitrosodiphenylamine
 Concen: 21.438 ng/ul
 RT: 15.581 min Scan# 2141
 Delta R.T. -0.006 min
 Lab File: BN031870.D
 Acq: 08 Jun 2024 23:04

Tgt Ion:169 Resp: 1248846
 Ion Ratio Lower Upper
 169 100
 168 67.6 52.7 79.1
 167 35.3 29.6 44.4









#68
4-Bromophenyl-phenylether
Concen: 21.415 ng/ul
RT: 16.257 min Scan# 21
Delta R.T. -0.012 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04




Instrument :
BNA_N
ClientSampleId :
SSTD020282


Tgt Ion:248 Resp: 438506
Ion Ratio Lower Upper
248 100
250 97.2 76.5 114.7
141 71.6 63.9 95.9



#69
Hexachlorobenzene
Concen: 20.888 ng/ul
RT: 16.363 min Scan# 2274
Delta R.T. -0.006 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04

Tgt Ion:284 Resp: 484760
Ion Ratio Lower Upper
284 100
142 35.9 31.1 46.7
249 31.5 25.7 38.5





#70

Atrazine

Concen: 18.777 ng/ul

RT: 16.533 min Scan# 21

Delta R.T. -0.012 min

Lab File: BN031870.D

Acq: 08 Jun 2024 23:04

Instrument :

BNA_N

ClientSampleId :

SSTD020282





Tgt Ion:200 Resp: 345366

Ion Ratio Lower Upper

200 100

173 27.3 21.2 31.8

215 50.1 38.9 58.3



#71

Pentachlorophenol

Concen: 18.980 ng/ul

RT: 16.716 min Scan# 2334

Delta R.T. -0.000 min

Lab File: BN031870.D

Acq: 08 Jun 2024 23:04




Tgt Ion:266 Resp: 281818


Ion Ratio Lower Upper

266 100

264 52.8 54.7 82.1#

268 57.3 53.4 80.2





#72

Phenanthrene

Concen: 20.188 ng/ul

RT: 17.104 min Scan# 2400

Delta R.T. -0.012 min

Lab File: BN031870.D

Acq: 08 Jun 2024 23:04

Instrument :

BNA_N

ClientSampleId :

SSTD020282





Tgt Ion:178 Resp: 2173424

Ion Ratio Lower Upper

178 100

179 15.5 12.6 19.0

176 18.9 15.8 23.6



#74

Anthracene

Concen: 20.273 ng/ul

RT: 17.198 min Scan# 2416

Delta R.T. -0.006 min

Lab File: BN031870.D

Acq: 08 Jun 2024 23:04




Tgt Ion:178 Resp: 2211498


Ion Ratio Lower Upper

178 100

179 15.9 11.8 17.6

176 18.5 14.8 22.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 22.319 ng/uL
RT: 13.104 min Scan# 1720
Delta R.T. -0.012 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04




Instrument :
BNA_N
ClientSampleId :
SSTD020282


Tgt Ion:216 Resp: 607111
Ion Ratio Lower Upper
216 100
214 79.5 62.1 93.1
218 46.2 40.2 60.4



#76
Pentachlorobenzene
Concen: 21.485 ng/uL
RT: 14.634 min Scan# 1980
Delta R.T. -0.006 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04

Tgt Ion:250 Resp: 569236
Ion Ratio Lower Upper
250 100
248 67.7 49.8 74.6
252 65.6 52.3 78.5









#77
Carbazole
Concen: 20.812 ng/ul
RT: 17.469 min Scan# 2462
Delta R.T. -0.006 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04




Instrument :
BNA_N
ClientSampleId :
SSTD020282


Tgt Ion:167 Resp: 1935848
Ion Ratio Lower Upper
167 100
166 21.0 17.9 26.9
139 13.2 10.5 15.7



#78
Di-n-butylphthalate
Concen: 19.909 ng/ul
RT: 18.033 min Scan# 2558
Delta R.T. -0.012 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04

Tgt Ion:149 Resp: 2547761
Ion Ratio Lower Upper
149 100
150 9.3 7.4 11.0
104 5.1 4.1 6.1









#80
Fluoranthene
Concen: 23.829 ng/u1
RT: 19.127 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04




Instrument :
BNA_N
ClientSampleId :
SSTD020282


Tgt Ion:202 Resp: 2133679
Ion Ratio Lower Upper
202 100
101 11.4 10.3 15.5
100 8.8 8.0 12.0



#82
Pyrene
Concen: 22.871 ng/u1
RT: 19.492 min Scan# 2806
Delta R.T. -0.000 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04

Tgt Ion:202 Resp: 2101129
Ion Ratio Lower Upper
202 100
101 14.4 12.6 18.8
100 11.6 10.2 15.2





#83

Butylbenzylphthalate

Concen: 21.973 ng/ul

RT: 20.398 min Scan# 2960

Delta R.T. -0.006 min

Lab File: BN031870.D

Acq: 08 Jun 2024 23:04

Instrument :

BNA_N

ClientSampleId :

SSTD020282





Tgt Ion:149 Resp: 953442

Ion Ratio Lower Upper

149 100

91 69.7 61.0 91.4

206 22.9 16.5 24.7



#84

3,3'-Dichlorobenzidine

Concen: 18.766 ng/ul

RT: 21.210 min Scan# 3098

Delta R.T. -0.006 min

Lab File: BN031870.D

Acq: 08 Jun 2024 23:04




Tgt Ion:252 Resp: 503863


Ion Ratio Lower Upper

252 100

254 60.7 53.6 80.4

126 11.4 10.2 15.4





#85

Benzo(a)anthracene

Concen: 19.662 ng/ul

RT: 21.280 min Scan# 3111

Delta R.T. -0.006 min

Lab File: BN031870.D

Acq: 08 Jun 2024 23:04

Instrument :

BNA_N

ClientSampleId :

SSTD020282





Tgt Ion:228 Resp: 1842659

Ion Ratio Lower Upper

228 100

229 18.6 16.7 25.1

226 26.8 22.8 34.2



#86

Bis(2-ethylhexyl)phthalate

Concen: 21.874 ng/ul

RT: 21.204 min Scan# 3097

Delta R.T. -0.018 min

Lab File: BN031870.D

Acq: 08 Jun 2024 23:04




Tgt Ion:149 Resp: 1383550


Ion Ratio Lower Upper

149 100

167 29.2 21.4 32.0

279 4.2 3.3 4.9





#87

Chrysene

Concen: 19.281 ng/ul

RT: 21.339 min Scan# 3121

Delta R.T. -0.012 min

Lab File: BN031870.D

Acq: 08 Jun 2024 23:04

Instrument :

BNA_N

ClientSampleId :

SSTD020282





Tgt Ion:228 Resp: 1687858

Ion Ratio Lower Upper

228 100

226 27.9 23.7 35.5

229 20.1 15.1 22.7



#89

Di-n-octyl phthalate

Concen: 21.325 ng/ul




RT: 22.315 min Scan# 3286


Delta R.T. -0.024 min

Lab File: BN031870.D

Acq: 08 Jun 2024 23:04

Tgt Ion:149 Resp: 2154288









#90
Benzo(b)fluoranthene
Concen: 18.915 ng/ul
RT: 23.304 min Scan# 3456
Delta R.T. -0.006 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04




Instrument :
BNA_N
ClientSampleId :
SSTD020282


Tgt Ion:252 Resp: 1706068
Ion Ratio Lower Upper
252 100
253 20.1 17.0 25.6
125 11.2 9.2 13.8



#91
Benzo(k)fluoranthene
Concen: 19.804 ng/ul
RT: 23.368 min Scan# 3465
Delta R.T. -0.006 min
Lab File: BN031870.D
Acq: 08 Jun 2024 23:04

Tgt Ion:252 Resp: 1752784
Ion Ratio Lower Upper
252 100
253 21.4 17.6 26.4
125 11.2 9.4 14.2





#93

Benzo(a)pyrene

Concen: 19.838 ng/ul

RT: 24.098 min Scan# 31

Delta R.T. -0.006 min

Lab File: BN031870.D

Acq: 08 Jun 2024 23:04

Instrument :

BNA_N

ClientSampleId :

SSTD020282





Tgt Ion:252 Resp: 1675796

Ion Ratio Lower Upper

252 100

253 21.1 16.6 24.8

125 12.1 9.8 14.8



#94

Indeno(1,2,3-cd)pyrene

Concen: 20.344 ng/ul

RT: 27.545 min Scan# 4175

Delta R.T. 0.012 min

Lab File: BN031870.D

Acq: 08 Jun 2024 23:04




Tgt Ion:276 Resp: 1980360


Ion Ratio Lower Upper

276 100

138 21.8 18.1 27.1

277 23.2 18.2 27.2





#95

Dibenzo(a,h)anthracene

Concen: 12.385 ng/ul

RT: 27.598 min Scan# 4186

Delta R.T. -0.006 min

Lab File: BN031870.D

Acq: 08 Jun 2024 23:04

Instrument :

BNA_N

ClientSampleId :

SSTD020282





Tgt Ion:278 Resp: 984294

Ion Ratio Lower Upper

278 100

139 18.3 15.0 22.6

279 22.6 17.4 26.0



#96

Benzo(g,h,i)perylene

Concen: 20.549 ng/ul

RT: 28.580 min Scan# 4351

Delta R.T. 0.012 min

Lab File: BN031870.D

Acq: 08 Jun 2024 23:04

Tgt Ion:276 Resp: 1590082

Ion Ratio Lower Upper

276 100

138 23.0 20.9 31.3

277 22.2 18.0 27.0

