

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061019\
 Data File : BN006238.D
 Acq On : 11 Jun 2019 00:03
 Operator : JU/SJ
 Sample : K3017-24
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 A51G2

Manual Integrations
 APPROVED

mohammad
 6/11/2019 4:29:11 PM

Quant Time: Jun 17 07:13:25 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN053019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Jun 09 05:10:30 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.65	152	4038	0.40	ng/ul	0.00
2) Naphthalene-d8	10.42	136	14980	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.30	164	8402	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.05	188	16045m	0.40	ng/ul	0.00
16) Chrysene-d12	21.27	240	12172m	0.40	ng/ul	0.00
20) Perylene-d12	23.49	264	11077m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.02	152	8911	0.39	ng/ul	0.00
14) Fluoranthene-d10	19.09	212	18811m	0.44	ng/ul	0.00
Target Compounds						
3) Naphthalene	10.47	128	4667	0.113	ng/ul#	Ovalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061019\
Data File : BN006238.D
Acq On : 11 Jun 2019 00:03
Operator : JU/SJ
Sample : K3017-24
Misc :
ALS Vial : 26 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
A51G2

Manual Integrations
APPROVED

mohammad
6/11/2019 4:29:11 PM

Quant Time: Jun 17 07:13:25 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN053019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Jun 09 05:10:30 2019
Response via : Initial Calibration





#3

Naphthalene

Concen: 0.113 ng/ul

RT: 10.47 min Scan# 2336

Delta R.T. -0.00 min

Lab File: BN006238.D


Acq: 11 Jun 2019 00:03

Instrument :

BNA_N

ClientSampleId :

A51G2



Tot Ion:	128	Resp:	4667
Ion Ratio	Lower	Upper	
128	100		
129	15.4	9.3	13.9#
127	15.5	10.7	16.1

Manual Integrations
APPROVED

mohammad
6/11/2019 4:29:11 PM

