

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN061022\  
 Data File : BN020084.D  
 Acq On : 09 Jun 2022 17:16  
 Operator : CG/JU  
 Sample : PB145382BL  
 Misc :  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 SBLK382

Quant Time: Jun 10 02:43:00 2022  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\SFAM-EPA-SIM-BN053122.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Jun 10 02:40:52 2022  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc  | Units | Dev(Min) |
|-----------------------------|--------|------|----------|-------|-------|----------|
| -----                       |        |      |          |       |       |          |
| Internal Standards          |        |      |          |       |       |          |
| 1) 1,4-Dichlorobenzene-d4   | 7.830  | 152  | 3724     | 0.400 | ng/ul | 0.00     |
| 4) Naphthalene-d8           | 10.611 | 136  | 12615    | 0.400 | ng/ul | 0.00     |
| 9) Acenaphthene-d10         | 14.450 | 164  | 8004     | 0.400 | ng/ul | 0.00     |
| 13) Phenanthrene-d10        | 17.190 | 188  | 17145    | 0.400 | ng/ul | 0.00     |
| 17) Chrysene-d12            | 21.382 | 240  | 13977    | 0.400 | ng/ul | 0.00     |
| 23) Perylene-d12            | 23.709 | 264  | 11978    | 0.400 | ng/ul | # 0.00   |
| System Monitoring Compounds |        |      |          |       |       |          |
| 3) 1,4-Dioxane-d8           | 3.286  | 96   | 22614    | 5.329 | ng/ul | 0.00     |
| 6) 2-Methylnaphthalene-d10  | 12.205 | 152  | 5884     | 0.292 | ng/ul | 0.00     |
| 18) Fluoranthene-d10        | 19.219 | 212  | 13159    | 0.311 | ng/ul | 0.00     |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN061022\  
Data File : BN020084.D  
Acq On : 09 Jun 2022 17:16  
Operator : CG/JU  
Sample : PB145382BL  
Misc :  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
SBLK382

Quant Time: Jun 10 02:43:00 2022  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\SFAM-EPA-SIM-BN053122.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri Jun 10 02:40:52 2022  
Response via : Initial Calibration

