

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061022\
 Data File : BN020089.D
 Acq On : 09 Jun 2022 20:19
 Operator : CG/JU
 Sample : N3212-01
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 F5A50

Manual Integrations
 APPROVED


Reviewed By :Jagrut Upadhyay 06/10/2022
 Supervised By :mohammad ahmed 06/10/2022

Quant Time: Jun 10 02:43:19 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN053122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 10 02:40:52 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.825	152	4283	0.400	ng/ul	0.00
4) Naphthalene-d8	10.611	136	14684	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.450	164	9140	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.190	188	19763	0.400	ng/ul	0.00
17) Chrysene-d12	21.376	240	17452	0.400	ng/ul	0.00
23) Perylene-d12	23.703	264	14534	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.286	96	22500	4.610	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.205	152	6475	0.276	ng/ul	0.00
18) Fluoranthene-d10	19.219	212	18149	0.344	ng/ul	0.00
Target Compounds						
2) 1,4-Dioxane	3.323	88	573m	0.112	ng/ul	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
 1,4-Dioxane
 Concen: 0.112 ng/ul m
 RT: 3.323 min Scan# 51
 Delta R.T. 0.000 min
 Lab File: BN020089.D
 Acq: 09 Jun 2022 20:19

Instrument :
 BNA_N
 ClientSampleId :
 F5A50



Tgt Ion: 88	Resp: 57
Ion Ratio	Lower Upper
88 100	
43 48.1	39.0 58.4
58 41.0	39.8 59.8

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 06/10/2022
 Supervised By :mohammad ahmed 06/10/2022

