


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061022\
 Data File : BN020081.D
 Acq On : 09 Jun 2022 15:27
 Operator : CG/JU
 Sample : DFTPP254
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP254

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN053122.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Jun 10 02:40:52 2022



AutoFind: Scans 2474, 2475, 2476; Background Corrected with Scan 2467

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	43.3	169813	PASS
68	69	0.00	2	1.8	2886	PASS
69	198	0.00	100	41.0	160811	PASS
70	69	0.00	2	0.6	991	PASS
127	198	10	80	52.7	206592	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	392320	PASS
199	198	5	9	7.0	27528	PASS
275	198	10	60	24.6	96627	PASS
365	198	1	100	3.1	12237	PASS
441	198	0.01	100	12.4	48712	PASS
442	198	50	100	83.1	326144	PASS
443	442	15	24	18.7	61013	PASS

m/z	Abundance
39.90	1252.0
41.00	264.0
42.90	272.0
43.90	843.0
49.00	275.0
55.00	186.0
57.00	212.0
67.00	155.0
72.90	374.0
83.90	161.0
207.00	845.0
207.90	203.0
281.00	205.0

Instrument :
BNA_N
ClientSampleId :
DFTPP254

m/z	Abundance
37.00	729.0
38.00	1773.0
39.00	11784.0
40.00	1253.0
41.00	627.0
43.10	261.0
44.00	834.0
45.00	379.0
49.00	1315.0
50.00	40592.0
51.00	160896.0
52.00	7897.0
53.00	315.0
54.90	1014.0
56.00	4578.0
57.00	10744.0
57.90	486.0
60.90	2086.0
62.00	2246.0
63.00	6657.0
64.00	761.0
65.00	2863.0
66.10	225.0
67.00	266.0
68.00	2927.0
69.00	149760.0
70.00	795.0
70.90	206.0
73.00	1385.0
74.00	15386.0
75.00	24280.0
76.10	8792.0
77.00	168128.0
78.00	11638.0
79.00	11677.0
80.00	8225.0
81.00	12255.0
81.90	3044.0
82.90	2794.0
83.90	548.0
85.00	3093.0
86.00	3594.0
87.00	1665.0
87.90	551.0
89.10	331.0
90.90	2673.0
91.90	3312.0
93.00	20216.0
93.90	1544.0
95.00	337.0
95.90	869.0
97.00	484.0
98.00	15645.0
99.00	13083.0
99.90	1124.0
101.00	7220.0
101.90	427.0
102.90	2082.0
104.00	4205.0
105.00	4354.0
106.00	1487.0
107.00	53032.0
108.00	8328.0
109.00	1937.0
110.00	97656.0
111.00	14239.0
111.90	1859.0
112.90	585.0
114.20	163.0
114.80	240.0
115.90	3110.0
117.00	45344.0
118.00	3535.0
118.90	449.0
120.00	777.0
120.80	250.0
121.90	3068.0
123.00	5198.0
123.90	2476.0
125.00	2193.0
127.00	185664.0
128.00	14000.0
129.00	70720.0
130.00	6074.0
130.90	1098.0
132.00	574.0
132.70	347.0
133.90	1964.0
134.90	6406.0
136.00	2347.0
137.00	3297.0
137.80	634.0

Instrument :
BNA_N
ClientSampleId :
DFTPP254

138.80	292.0
139.90	992.0
140.90	8971.0
142.00	3043.0
142.90	2178.0
143.80	539.0
144.90	485.0
145.90	1591.0
146.90	4785.0
147.90	10828.0
148.80	2383.0
150.00	749.0
151.00	1333.0
151.70	741.0
153.00	3192.0
153.90	2194.0
155.00	5783.0
156.00	7882.0
157.00	1248.0
157.90	1516.0
158.90	1450.0
160.00	2849.0
160.90	4064.0
162.00	1187.0
162.90	495.0
163.90	266.0
164.90	3574.0
166.00	2758.0
167.00	18368.0
168.00	10813.0
168.90	1702.0
170.00	623.0
170.90	745.0
171.90	1394.0
172.90	1902.0
174.00	4114.0
175.00	6689.0
176.00	1684.0
176.90	3543.0
178.00	1113.0
178.90	13678.0
180.00	9360.0
181.00	4245.0
182.00	641.0
182.90	314.0
183.90	965.0
184.90	6364.0
186.00	47592.0
187.00	13164.0
188.00	1701.0
189.00	3111.0
190.00	590.0
190.90	1508.0
191.90	3947.0
193.00	4955.0
193.90	1098.0
195.00	660.0
196.00	8143.0
197.90	320960.0
198.90	23256.0
199.90	2179.0
201.50	1445.0
202.90	2383.0
204.00	11777.0
205.00	20432.0
206.00	87104.0
207.00	11161.0
208.00	3112.0
208.90	1060.0
210.10	1741.0
211.00	3367.0
212.00	516.0
212.90	268.0
214.80	1007.0
215.90	1946.0
216.90	21400.0
218.00	2905.0
219.10	222.0
221.00	12453.0
221.80	2807.0
222.90	4564.0
224.00	45168.0
224.90	11259.0
226.00	1567.0
227.00	20240.0
227.90	3025.0
229.00	4016.0
229.90	596.0
231.00	2106.0
231.90	336.0
233.00	277.0
233.90	1172.0
234.90	1257.0
235.90	808.0
236.90	2063.0
238.10	238.0

Instrument :
BNA_N
ClientSampleId :
DFTPP254

238.90	904.0
239.90	813.0
240.90	1103.0
242.00	2367.0
243.00	2609.0
244.00	34512.0
245.00	4724.0
245.90	6784.0
246.90	1289.0
248.00	434.0
248.90	1498.0
250.00	312.0
250.90	393.0
251.90	267.0
252.90	832.0
255.00	179392.0
256.00	26392.0
257.00	2463.0
258.00	10238.0
259.00	1770.0
259.90	358.0
260.80	224.0
264.00	338.0
265.00	4108.0
265.90	652.0
269.90	226.0
270.80	289.0
272.00	474.0
272.90	4331.0
274.00	13374.0
275.00	71680.0
276.00	10634.0
277.00	6507.0
277.80	1248.0
278.90	283.0
280.90	313.0
283.00	597.0
283.90	515.0
285.00	1124.0
286.00	207.0
289.00	232.0
289.80	242.0
290.90	157.0
291.90	167.0
292.90	1556.0
294.00	341.0
295.00	428.0
295.90	20904.0
297.00	2997.0
297.90	244.0
301.00	326.0
301.90	424.0
303.00	2430.0
303.90	687.0
307.90	382.0
308.90	159.0
309.90	393.0
312.70	151.0
314.00	781.0
314.90	2079.0
316.00	1609.0
317.00	270.0
320.90	493.0
322.00	374.0
323.00	6113.0
324.00	1233.0
326.90	1211.0
327.80	516.0
331.90	557.0
333.00	830.0
334.00	4179.0
334.90	1046.0
340.90	732.0
342.00	239.0
345.80	1287.0
346.90	377.0
352.00	1852.0
353.00	1190.0
354.00	2154.0
355.00	343.0
359.00	183.0
363.90	213.0
364.90	6702.0
365.90	1257.0
369.90	213.0
370.90	631.0
372.00	2777.0
372.90	736.0
382.90	862.0
384.00	199.0
384.80	174.0
389.90	483.0
390.90	331.0
391.80	213.0
402.00	1450.0
403.00	1884.0

Instrument :
BNA_N
ClientSampleId :
DFTPP254

403.90	830.0
404.80	216.0
420.90	1320.0
422.00	1522.0
423.00	10686.0
423.90	3369.0
424.90	534.0
441.00	29904.0
442.00	197568.0
443.00	41176.0
443.90	3828.0
444.90	273.0

Instrument :
BNA_N
ClientSampleId :
DFTPP254

m/z	Abundance
37.00	797.0
38.00	2145.0
39.00	14001.0
40.00	1269.0
41.00	557.0
43.10	304.0
43.90	649.0
44.90	321.0
49.10	1470.0
50.00	49776.0
51.00	195840.0
52.00	10129.0
53.00	371.0
54.90	965.0
56.00	5756.0
57.00	13392.0
57.90	676.0
58.90	205.0
59.90	186.0
61.00	2428.0
62.00	2809.0
63.00	7992.0
64.00	1011.0
65.00	3960.0
66.00	206.0
67.00	334.0
68.00	3088.0
69.00	186688.0
69.90	1317.0
73.00	1725.0
74.00	20016.0
75.00	30136.0
76.00	10099.0
77.00	215296.0
78.00	14812.0
79.00	15250.0
80.00	11127.0
80.90	16373.0
82.00	3947.0
83.00	3622.0
83.90	459.0
84.90	3638.0
86.00	4508.0
86.90	1934.0
88.00	784.0
88.90	323.0
91.00	3507.0
92.00	4163.0
93.00	26064.0
94.00	1758.0
95.10	444.0
96.00	1144.0
97.10	585.0
98.00	20552.0
99.00	15613.0
100.00	1704.0
101.00	9492.0
101.90	382.0
103.00	2985.0
104.00	5008.0
104.90	5362.0
106.10	2100.0
107.00	70776.0
108.00	11003.0
108.90	2358.0
110.00	126568.0
111.00	19032.0
112.00	2663.0
112.90	824.0
114.10	268.0
114.90	327.0
116.00	4040.0
117.00	60216.0
118.00	3814.0
118.90	664.0
120.00	784.0
120.90	420.0
121.90	4591.0
123.00	7406.0
124.00	3187.0
125.00	3066.0
126.00	1568.0
127.00	239488.0
128.00	18720.0
129.00	95648.0
129.90	7692.0
131.00	1538.0
132.00	763.0
133.10	332.0
133.90	2653.0
134.90	7740.0
136.00	2745.0

Instrument :
BNA_N
ClientSampleId :
DFTPP254

137.00	4424.0
137.90	973.0
138.90	432.0
140.00	1256.0
140.90	12935.0
142.00	4140.0
142.90	2614.0
143.90	693.0
144.90	690.0
145.90	2344.0
147.00	6335.0
147.90	13843.0
149.00	2826.0
149.90	746.0
151.00	1657.0
151.70	809.0
151.90	777.0
153.00	3780.0
154.00	2793.0
155.00	6679.0
156.00	10498.0
156.90	2002.0
157.90	2278.0
158.90	1630.0
160.00	4097.0
161.00	5293.0
162.00	1348.0
162.90	436.0
163.90	680.0
165.00	5243.0
166.00	3623.0
167.00	24448.0
168.00	12860.0
169.00	2356.0
169.90	820.0
170.90	1066.0
171.90	2354.0
173.00	2817.0
174.00	5336.0
175.00	9763.0
176.00	2473.0
176.90	4977.0
178.00	1437.0
178.90	17712.0
180.00	11895.0
180.90	5528.0
182.00	905.0
182.90	378.0
184.00	1634.0
185.00	8164.0
186.00	67032.0
187.00	19856.0
188.00	2514.0
189.00	4307.0
190.00	678.0
191.00	1941.0
191.90	5839.0
193.00	6347.0
194.00	1445.0
196.00	14235.0
197.90	450176.0
199.00	33688.0
200.00	2054.0
201.30	1726.0
202.90	3158.0
204.00	16400.0
205.00	28632.0
206.00	122624.0
207.00	16357.0
208.00	4182.0
208.90	1334.0
210.00	1929.0
210.90	4633.0
212.00	912.0
213.00	430.0
214.10	191.0
214.90	1393.0
216.00	2909.0
216.90	32528.0
217.90	3946.0
218.90	442.0
221.00	15903.0
221.70	4289.0
223.00	7564.0
224.00	66608.0
225.00	16254.0
226.10	1781.0
227.00	28288.0
227.90	4053.0
229.00	5690.0
230.00	732.0
231.00	2408.0
232.00	346.0
232.90	497.0
233.90	1768.0
234.90	1905.0

Instrument :
BNA_N
ClientSampleId :
DFTPP254

235.90	1418.0
237.00	2355.0
237.90	298.0
238.90	814.0
239.90	1161.0
241.00	1637.0
242.00	3784.0
243.00	3527.0
244.00	50360.0
245.00	6221.0
246.00	10539.0
246.90	2037.0
248.00	608.0
249.00	1669.0
249.90	347.0
251.00	403.0
252.00	442.0
252.90	1016.0
254.00	2710.0
254.90	238272.0
256.00	37336.0
257.00	3194.0
257.90	15670.0
259.00	2201.0
259.90	451.0
260.90	339.0
263.80	267.0
264.20	324.0
264.90	6371.0
266.00	717.0
269.90	219.0
271.00	573.0
272.00	904.0
273.00	7838.0
274.00	20648.0
275.00	110776.0
276.00	16161.0
277.00	10304.0
277.90	1801.0
279.10	349.0
280.80	239.0
282.90	957.0
284.00	989.0
285.00	1782.0
285.90	279.0
289.00	340.0
291.10	284.0
291.90	377.0
292.90	2305.0
294.00	521.0
296.00	33192.0
297.00	4213.0
297.90	392.0
300.90	406.0
301.90	638.0
303.00	3782.0
304.00	923.0
304.90	213.0
307.90	550.0
308.90	343.0
309.90	504.0
313.00	267.0
314.00	1300.0
314.90	3066.0
316.00	1959.0
317.00	309.0
321.00	982.0
322.00	582.0
323.00	9631.0
324.00	2275.0
324.90	238.0
325.90	229.0
327.00	1935.0
327.90	1133.0
331.90	848.0
332.90	1066.0
334.00	5848.0
334.90	1544.0
336.00	272.0
341.00	947.0
342.00	398.0
345.80	2107.0
347.00	429.0
351.00	175.0
351.90	3023.0
353.00	2071.0
354.00	3034.0
354.90	788.0
358.90	233.0
364.90	14435.0
366.00	1995.0
370.00	204.0
371.00	1024.0
372.00	5087.0
372.90	1103.0
383.00	1428.0

Instrument :
BNA_N
ClientSampleId :
DFTPP254

383.90	472.0
389.90	706.0
390.90	511.0
391.90	252.0
400.90	412.0
402.00	1863.0
402.90	2916.0
403.90	991.0
404.90	281.0
409.90	155.0
421.00	2984.0
422.00	2831.0
423.00	18664.0
424.00	5042.0
425.00	640.0
440.90	58240.0
442.00	391040.0
443.00	61624.0
443.90	6853.0
445.00	491.0

Instrument :
BNA_N
ClientSampleId :
DFTPP254

m/z	Abundance
36.90	534.0
38.00	1712.0
39.00	10837.0
40.00	1492.0
41.00	411.0
42.90	186.0
43.90	634.0
44.90	337.0
49.10	1210.0
50.00	38232.0
51.00	152704.0
52.00	7477.0
53.00	423.0
55.00	752.0
56.00	4450.0
57.00	10957.0
58.00	333.0
58.90	194.0
61.00	1778.0
62.00	2048.0
63.00	5768.0
63.90	986.0
65.00	2643.0
65.90	168.0
67.00	194.0
68.00	2642.0
69.00	145984.0
70.00	862.0
73.00	1228.0
74.00	15180.0
75.00	23760.0
76.00	8199.0
77.00	170880.0
78.00	10959.0
79.00	11136.0
79.90	8227.0
81.00	12663.0
81.90	2660.0
83.00	2673.0
83.90	431.0
85.00	2926.0
86.00	3915.0
87.00	1658.0
87.90	636.0
88.90	330.0
91.00	2770.0
92.00	3108.0
93.00	21608.0
93.90	1111.0
94.90	363.0
96.10	659.0
97.00	508.0
97.90	15830.0
98.90	13175.0
100.00	980.0
101.00	7068.0
102.00	345.0
103.00	2679.0
104.00	4523.0
105.00	4391.0
105.90	1530.0
107.00	56944.0
107.90	8763.0
109.00	1560.0
109.90	103152.0
111.00	14858.0
111.90	1917.0
113.00	609.0
114.80	313.0
116.00	2924.0
117.00	47296.0
118.00	3833.0
119.00	385.0
120.00	706.0
120.90	369.0
122.00	4090.0
123.00	5978.0
124.00	2561.0
124.90	2593.0
126.20	1781.0
127.00	194624.0
128.00	14401.0
129.00	78536.0
129.90	6508.0
131.00	1344.0
132.00	759.0
132.90	480.0
133.90	2530.0
135.00	6140.0
135.90	2816.0
136.90	3361.0
137.90	668.0

Instrument :
BNA_N
ClientSampleId :
DFTPP254

138.80	452.0
140.00	1003.0
140.90	9851.0
142.00	3124.0
142.90	2246.0
143.90	608.0
144.90	631.0
145.90	1880.0
147.00	5305.0
148.00	12205.0
148.90	2861.0
150.00	667.0
151.00	1226.0
151.70	954.0
152.90	3296.0
154.00	2496.0
155.00	6431.0
156.00	8330.0
157.00	1796.0
157.90	1673.0
159.00	1686.0
160.00	3164.0
160.90	4724.0
161.90	1198.0
163.00	419.0
164.00	474.0
164.90	4067.0
166.00	3214.0
167.00	22000.0
168.00	10931.0
169.00	1786.0
169.90	705.0
171.00	868.0
171.90	1877.0
173.00	2774.0
174.00	4223.0
175.00	8096.0
176.00	2234.0
176.90	3945.0
178.00	1277.0
178.90	15947.0
180.00	11166.0
180.90	4643.0
182.00	823.0
183.00	347.0
183.90	1488.0
185.00	7754.0
186.00	53680.0
187.00	16792.0
188.00	1551.0
188.90	3961.0
189.90	701.0
191.00	1889.0
191.90	4639.0
193.10	5164.0
194.00	1068.0
195.00	755.0
195.90	10576.0
197.90	405824.0
199.00	25640.0
200.00	2350.0
201.40	1648.0
203.00	3189.0
204.00	14810.0
205.00	24176.0
206.00	107928.0
207.00	13670.0
208.00	3158.0
208.90	1255.0
210.10	1639.0
210.90	4670.0
212.90	359.0
215.00	1380.0
216.00	2426.0
216.90	26928.0
218.00	4151.0
219.00	247.0
221.00	13780.0
221.80	3537.0
223.00	6188.0
224.00	58400.0
225.00	14685.0
226.10	1740.0
227.00	27040.0
228.00	4101.0
229.00	5150.0
229.90	723.0
231.00	1813.0
231.90	514.0
233.00	365.0
233.90	1929.0
235.00	1799.0
236.00	937.0
237.00	2232.0
237.80	444.0
239.00	891.0

Instrument :
BNA_N
ClientSampleId :
DFTPP254

240.00	851.0
240.80	1203.0
242.00	2773.0
242.90	2838.0
244.00	43112.0
245.00	6387.0
246.00	9571.0
247.00	1925.0
247.80	481.0
249.00	1588.0
249.90	315.0
250.80	363.0
251.90	409.0
253.00	1011.0
255.00	248640.0
256.00	31584.0
257.00	2698.0
257.90	13357.0
258.90	2384.0
260.00	401.0
260.90	334.0
264.00	334.0
264.90	5537.0
266.00	839.0
269.90	286.0
270.90	623.0
272.00	575.0
273.00	6277.0
274.00	18880.0
275.00	107424.0
276.00	14283.0
277.00	9951.0
277.90	1799.0
279.10	269.0
282.10	195.0
282.90	945.0
284.00	803.0
285.00	1393.0
285.90	232.0
288.90	343.0
290.00	255.0
290.90	213.0
292.00	394.0
292.90	1645.0
293.90	506.0
296.00	28904.0
297.00	4304.0
298.00	186.0
300.80	302.0
302.00	651.0
303.00	3581.0
304.00	1064.0
307.80	466.0
308.90	253.0
310.00	444.0
312.90	281.0
314.00	1486.0
315.00	3075.0
315.90	1660.0
316.90	295.0
321.00	962.0
322.10	462.0
323.00	9726.0
324.00	2246.0
326.00	184.0
326.90	1840.0
327.90	935.0
328.90	241.0
331.90	750.0
333.00	963.0
334.00	6030.0
335.00	1721.0
336.00	235.0
338.90	162.0
341.10	1100.0
341.90	343.0
345.90	1638.0
347.00	527.0
350.00	162.0
351.10	377.0
352.00	2725.0
352.90	2068.0
354.00	3493.0
354.80	570.0
359.00	221.0
364.10	255.0
364.90	15574.0
366.00	1978.0
366.90	166.0
370.10	206.0
371.00	986.0
372.00	6167.0
373.00	1689.0
373.80	190.0
382.90	1510.0
383.90	517.0

Instrument :
BNA_N
ClientSampleId :
DFTPP254

390.00	671.0
390.90	495.0
392.00	389.0
400.90	311.0
401.90	2206.0
403.00	3067.0
403.90	1208.0
404.90	257.0
420.90	2737.0
422.00	3038.0
423.00	20096.0
423.90	4694.0
425.00	607.0
441.00	57992.0
441.90	389824.0
442.90	80240.0
444.00	6738.0
444.90	385.0

Instrument :
BNA_N
ClientSampleId :
DFTPP254