

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061119\
 Data File : BN006247.D
 Acq On : 11 Jun 2019 15:03
 Operator : JU/SJ
 Sample : K3083-12
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 A51H4

Quant Time: Jun 11 15:38:55 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN053019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 11 12:32:25 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.65	152	4753	0.40	ng/ul	0.00
2) Naphthalene-d8	10.43	136	19773	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.30	164	10724	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.06	188	20685	0.40	ng/ul	0.00
16) Chrysene-d12	21.30	240	12665	0.40	ng/ul	0.03
20) Perylene-d12	23.56	264	13730	0.40	ng/ul	0.07

System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.03	152	8113	0.27	ng/ul	0.00
14) Fluoranthene-d10	19.10	212	16146	0.29	ng/ul	0.00

Target Compounds				Ovalue		
3) Naphthalene	10.48	128	54356	0.996	ng/ul	99
7) Acenaphthylene	14.02	152	27023	0.453	ng/ul	100
8) Acenaphthene	14.36	153	23185	0.561	ng/ul	99
9) Fluorene	15.36	166	19605	0.405	ng/ul	99
11) Pentachlorophenol	16.72	266	1819	0.346	ng/ul	97
12) Phenanthrene	17.10	178	32436	0.489	ng/ul	99
13) Anthracene	17.19	178	29085	0.459	ng/ul	99
15) Fluoranthene	19.13	202	34341	0.468	ng/ul	97
17) Pyrene	19.49	202	46425	0.696	ng/ul	100
18) Benzo(a)anthracene	21.28	228	28265	0.505	ng/ul	98
21) Benzo(b)fluoranthene	22.89	252	28137	0.491	ng/ul#	29
23) Benzo(a)pyrene	23.46	252	17540	0.329	ng/ul#	7
24) Indeno(1,2,3-cd)pyrene	25.80	276	38929	0.675	ng/ul#	96
25) Dibenzo(a,h)anthracene	25.81	278	29357	0.636	ng/ul	93
26) Benzo(g,h,i)perylene	26.49	276	34300	0.710	ng/ul	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061119\
Data File : BN006247.D
Acq On : 11 Jun 2019 15:03
Operator : JU/SJ
Sample : K3083-12
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
A51H4

Quant Time: Jun 11 15:38:55 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN053019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 11 12:32:25 2019
Response via : Initial Calibration

