


Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061119\
Data File : BN006269.D
Acq On : 12 Jun 2019 05:00
Operator : JU/SJ
Sample : K3215-13
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DB8J8

Manual Integrations
APPROVED

mohammad
6/12/2019 8:19:43 AM

Quant Time: Jun 12 07:34:49 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN053019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Jun 09 05:10:30 2019
Response via : Initial Calibration



Quantitation Report (Qedit)




Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061119\
Data File : BN006269.D
Acq On : 12 Jun 2019 05:00
Operator : JU/SJ
Sample : K3215-13
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampled :
DB8J8

Manual Integrations
APPROVED

mohammad
6/12/2019 8:19:43 AM

Quant Time: Jun 12 06:52:22 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN053019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Jun 09 05:10:30 2019
Response via : Initial Calibration



(2) Naphthalene-d8 (I)

10.554min (+0.127) 0.40ng/ul

response 1840231

Ion	Exp%	Act%
136.00	100	100
137.00	11.10	12.30
54.00	9.90	14.17#
68.00	8.30	38.83#

Quantitation Report (Qedit)




Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061119\
 Data File : BN006269.D
 Acq On : 12 Jun 2019 05:00
 Operator : JU/SJ
 Sample : K3215-13
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DB8J8

Manual Integrations
 APPROVED

mohammad
 6/12/2019 8:19:43 AM

Quant Time: Jun 12 06:52:22 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN053019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Jun 09 05:10:30 2019
 Response via : Initial Calibration



(2) Naphthalene-d8 (I)

10.439min (+0.011) 0.40ng/ul m *JU 06/12/19*

response 38661

Ion	Exp%	Act%
136.00	100	100
137.00	11.10	15.55#
54.00	9.90	84.21#
68.00	8.30	83.72#

Quantitation Report (Qedit)




Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061119\
 Data File : BN006269.D
 Acq On : 12 Jun 2019 05:00
 Operator : JU/SJ
 Sample : K3215-13
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DB8J8

Manual Integrations
 APPROVED

mohammad
 6/12/2019 8:19:43 AM

Quant Time: Jun 12 07:30:20 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN053019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Jun 09 05:10:30 2019
 Response via : Initial Calibration



(3) Naphthalene

10.494min (+0.022) 14.38ng/ul

response 1533684

Ion	Exp%	Act%
128.00	100	100
129.00	11.60	11.22
127.00	13.40	13.04
0.00	0.00	0.00

Quantitation Report (Qedit)




Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061119\
 Data File : BN006269.D
 Acq On : 12 Jun 2019 05:00
 Operator : JU/SJ
 Sample : K3215-13
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DB8J8

Manual Integrations
 APPROVED

mohammad
 6/12/2019 8:19:43 AM

Quant Time: Jun 12 07:30:20 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN053019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Jun 09 05:10:30 2019
 Response via : Initial Calibration



(3) Naphthalene

10.444min (-0.027) 68.79ng/ul m 06/12/19

response 7339235

Ion	Exp%	Act%
128.00	100	100
129.00	11.60	7.74#
127.00	13.40	1.48#
0.00	0.00	0.00

Quantitation Report (Qedit)




Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061119\
 Data File : BN006269.D
 Acq On : 12 Jun 2019 05:00
 Operator : JU/SJ
 Sample : K3215-13
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampled :
 DB8J8

Manual Integrations
 APPROVED

mohammad
 6/12/2019 8:19:43 AM

Quant Time: Jun 12 07:30:20 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN053019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Jun 09 05:10:30 2019
 Response via : Initial Calibration



(4) 2-Methylnaphthalene-d10 (SURR)

12.248min (+0.226) 0.71ng/ui

response 42155

Ion	Exp%	Act%
152.00	100	100
151.00	19.30	16.28
0.00	0.00	0.00
0.00	0.00	0.00

Quantitation Report (Qedit)




Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061119\
 Data File : BN006269.D
 Acq On : 12 Jun 2019 05:00
 Operator : JU/SJ
 Sample : K3215-13
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampled :
 DB8J8

Manual Integrations
 APPROVED

mohammad
 6/12/2019 8:19:43 AM

Quant Time: Jun 12 07:30:20 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN053019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Jun 09 05:10:30 2019
 Response via : Initial Calibration



(4) 2-Methylnaphthalene-d10 (SURR)

12.034min (+0.011) 0.33ng/ul m ⁵⁰
 06/12/19

response 19501

Ion	Exp%	Act%
152.00	100	100
151.00	19.30	35.19#
0.00	0.00	0.00
0.00	0.00	0.00

Quantitation Report (Qedit)




Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061119\
 Data File : BN006269.D
 Acq On : 12 Jun 2019 05:00
 Operator : JU/SJ
 Sample : K3215-13
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DB8J8

Manual Integrations
 APPROVED

mohammad
 6/12/2019 8:19:43 AM

Quant Time: Jun 12 07:30:20 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN053019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Jun 09 05:10:30 2019
 Response via : Initial Calibration



(5) 2-Methylnaphthalene

12.331min (+0.231) 6.10ng/ul

response 453483

Ion	Exp%	Act%
142.00	100	100
141.00	87.90	95.68
0.00	0.00	0.00
0.00	0.00	0.00

Quantitation Report (Qedit)




Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061119\
 Data File : BN006269.D
 Acq On : 12 Jun 2019 05:00
 Operator : JU/SJ
 Sample : K3215-13
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DB8J8

Manual Integrations
 APPROVED

mohammad
 6/12/2019 8:19:43 AM

Quant Time: Jun 12 07:30:20 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN053019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Jun 09 05:10:30 2019
 Response via : Initial Calibration



(5) 2-Methylnaphthalene

12.111min (+0.011) 8.30ng/ul m *JU* 06/12/19

response 617117

Ion	Exp%	Act%
142.00	100	100
141.00	87.90	70.31#
0.00	0.00	0.00
0.00	0.00	0.00

Quantitation Report (Qedit)




Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061119\
 Data File : BN006269.D
 Acq On : 12 Jun 2019 05:00
 Operator : JU/SJ
 Sample : K3215-13
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DB8J8

Manual Integrations
 APPROVED

mohammad
 6/12/2019 8:19:43 AM

Quant Time: Jun 12 07:30:20 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN053019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Jun 09 05:10:30 2019
 Response via : Initial Calibration



(8) Acenaphthene (C)

14.445min (+0.083) 0.05ng/ul

response 4294

Ion	Exp%	Act%
153.00	100	100
152.00	48.90	44.59
154.00	88.90	100.50
0.00	0.00	0.00

Quantitation Report (Qedit)




Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061119\
 Data File : BN006269.D
 Acq On : 12 Jun 2019 05:00
 Operator : JU/SJ
 Sample : K3215-13
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DB8J8

Manual Integrations
 APPROVED

mohammad
 6/12/2019 8:19:43 AM

Quant Time: Jun 12 07:30:20 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN053019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Jun 09 05:10:30 2019
 Response via : Initial Calibration



(8) Acenaphthene (C)

14.366min (+0.005) 0.51ng/ul m *JU* 06/12/19

response 43965

Ion	Exp%	Act%
153.00	100	100
152.00	48.90	29.14#
154.00	88.90	55.28#
0.00	0.00	0.00

Quantitation Report (Qedit)




Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061119\
 Data File : BN006269.D
 Acq On : 12 Jun 2019 05:00
 Operator : JU/SJ
 Sample : K3215-13
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DB8J8

Manual Integrations
 APPROVED

mohammad
 6/12/2019 8:19:43 AM

Quant Time: Jun 12 07:30:20 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN053019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Jun 09 05:10:30 2019
 Response via : Initial Calibration



(10) Phenanthrene-d10 (I)
 17.162min (+0.110) 0.40ng/ul
 response 2100651

Ion	Exp%	Act%
188.00	100	100
94.00	13.00	15.87#
80.00	13.60	14.39
0.00	0.00	0.00

Quantitation Report (Qedit)




Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061119\
 Data File : BN006269.D
 Acq On : 12 Jun 2019 05:00
 Operator : JU/SJ
 Sample : K3215-13
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DB8J8

Manual Integrations
 APPROVED

mohammad
 6/12/2019 8:19:43 AM

Quant Time: Jun 12 07:30:20 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN053019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Jun 09 05:10:30 2019
 Response via : Initial Calibration



(10) Phenanthrene-d10 (I)

17.065min (+0.013) 0.40ng/ul m *JU* 06/12/19

response 38593

Ion	Exp%	Act%
188.00	100	100
94.00	13.00	30.11#
80.00	13.60	30.78#
0.00	0.00	0.00

Quantitation Report (Qedit)




Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061119\
 Data File : BN006269.D
 Acq On : 12 Jun 2019 05:00
 Operator : JU/SJ
 Sample : K3215-13
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DB8J8

Manual Integrations
 APPROVED

mohammad
 6/12/2019 8:19:43 AM

Quant Time: Jun 12 07:33:05 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN053019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Jun 09 05:10:30 2019
 Response via : Initial Calibration



(11) Pentachlorophenol

16.765min (+0.051) 698.93ng/ul

response 6854085

Ion	Exp%	Act%
266.00	100	100
264.00	62.90	63.16
268.00	64.10	65.33
0.00	0.00	0.00

Quantitation Report (Qedit)




Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061119\
 Data File : BN006269.D
 Acq On : 12 Jun 2019 05:00
 Operator : JU/SJ
 Sample : K3215-13
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampled :
 DB8J8

Manual Integrations
 APPROVED

mohammad
 6/12/2019 8:19:43 AM

Quant Time: Jun 12 07:33:05 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN053019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Jun 09 05:10:30 2019
 Response via : Initial Calibration



(11) Pentachlorophenol

16.765min (+0.051) 878.42ng/ul m *Ju* 06/12/19

response 8614255

Ion	Exp%	Act%
266.00	100	100
264.00	62.90	50.25#
268.00	64.10	51.98
0.00	0.00	0.00

Quantitation Report (Qedit)




Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061119\
 Data File : BN006269.D
 Acq On : 12 Jun 2019 05:00
 Operator : JU/SJ
 Sample : K3215-13
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DB8J8

Manual Integrations
 APPROVED

mohammad
 6/12/2019 8:19:43 AM

Quant Time: Jun 12 07:33:05 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN053019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Jun 09 05:10:30 2019
 Response via : Initial Calibration



(14) Fluoranthene-d10 (SURR)

19.206min (+0.116) 0.00ng/ul

response 457

Ion	Exp%	Act%
212.00	100	100
106.00	18.10	0.00#
104.00	9.90	0.00#
0.00	0.00	0.00

Quantitation Report (Qedit)




Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061119\
 Data File : BN006269.D
 Acq On : 12 Jun 2019 05:00
 Operator : JU/SJ
 Sample : K3215-13
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DB8J8

Manual Integrations
 APPROVED

mohammad
 6/12/2019 8:19:43 AM

Quant Time: Jun 12 07:33:05 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN053019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Jun 09 05:10:30 2019
 Response via : Initial Calibration



(14) Fluoranthene-d10 (SURR)

19.095min (+0.005) 0.25ng/ul m *JU* 06/12/19

response 26354

Ion	Exp%	Act%
212.00	100	100
106.00	18.10	0.00#
104.00	9.90	0.00#
0.00	0.00	0.00

Quantitation Report (Qedit)




Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061119\
 Data File : BN006269.D
 Acq On : 12 Jun 2019 05:00
 Operator : JU/SJ
 Sample : K3215-13
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DB8J8

Manual Integrations
 APPROVED

mohammad
 6/12/2019 8:19:43 AM

Quant Time: Jun 12 07:33:05 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN053019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Jun 09 05:10:30 2019
 Response via : Initial Calibration



(16) Chrysene-d12 (I)

21.365min (+0.102) 0.40ng/ui

response 519

Ion	Exp%	Act%
240.00	100	100
120.00	20.10	151.04#
236.00	30.10	70.12#
0.00	0.00	0.00

Quantitation Report (Qedit)




Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061119\
 Data File : BN006269.D
 Acq On : 12 Jun 2019 05:00
 Operator : JU/SJ
 Sample : K3215-13
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampled :
 DB8J8

Manual Integrations
 APPROVED

mohammad
 6/12/2019 8:19:43 AM

Quant Time: Jun 12 07:33:05 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN053019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Jun 09 05:10:30 2019
 Response via : Initial Calibration



(16) Chrysene-d12 (I)

21.262min (+0.000) 0.40ng/ul m *Ju* 06/12/19

response 17643

Ion	Exp%	Act%
240.00	100	100
120.00	20.10	26.35#
236.00	30.10	31.68
0.00	0.00	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061119\
 Data File : BN006269.D
 Acq On : 12 Jun 2019 05:00
 Operator : JU/SJ
 Sample : K3215-13
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DB8J8

Manual Integrations
 APPROVED

mohammad
 6/12/2019 8:19:43 AM

Quant Time: Jun 12 07:34:49 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN053019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Jun 09 05:10:30 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.65	152	6620	0.40	ng/ul	0.00
2) Naphthalene-d8	10.44	136	38661m>	0.40	ng/ul	0.01
6) Acenaphthene-d10	14.30	164	22363	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.07	188	38593m>	0.40	ng/ul	0.01
16) Chrysene-d12	21.26	240	17643m>	0.40	ng/ul	0.00
20) Perylene-d12	23.49	264	25727	0.40	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Methylnaphthalene-d10	12.03	152	19501m	0.33	ng/ul	0.01
14) Fluoranthene-d10	19.09	212	26354m	0.25	ng/ul	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)	Ovalue
3) Naphthalene	10.44	128	7339235m>	68.794	ng/ul		
5) 2-Methylnaphthalene	12.11	142	617117m>	8.301	ng/ul		
8) Acenaphthene	14.37	153	43965m>	0.510	ng/ul		
9) Fluorene	15.36	166	46940	0.465	ng/ul	#	85
11) Pentachlorophenol	16.77	266	8614255m>	878.423	ng/ul		

(#) = qualifier out of range (m) = manual integration (+) = signals summed

JU 06/12/19