

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061119\
 Data File : BN006241.D
 Acq On : 11 Jun 2019 09:09
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTD0.409

Quant Time: Jun 11 12:09:51 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN053019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Jun 09 05:10:30 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	50	0.00
2 I	Naphthalene-d8	0.400	0.400	0.0	53	0.00
3	Naphthalene	0.400	0.413	-3.2	55	0.00
4 SURR	2-Methylnaphthalene-d10	0.400	0.377	5.8	50	0.00
5	2-Methylnaphthalene	0.400	0.376	6.0	50	0.00
6 I	Acenaphthene-d10	0.400	0.400	0.0	47	0.00
7	Acenaphthylene	0.400	0.384	4.0	45	0.00
8 C	Acenaphthene	0.400	0.411	-2.7	49	0.00
9	Fluorene	0.400	0.360	10.0	43	0.00
10 I	Phenanthrene-d10	0.400	0.400	0.0	39	0.00
11	Pentachlorophenol	0.400	0.314	21.5#	37	0.00
12	Phenanthrene	0.400	0.389	2.8	39	0.00
13	Anthracene	0.400	0.349	12.8	35	0.00
14 SURR	Fluoranthene-d10	0.400	0.398	0.5	39	0.00
15 C	Fluoranthene	0.400	0.396	1.0	39	0.00
16 I	Chrysene-d12	0.400	0.400	0.0	42	0.00
17	Pyrene	0.400	0.382	4.5	41	0.00
18	Benzo(a)anthracene	0.400	0.364	9.0	39	0.00
19	Chrysene	0.400	0.473	-18.2	50	0.00
20 I	Perylene-d12	0.400	0.400	0.0	55	0.01
21	Benzo(b)fluoranthene	0.400	0.372	7.0	52	0.01
22	Benzo(k)fluoranthene	0.400	0.417	-4.2	57	0.01
23 C	Benzo(a)pyrene	0.400	0.412	-3.0	57	0.00
24	Indeno(1,2,3-cd)pyrene	0.400	0.442	-10.5	61	0.01
25	Dibenzo(a,h)anthracene	0.400	0.433	-8.2	60	0.01
26	Benzo(g,h,i)perylene	0.400	0.452	-13.0	63	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0