

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061221\
 Data File : BN014891.D
 Acq On : 12 Jun 2021 13:00
 Operator : CG/JU
 Sample : SSTD16056
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :

Quant Time: Jun 14 01:51:42 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-BN061221MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 14 01:48:30 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.692	152	108500	20.000	ng/ul	0.00
20) Naphthalene-d8	10.475	136	397768	20.000	ng/ul	# 0.00
38) Acenaphthene-d10	14.345	164	277431	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.098	188	651986	20.000	ng/ul	0.00
79) Chrysene-d12	21.309	240	670902	20.000	ng/ul	0.01
88) Perylene-d12	23.556	264	857600	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.199	96	124634	46.778	ng/uL	0.00
4) Pyridine-d5	3.599	84	849892	126.658	ng/ul	-0.01
7) Phenol-d5	6.875	99	1260192	118.616	ng/ul	0.01
9) Bis-(2-Chloroethyl)eth...	7.034	67	611890	115.825	ng/ul	0.00
11) 2-Chlorophenol-d4	7.234	132	1015318	154.311	ng/ul	0.01
15) 4-Methylphenol-d8	8.416	113	1057553	128.032	ng/ul	0.02
21) Nitrobenzene-d5	8.851	128	500201	166.939	ng/ul	0.01
24) 2-Nitrophenol-d4	9.569	143	589477	183.028	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.110	165	1130786	153.989	ng/ul	0.01
31) 4-Chloroaniline-d4	10.622	131	1280684	147.932	ng/ul	0.01
46) Dimethylphthalate-d6	13.769	166	3339544	151.513	ng/ul	0.02
49) Acenaphthylene-d8	14.039	160	3743159	146.862	ng/ul	0.00
54) 4-Nitrophenol-d4	14.580	143	546126	155.970	ng/ul	0.02
60) Fluorene-d10	15.351	176	2690710	141.680	ng/ul	0.01
65) 4,6-Dinitro-2-methylph...	15.480	200	682101	191.275	ng/ul	0.02
73) Anthracene-d10	17.204	188	4213174	138.332	ng/ul	0.01
81) Pyrene-d10	19.509	212	4920217	155.536	ng/ul	0.01
92) Benzo(a)pyrene-d12	23.427	264	6006545	131.488	ng/ul	0.02
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.234	88	135689	43.377	ng/uL#	90
5) Pyridine	3.622	79	828628	117.847	ng/ul#	93
6) Benzaldehyde	6.834	77	477208	83.413	ng/ul	99
8) Phenol	6.904	94	1259725	115.235	ng/ul#	79
10) Bis(2-Chloroethyl)ether	7.128	93	935094	110.828	ng/ul	94
12) 2-Chlorophenol	7.263	128	1001328	145.400	ng/ul	97
13) 2-Methylphenol	8.139	108	983676	122.530	ng/ul	98
14) 2,2'-oxybis(1-Chloropr...	8.216	45	324963	61.053	ng/ul	94
16) Acetophenone	8.516	105	1665026	121.771	ng/ul#	96
17) N-Nitroso-di-n-propyla...	8.528	70	797668	121.114	ng/ul#	89
18) 4-Methylphenol	8.486	108	1042348	119.045	ng/ul	94
19) Hexachloroethane	8.763	117	509968	147.907	ng/ul	92
22) Nitrobenzene	8.892	77	1339496	143.243	ng/ul#	97
23) Isophorone	9.428	82	2308566	130.009	ng/ul#	93
25) 2-Nitrophenol	9.598	139	578085	168.902	ng/ul#	88
26) 2,4-Dimethylphenol	9.669	107	1503169	142.671	ng/ul	93
27) Bis(2-Chloroethoxy)met...	9.945	93	7400	0.686	ng/ul	93
29) 2,4-Dichlorophenol	10.139	162	1060783	150.622	ng/ul	96
30) Naphthalene	10.528	128	3028891	140.097	ng/ul	98
32) 4-Chloroaniline	10.645	127	1256438	136.368	ng/ul	96
33) Hexachlorobutadiene	10.816	225	975808	154.956	ng/ul	95
34) Caprolactam	11.451	113	165372	72.678	ng/ul#	75
35) 4-Chloro-3-methylphenol	11.792	107	1425948	139.720	ng/ul	95
36) 2-Methylnaphthalene	12.151	142	2250878	139.671	ng/ul	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061221\
 Data File : BN014891.D
 Acq On : 12 Jun 2021 13:00
 Operator : CG/JU
 Sample : SSTD16056
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :

Quant Time: Jun 14 01:51:42 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-BN061221MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 14 01:48:30 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.375	142	2203000	139.042	ng/ul#	94
39) 1,2,4,5-Tetrachloroben...	12.527	216	1393064	132.803	ng/ul#	94
40) Hexachlorocyclopentadiene	12.504	237	1140468	164.118	ng/ul	94
41) 2,4,6-Trichlorophenol	12.769	196	940320	159.635	ng/ul	94
42) 2,4,5-Trichlorophenol	12.851	196	1035583	161.094	ng/ul	97
43) 1,1'-Biphenyl	13.180	154	2907326	146.491	ng/ul#	94
44) 2-Chloronaphthalene	13.216	162	2346394	147.549	ng/ul	96
45) 2-Nitroaniline	13.427	65	838366	173.088	ng/ul	92
47) Dimethylphthalate	13.816	163	3157411	144.831	ng/ul	97
48) 2,6-Dinitrotoluene	13.939	165	680833	168.056	ng/ul	96
50) Acenaphthylene	14.069	152	3334355	142.288	ng/ul	99
51) 3-Nitroaniline	14.263	138	551594	140.320	ng/ul	90
52) Acenaphthene	14.416	153	2459164	144.830	ng/ul	98
53) 2,4-Dinitrophenol	14.469	184	476746	162.086	ng/ul#	76
55) 4-Nitrophenol	14.598	109	914414	136.225	ng/ul#	85
56) Dibenzofuran	14.751	168	3458990	136.943	ng/ul#	83
57) 2,4-Dinitrotoluene	14.727	165	1026029	158.922	ng/ul#	92
58) 2,3,4,6-Tetrachlorophenol	14.986	232	947594	143.298	ng/ul#	88
59) Diethylphthalate	15.186	149	3250246	155.083	ng/ul	96
61) Fluorene	15.404	166	2631870	138.495	ng/ul	97
62) 4-Chlorophenyl-phenyle...	15.398	204	1455507	129.462	ng/ul#	80
63) 4-Nitroaniline	15.521	138	608	0.152	ng/ul	94
66) 4,6-Dinitro-2-methylph...	15.498	198	629985	180.582	ng/ul#	76
67) N-Nitrosodiphenylamine	15.616	169	2459600	140.060	ng/ul	98
68) 4-Bromophenyl-phenylether	16.298	248	1103526	129.979	ng/ul#	77
69) Hexachlorobenzene	16.410	284	1344171	126.556	ng/ul#	87
70) Atrazine	16.580	200	1131377	155.803	ng/ul	99
71) Pentachlorophenol	16.757	266	879763	143.868	ng/ul	95
72) Phenanthrene	17.145	178	4737616	139.069	ng/ul	99
74) Anthracene	17.239	178	4622155	134.622	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.139	216	1480591	137.428	ng/ul	92
76) Pentachlorobenzene	14.674	250	1526898	137.031	ng/ul	96
77) Carbazole	17.510	167	4075488	140.841	ng/ul#	94
78) Di-n-butylphthalate	18.080	149	5358821	157.165	ng/ul	98
80) Fluoranthene	19.168	202	5888293	157.852	ng/ul#	94
82) Pyrene	19.539	202	5794059	151.041	ng/ul#	94
83) Butylbenzylphthalate	20.445	149	2524090	195.824	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.227	252	1642527	129.854	ng/ul#	95
85) Benzo(a)anthracene	21.292	228	5931530	141.833	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.227	149	3321353	178.768	ng/ul	97
87) Chrysene	21.345	228	5737637	136.238	ng/ul	100
89) Di-n-octyl phthalate	22.109	149	6851438	180.576	ng/ul	100
90) Benzo(b)fluoranthene	22.892	252	6892911	133.927	ng/ul	98
91) Benzo(k)fluoranthene	22.939	252	6622126	129.141	ng/ul	99
93) Benzo(a)pyrene	23.480	252	6022618	129.303	ng/ul	96
94) Indeno(1,2,3-cd)pyrene	25.862	276	8145155	127.630	ng/ul	94
95) Dibenzo(a,h)anthracene	25.880	278	6308790	119.451	ng/ul	98
96) Benzo(g,h,i)perylene	26.568	276	7512327	133.006	ng/ul	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061221\
 Data File : BN014891.D
 Acq On : 12 Jun 2021 13:00
 Operator : CG/JU
 Sample : SSTD16056
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampled :

Quant Time: Jun 14 01:51:42 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-BN061221MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 14 01:48:30 2021
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 43.377 ng/uL
RT: 3.234 min Scan# 25
Delta R.T. -0.006 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00




Instrument :
BNA_N
ClientSampled :


Tgt Ion: 88 Resp: 135689
Ion Ratio Lower Upper
88 100
43 22.5 13.8 20.6#
58 52.8 37.1 55.7



#5
Pyridine
Concen: 117.847 ng/uL
RT: 3.622 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00

Tgt Ion: 79 Resp: 828628
Ion Ratio Lower Upper
79 100
52 60.9 45.9 68.9
51 34.5 22.2 33.4#









#6
Benzaldehyde
Concen: 83.413 ng/ul
RT: 6.834 min Scan# 637
Delta R.T. -0.000 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00




Instrument :
BNA_N
ClientSampled :


Tgt Ion: 77 Resp: 477208
Ion Ratio Lower Upper
77 100
105 88.8 72.2 108.4
106 86.8 70.2 105.4



#8
Phenol
Concen: 115.235 ng/ul
RT: 6.904 min Scan# 649
Delta R.T. 0.011 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00

Tgt Ion: 94 Resp: 1259725
Ion Ratio Lower Upper
94 100
65 34.0 22.7 34.1
66 67.6 39.2 58.8#









#10
 Bis(2-Chloroethyl)ether
 Concen: 110.828 ng/ul
 RT: 7.128 min Scan# 687
 Delta R.T. 0.011 min
 Lab File: BN014891.D
 Acq: 12 Jun 2021 13:00




Instrument :
 BNA_N
 ClientSampleId :


Tgt Ion: 93 Resp: 935094
 Ion Ratio Lower Upper
 93 100
 63 60.7 45.3 67.9
 95 32.5 22.8 34.2



#12
 2-Chlorophenol
 Concen: 145.400 ng/ul
 RT: 7.263 min Scan# 710
 Delta R.T. 0.011 min
 Lab File: BN014891.D
 Acq: 12 Jun 2021 13:00

Tgt Ion:128 Resp: 1001328
 Ion Ratio Lower Upper
 128 100
 64 50.1 42.3 63.5
 130 32.3 26.1 39.1









#13
2-Methylphenol
Concen: 122.530 ng/ul
RT: 8.139 min Scan# 859
Delta R.T. 0.006 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00




Instrument :
BNA_N
ClientSampled :


Tgt Ion:108 Resp: 983676
Ion Ratio Lower Upper
108 100
107 94.1 74.1 111.1



#14
2,2'-oxybis(1-Chloropropane)
Concen: 61.053 ng/ul
RT: 8.216 min Scan# 872
Delta R.T. 0.006 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00

Tgt Ion: 45 Resp: 324963
Ion Ratio Lower Upper
45 100
77 35.8 25.8 38.8
79 24.0 21.4 32.2









#16
Acetophenone
Concen: 121.771 ng/ul
RT: 8.516 min Scan# 923
Delta R.T. 0.011 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:105 Resp: 1665026
Ion Ratio Lower Upper
105 100
77 86.9 70.8 106.2
51 25.0 15.6 23.4#



#17
N-Nitroso-di-n-propylamine
Concen: 121.114 ng/ul
RT: 8.528 min Scan# 925
Delta R.T. 0.035 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00

Tgt Ion: 70 Resp: 797668
Ion Ratio Lower Upper
70 100
42 41.2 27.1 40.7#
101 10.1 10.9 16.3#
130 19.1 11.8 17.8#









#18
4-Methylphenol
Concen: 119.045 ng/ul
RT: 8.486 min Scan# 918
Delta R.T. 0.029 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:108 Resp: 1042348
Ion Ratio Lower Upper
108 100
107 110.1 93.4 140.0



#19
Hexachloroethane
Concen: 147.907 ng/ul
RT: 8.763 min Scan# 965
Delta R.T. -0.000 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00

Tgt Ion:117 Resp: 509968
Ion Ratio Lower Upper
117 100
201 112.7 91.0 136.4
199 71.4 70.1 105.1









#22
Nitrobenzene
Concen: 143.243 ng/ul
RT: 8.892 min Scan# 987
Delta R.T. 0.011 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00




Instrument :
BNA_N
ClientSampleId :


Tgt Ion: 77 Resp: 1339496
Ion Ratio Lower Upper
77 100
123 36.6 29.4 44.2
65 16.7 9.7 14.5#



#23
Isophorone
Concen: 130.009 ng/ul
RT: 9.428 min Scan# 1078
Delta R.T. 0.023 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00

Tgt Ion: 82 Resp: 2308566
Ion Ratio Lower Upper
82 100
95 12.7 7.7 11.5#
138 16.3 11.0 16.4









#25
2-Nitrophenol
Concen: 168.902 ng/ul
RT: 9.598 min Scan# 1107
Delta R.T. 0.006 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00




Instrument :
BNA_N
ClientSampled :


Tgt Ion:139 Resp: 578085
Ion Ratio Lower Upper
139 100
65 64.4 55.1 82.7
109 76.6 49.9 74.9#



#26
2,4-Dimethylphenol
Concen: 142.671 ng/ul
RT: 9.669 min Scan# 1119
Delta R.T. 0.011 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00

Tgt Ion:107 Resp: 1503169
Ion Ratio Lower Upper
107 100
121 40.4 31.8 47.6
122 63.5 57.5 86.3









#27

Bis(2-Chloroethoxy)methane
Concen: 0.686 ng/ul
RT: 9.945 min Scan# 1166
Delta R.T. 0.053 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00

Instrument :
BNA_N
ClientSampled :




Tgt Ion: 93	Resp:	7400
Ion Ratio	Lower	Upper
93	100	
95	37.5	26.4 39.6
123	8.6	6.4 9.6




#29

2,4-Dichlorophenol
Concen: 150.622 ng/ul
RT: 10.139 min Scan# 1199
Delta R.T. 0.011 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00

Tgt Ion: 162	Resp:	1060783
Ion Ratio	Lower	Upper
162	100	
164	63.0	48.8 73.2
98	45.8	40.3 60.5









#30
Naphthalene
Concen: 140.097 ng/ul
RT: 10.528 min Scan# 1265
Delta R.T. 0.006 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00




Instrument :
BNA_N
ClientSampled :


Tgt Ion:128 Resp: 3028891
Ion Ratio Lower Upper
128 100
129 11.3 8.1 12.1
127 14.2 11.0 16.4



#32
4-Chloroaniline
Concen: 136.368 ng/ul
RT: 10.645 min Scan# 1285
Delta R.T. 0.011 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00


Tgt Ion:127 Resp: 1256438
Ion Ratio Lower Upper
127 100
129 31.5 27.0 40.4








#33
Hexachlorobutadiene
Concen: 154.956 ng/ul
RT: 10.816 min Scan# 1314
Delta R.T. -0.000 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00


Instrument :
BNA_N
ClientSampled :





Tgt Ion:225 Resp: 975808
Ion Ratio Lower Upper
225 100
223 61.5 53.3 79.9
227 64.1 53.3 79.9




#34
Caprolactam
Concen: 72.678 ng/ul
RT: 11.451 min Scan# 1422
Delta R.T. 0.058 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00



Tgt Ion:113 Resp: 165372
Ion Ratio Lower Upper
113 100
55 96.9 94.8 142.2
56 101.6 58.6 87.8#









#35
 4-Chloro-3-methylphenol
 Concen: 139.720 ng/ul
 RT: 11.792 min Scan# 1480
 Delta R.T. 0.023 min
 Lab File: BN014891.D
 Acq: 12 Jun 2021 13:00




Instrument :
 BNA_N
 ClientSampled :


Tgt Ion:107 Resp: 1425948
 Ion Ratio Lower Upper
 107 100
 144 17.9 13.4 20.2
 142 57.5 42.6 63.8



#36
 2-Methylnaphthalene
 Concen: 139.671 ng/ul
 RT: 12.151 min Scan# 1541
 Delta R.T. 0.006 min
 Lab File: BN014891.D
 Acq: 12 Jun 2021 13:00

Tgt Ion:142 Resp: 2250878
 Ion Ratio Lower Upper
 142 100
 141 89.0 73.6 110.4









#37
1-Methylnaphthalene
Concen: 139.042 ng/ul
RT: 12.375 min Scan# 1579
Delta R.T. 0.006 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00




Instrument :
BNA_N
ClientSampled :


Tgt Ion:142 Resp: 2203000
Ion Ratio Lower Upper
142 100
141 95.7 72.0 108.0
116 6.9 3.8 5.6#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 132.803 ng/ul
RT: 12.527 min Scan# 1605
Delta R.T. 0.006 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00

Tgt Ion:216 Resp: 1393064
Ion Ratio Lower Upper
216 100
214 79.8 65.8 98.8
179 27.0 16.7 25.1#
108 20.4 12.6 18.8#









#40
Hexachlorocyclopentadiene
Concen: 164.118 ng/ul
RT: 12.504 min Scan# 1601
Delta R.T. 0.006 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00




Instrument :
BNA_N
ClientSampled :


Tgt Ion:237 Resp: 1140468
Ion Ratio Lower Upper
237 100
235 57.1 49.8 74.6
272 11.1 9.9 14.9



#41
2,4,6-Trichlorophenol
Concen: 159.635 ng/ul
RT: 12.769 min Scan# 1646
Delta R.T. 0.006 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00

Tgt Ion:196 Resp: 940320
Ion Ratio Lower Upper
196 100
198 96.0 81.4 122.2
200 30.6 26.7 40.1









#42
2,4,5-Trichlorophenol
Concen: 161.094 ng/ul
RT: 12.851 min Scan# 1660
Delta R.T. 0.011 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00




Instrument :
BNA_N
ClientSampled :


Tgt Ion:196 Resp: 1035583
Ion Ratio Lower Upper
196 100
198 94.4 77.8 116.8
200 29.7 24.3 36.5



#43
1,1'-Biphenyl
Concen: 146.491 ng/ul
RT: 13.180 min Scan# 1716
Delta R.T. 0.011 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00

Tgt Ion:154 Resp: 2907326
Ion Ratio Lower Upper
154 100
153 42.9 32.2 48.4
76 11.4 12.3 18.5#









#44
2-Chloronaphthalene
Concen: 147.549 ng/ul
RT: 13.216 min Scan# 1722
Delta R.T. 0.011 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00




Instrument :
BNA_N
ClientSampled :


Tgt Ion:162 Resp: 2346394
Ion Ratio Lower Upper
162 100
164 34.3 25.9 38.9
127 52.2 39.1 58.7



#45
2-Nitroaniline
Concen: 173.088 ng/ul
RT: 13.427 min Scan# 1758
Delta R.T. 0.011 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00

Tgt Ion: 65 Resp: 838366
Ion Ratio Lower Upper
65 100
92 63.2 60.5 90.7
138 86.0 67.0 100.6





#47

Dimethylphthalate

Concen: 144.831 ng/ul

RT: 13.816 min Scan# 1824

Delta R.T. 0.017 min

Lab File: BN014891.D

Acq: 12 Jun 2021 13:00





Tgt Ion:163 Resp: 3157411

Ion Ratio Lower Upper

163 100

194 4.2 3.8 5.8

164 9.4 8.7 13.1



#48

2,6-Dinitrotoluene

Concen: 168.056 ng/ul

RT: 13.939 min Scan# 1845

Delta R.T. 0.023 min

Lab File: BN014891.D

Acq: 12 Jun 2021 13:00




Tgt Ion:165 Resp: 680833


Ion Ratio Lower Upper

165 100

89 82.2 63.0 94.6

121 21.6 15.8 23.8









#50
Acenaphthylene
Concen: 142.288 ng/ul
RT: 14.069 min Scan# 1867
Delta R.T. 0.006 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:152 Resp: 3334355
Ion Ratio Lower Upper
152 100
151 22.4 17.8 26.8
153 12.9 11.2 16.8



#51
3-Nitroaniline
Concen: 140.320 ng/ul
RT: 14.263 min Scan# 1900
Delta R.T. 0.017 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00

Tgt Ion:138 Resp: 551594
Ion Ratio Lower Upper
138 100
108 19.1 17.0 25.4
92 152.8 133.6 200.4









#52
Acenaphthene
Concen: 144.830 ng/ul
RT: 14.416 min Scan# 1926
Delta R.T. 0.011 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00




Instrument :
BNA_N
ClientSampled :


Tgt Ion:153 Resp: 2459164
Ion Ratio Lower Upper
153 100
152 49.3 37.7 56.5
154 83.6 67.8 101.6



#53
2,4-Dinitrophenol
Concen: 162.086 ng/ul
RT: 14.469 min Scan# 1935
Delta R.T. 0.006 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00

Tgt Ion:184 Resp: 476746
Ion Ratio Lower Upper
184 100
63 81.9 54.3 81.5#
154 115.3 70.0 105.0#









#55
4-Nitrophenol
Concen: 136.225 ng/ul
RT: 14.598 min Scan# 1957
Delta R.T. 0.029 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:109 Resp: 914414
Ion Ratio Lower Upper
109 100
139 58.7 59.8 89.8#
65 74.6 68.6 102.8



#56
Dibenzofuran
Concen: 136.943 ng/ul
RT: 14.751 min Scan# 1983
Delta R.T. 0.006 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00

Tgt Ion:168 Resp: 3458990
Ion Ratio Lower Upper
168 100
139 47.0 29.5 44.3#









#57
2,4-Dinitrotoluene
Concen: 158.922 ng/ul
RT: 14.727 min Scan# 1979
Delta R.T. 0.017 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:165 Resp: 1026029
Ion Ratio Lower Upper
165 100
63 52.7 37.6 56.4
182 2.9 1.8 2.6#



#58
2,3,4,6-Tetrachlorophenol
Concen: 143.298 ng/ul
RT: 14.986 min Scan# 2023
Delta R.T. 0.011 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00

Tgt Ion:232 Resp: 947594
Ion Ratio Lower Upper
232 100
131 42.9 27.8 41.8#
130 2.6 1.7 2.5#
166 30.8 20.6 30.8









#59
Diethylphthalate
Concen: 155.083 ng/ul
RT: 15.186 min Scan# 2057
Delta R.T. 0.011 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00




Instrument :
BNA_N
ClientSampled :


Tgt Ion:149 Resp: 3250246
Ion Ratio Lower Upper
149 100
177 23.3 20.6 31.0
150 13.8 11.4 17.2



#61
Fluorene
Concen: 138.495 ng/ul
RT: 15.404 min Scan# 2094
Delta R.T. 0.011 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00

Tgt Ion:166 Resp: 2631870
Ion Ratio Lower Upper
166 100
165 96.0 74.6 111.8
167 14.2 11.1 16.7









#62
4-Chlorophenyl-phenylether
Concen: 129.462 ng/ul
RT: 15.398 min Scan# 2093
Delta R.T. 0.006 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00




Instrument :
BNA_N
ClientSampled :


Tgt Ion:204 Resp: 1455507
Ion Ratio Lower Upper
204 100
206 31.9 27.2 40.8
141 87.7 52.1 78.1#



#63
4-Nitroaniline
Concen: 0.152 ng/ul
RT: 15.521 min Scan# 2114
Delta R.T. 0.106 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00

Tgt Ion:138 Resp: 608
Ion Ratio Lower Upper
138 100
92 74.8 57.5 86.3
108 145.7 109.3 163.9









#66
 4,6-Dinitro-2-methylphenol
 Concen: 180.582 ng/ul
 RT: 15.498 min Scan# 2110
 Delta R.T. 0.023 min
 Lab File: BN014891.D
 Acq: 12 Jun 2021 13:00




Instrument :
 BNA_N
 ClientSampled :


Tgt Ion:198 Resp: 629985
 Ion Ratio Lower Upper
 198 100
 182 4.0 4.1 6.1#
 77 40.4 20.8 31.2#



#67
 N-Nitrosodiphenylamine
 Concen: 140.060 ng/ul
 RT: 15.616 min Scan# 2130
 Delta R.T. 0.006 min
 Lab File: BN014891.D
 Acq: 12 Jun 2021 13:00

Tgt Ion:169 Resp: 2459600
 Ion Ratio Lower Upper
 169 100
 168 70.8 54.7 82.1
 167 38.1 30.6 46.0









#68
4-Bromophenyl-phenylether
Concen: 129.979 ng/ul
RT: 16.298 min Scan# 2246
Delta R.T. 0.006 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00




Instrument :
BNA_N
ClientSampled :


Tgt Ion:248 Resp: 1103526
Ion Ratio Lower Upper
248 100
250 98.0 74.2 111.2
141 117.4 63.4 95.2#



#69
Hexachlorobenzene
Concen: 126.556 ng/ul
RT: 16.410 min Scan# 2265
Delta R.T. 0.006 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00

Tgt Ion:284 Resp: 1344171
Ion Ratio Lower Upper
284 100
142 36.4 20.6 30.8#
249 38.2 27.3 40.9









#70
 Atrazine
 Concen: 155.803 ng/ul
 RT: 16.580 min Scan# 2294
 Delta R.T. 0.017 min
 Lab File: BN014891.D
 Acq: 12 Jun 2021 13:00

Instrument :
 BNA_N
 ClientSampled :

Tgt Ion:200 Resp: 1131377




Ion	Ratio	Lower	Upper
200	100		
173	27.4	22.0	33.0
215	49.0	38.3	57.5




#71
 Pentachlorophenol
 Concen: 143.868 ng/ul
 RT: 16.757 min Scan# 2324
 Delta R.T. 0.006 min
 Lab File: BN014891.D
 Acq: 12 Jun 2021 13:00

Tgt Ion:266 Resp: 879763

Ion	Ratio	Lower	Upper
266	100		
264	64.4	47.8	71.6
268	63.7	48.4	72.6









#72
Phenanthrene
Concen: 139.069 ng/u1
RT: 17.145 min Scan# 2390
Delta R.T. 0.006 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00




Instrument :
BNA_N
ClientSampled :


Tgt Ion:178 Resp: 4737616
Ion Ratio Lower Upper
178 100
179 15.7 12.9 19.3
176 19.8 15.4 23.0



#74
Anthracene
Concen: 134.622 ng/u1
RT: 17.239 min Scan# 2406
Delta R.T. 0.011 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00

Tgt Ion:178 Resp: 4622155
Ion Ratio Lower Upper
178 100
179 15.4 12.9 19.3
176 20.5 15.3 22.9









#75
1,2,3,4-Tetrachlorobenzene
Concen: 137.428 ng/uL
RT: 13.139 min Scan# 1709
Delta R.T. 0.006 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00




Instrument :
BNA_N
ClientSampled :


Tgt Ion:216 Resp: 1480591
Ion Ratio Lower Upper
216 100
214 79.2 58.6 87.8
218 47.4 43.5 65.3



#76
Pentachlorobenzene
Concen: 137.031 ng/uL
RT: 14.674 min Scan# 1970
Delta R.T. 0.011 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00

Tgt Ion:250 Resp: 1526898
Ion Ratio Lower Upper
250 100
248 57.7 49.5 74.3
252 63.2 51.7 77.5









#77
Carbazole
Concen: 140.841 ng/ul
RT: 17.510 min Scan# 2452
Delta R.T. 0.006 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00




Instrument :
BNA_N
ClientSampled :


Tgt Ion:167 Resp: 4075488
Ion Ratio Lower Upper
167 100
166 23.9 17.4 26.0
139 16.2 10.6 15.8#



#78
Di-n-butylphthalate
Concen: 157.165 ng/ul
RT: 18.080 min Scan# 2549
Delta R.T. 0.006 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00

Tgt Ion:149 Resp: 5358821
Ion Ratio Lower Upper
149 100
150 9.0 6.5 9.7
104 7.2 5.6 8.4









#80
Fluoranthene
Concen: 157.852 ng/ul
RT: 19.168 min Scan# 2734
Delta R.T. 0.006 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00




Instrument :
BNA_N
ClientSampled :


Tgt Ion:202 Resp: 5888293
Ion Ratio Lower Upper
202 100
101 6.9 7.6 11.4#
100 5.7 5.6 8.4



#82
Pyrene
Concen: 151.041 ng/ul
RT: 19.539 min Scan# 2797
Delta R.T. 0.011 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00

Tgt Ion:202 Resp: 5794059
Ion Ratio Lower Upper
202 100
101 7.7 8.7 13.1#
100 6.7 6.2 9.4









#83
Butylbenzylphthalate
Concen: 195.824 ng/ul
RT: 20.445 min Scan# 2951
Delta R.T. 0.006 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00




Instrument :
BNA_N
ClientSampled :


Tgt Ion:149 Resp: 2524090
Ion Ratio Lower Upper
149 100
91 77.9 65.7 98.5
206 25.6 23.8 35.8



#84
3,3'-Dichlorobenzidine
Concen: 129.854 ng/ul
RT: 21.227 min Scan# 3084
Delta R.T. 0.006 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00

Tgt Ion:252 Resp: 1642527
Ion Ratio Lower Upper
252 100
254 66.6 50.1 75.1
126 9.8 6.3 9.5#









#85
Benzo(a)anthracene
Concen: 141.833 ng/ul
RT: 21.292 min Scan# 3095
Delta R.T. 0.011 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00




Instrument :
BNA_N
ClientSampled :


Tgt Ion:228 Resp: 5931530
Ion Ratio Lower Upper
228 100
229 19.4 16.2 24.4
226 28.3 21.6 32.4



#86
Bis(2-ethylhexyl)phthalate
Concen: 178.768 ng/ul
RT: 21.227 min Scan# 3084
Delta R.T. 0.006 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00

Tgt Ion:149 Resp: 3321353
Ion Ratio Lower Upper
149 100
167 36.3 30.7 46.1
279 5.8 5.1 7.7









#87
 Chrysene
 Concen: 136.238 ng/ul
 RT: 21.345 min Scan# 3104
 Delta R.T. 0.011 min
 Lab File: BN014891.D
 Acq: 12 Jun 2021 13:00




Instrument :
 BNA_N
 ClientSampleId :


Tgt Ion:228 Resp: 5737637
 Ion Ratio Lower Upper
 228 100
 226 29.8 23.8 35.6
 229 18.4 14.8 22.2



#89
 Di-n-octyl phthalate
 Concen: 180.576 ng/ul
 RT: 22.109 min Scan# 3234
 Delta R.T. 0.006 min
 Lab File: BN014891.D
 Acq: 12 Jun 2021 13:00

Tgt Ion:149 Resp: 6851438









#90
Benzo(b)fluoranthene
Concen: 133.927 ng/ul
RT: 22.892 min Scan# 3367
Delta R.T. 0.017 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00




Instrument :
BNA_N
ClientSampleId :


Tgt Ion:252 Resp: 6892911
Ion Ratio Lower Upper
252 100
253 21.1 15.8 23.8
125 5.2 4.3 6.5



#91
Benzo(k)fluoranthene
Concen: 129.141 ng/ul
RT: 22.939 min Scan# 3375
Delta R.T. 0.017 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00

Tgt Ion:252 Resp: 6622126
Ion Ratio Lower Upper
252 100
253 20.6 17.0 25.4
125 5.7 4.8 7.2









#93
Benzo(a)pyrene
Concen: 129.303 ng/ul
RT: 23.480 min Scan# 3467
Delta R.T. 0.023 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00




Instrument :
BNA_N
ClientSampled :


Tgt Ion:252 Resp: 6022618
Ion Ratio Lower Upper
252 100
253 21.6 18.8 28.2
125 5.8 5.4 8.2



#94
Indeno(1,2,3-cd)pyrene
Concen: 127.630 ng/ul
RT: 25.862 min Scan# 3872
Delta R.T. 0.041 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00

Tgt Ion:276 Resp: 8145155
Ion Ratio Lower Upper
276 100
138 14.4 10.0 15.0
277 27.6 19.5 29.3









#95
Dibenzo(a,h)anthracene
Concen: 119.451 ng/ul
RT: 25.880 min Scan# 3875
Delta R.T. 0.041 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00

Instrument :
BNA_N
ClientSampled :

Tgt Ion:278 Resp: 6308790
Ion Ratio Lower Upper
278 100
139 9.4 7.6 11.4
279 25.0 19.0 28.6



#96
Benzo(g,h,i)perylene
Concen: 133.006 ng/ul
RT: 26.568 min Scan# 3992
Delta R.T. 0.047 min
Lab File: BN014891.D
Acq: 12 Jun 2021 13:00

Tgt Ion:276 Resp: 7512327
Ion Ratio Lower Upper
276 100
138 11.0 10.2 15.4
277 23.1 15.7 23.5

