


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061221\
 Data File : BN014884.D
 Acq On : 12 Jun 2021 08:42
 Operator : CG/JU
 Sample : DFTPP99
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP099

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-BN061221MA.M
 Title : SVOA CALIBRATION
 Last Update : Mon Jun 14 10:46:06 2021



AutoFind: Scans 2459, 2460, 2461; Background Corrected with Scan 2453

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	16.5	94227	PASS
68	69	0.00	2	1.3	1573	PASS
69	198	0.00	100	21.7	123995	PASS
70	69	0.00	2	0.9	1153	PASS
127	198	10	80	27.1	154325	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	570432	PASS
199	198	5	9	6.9	39088	PASS
275	198	10	60	25.8	147264	PASS
365	198	1	100	2.9	16790	PASS
441	198	0.01	100	9.4	53693	PASS
442	198	50	100	60.0	342144	PASS
443	442	15	24	19.3	66013	PASS

m/z	Abundance
39.80	304.0
43.80	1015.0
48.80	334.0
50.90	618.0
72.90	1188.0
84.00	435.0
190.90	362.0
192.80	585.0
207.00	1731.0
220.90	360.0
280.70	384.0
281.10	692.0
282.90	376.0
354.80	360.0

Instrument :
BNA_N
ClientSampleId :
DFTPP099

m/z	Abundance
36.80	610.0
37.40	537.0
38.10	1899.0
39.00	7090.0
41.10	316.0
43.70	399.0
48.90	1949.0
50.00	33296.0
51.00	114840.0
52.00	5605.0
52.90	507.0
54.40	621.0
55.00	1045.0
56.00	3491.0
56.90	5571.0
61.00	2108.0
62.00	2211.0
63.00	4532.0
64.00	875.0
65.20	1531.0
67.00	610.0
68.10	4719.0
69.00	136512.0
70.00	1103.0
71.10	335.0
73.10	2518.0
74.00	22448.0
75.00	26456.0
76.10	10101.0
77.00	142016.0
78.10	9206.0
79.00	19264.0
80.00	10510.0
81.00	15943.0
81.90	3264.0
83.10	2558.0
85.00	2815.0
86.00	4222.0
87.00	2573.0
87.70	942.0
91.00	3690.0
92.00	5003.0
93.00	25480.0
94.00	1179.0
95.10	555.0
95.90	815.0
96.90	1377.0
98.00	20712.0
99.00	13738.0
100.00	1213.0
100.90	7381.0
102.20	329.0
102.80	2902.0
104.00	6275.0
105.00	6879.0
106.10	2798.0
107.00	70568.0
107.90	6156.0
108.20	6062.0
108.80	2160.0
109.90	125168.0
111.00	17184.0
112.00	2219.0
112.80	583.0
115.90	4781.0
116.90	64120.0
117.90	4753.0
118.90	1522.0
119.90	901.0
122.10	5631.0
123.10	5476.0
124.00	2910.0
125.00	2768.0
127.00	159488.0
128.00	13247.0
129.00	71336.0
130.00	5868.0
130.90	360.0
132.00	388.0
133.10	893.0
134.00	3903.0
135.00	9701.0
136.00	3085.0
137.00	4100.0
138.30	708.0
139.10	941.0
140.00	1140.0
140.90	7545.0
142.00	1862.0
142.90	2693.0
144.00	613.0
145.10	870.0

Instrument :
BNA_N
ClientSampleId :
DFTPP099

145.90	2405.0
146.90	5394.0
148.00	11838.0
148.80	2464.0
149.80	586.0
150.90	1831.0
151.60	681.0
152.80	2849.0
153.80	2470.0
154.90	7156.0
156.00	8620.0
156.80	1973.0
157.90	2742.0
158.80	1575.0
159.90	3538.0
160.90	5677.0
162.00	1513.0
163.20	1181.0
163.90	971.0
165.00	4419.0
165.90	2634.0
167.00	23336.0
168.00	10253.0
168.90	1834.0
169.90	726.0
171.00	1170.0
172.10	1766.0
173.20	2687.0
174.10	5274.0
175.10	6256.0
176.10	2546.0
177.10	4507.0
177.70	1683.0
178.20	1185.0
178.90	15013.0
180.00	11198.0
180.90	4734.0
181.90	799.0
182.90	406.0
184.10	1964.0
185.00	10695.0
186.00	76064.0
187.00	23392.0
188.00	2911.0
189.00	5115.0
189.80	1221.0
191.00	1730.0
192.00	6812.0
193.10	5625.0
193.80	785.0
194.30	1211.0
195.00	1310.0
196.10	11758.0
197.90	570944.0
199.00	39056.0
200.00	3275.0
201.40	1104.0
202.90	4850.0
204.00	16145.0
205.00	28832.0
206.10	121720.0
207.00	13119.0
208.00	4028.0
208.90	3016.0
209.80	445.0
211.10	5506.0
212.90	841.0
214.90	2090.0
215.90	2592.0
217.00	27296.0
218.10	4615.0
221.00	8805.0
221.70	2933.0
222.10	3002.0
223.00	6394.0
224.00	70064.0
225.00	23696.0
226.20	4014.0
227.00	33776.0
227.90	5409.0
229.10	5840.0
230.10	863.0
231.10	2304.0
232.20	1198.0
232.90	640.0
234.10	1585.0
234.90	3184.0
235.90	1484.0
237.00	2397.0
238.30	576.0
239.00	1734.0
239.80	2619.0
240.30	499.0
241.00	1527.0
242.10	3157.0
243.00	5080.0

Instrument :
BNA_N
ClientSampleId :
DFTPP099

244.00	52512.0
245.00	6392.0
245.90	14122.0
246.90	2551.0
248.00	458.0
248.80	2946.0
249.80	435.0
250.70	937.0
252.10	1741.0
252.80	1745.0
253.60	2919.0
255.00	299584.0
256.00	40864.0
257.10	3524.0
258.00	16512.0
259.10	5115.0
259.90	687.0
261.10	1312.0
263.80	856.0
265.00	7350.0
265.80	547.0
266.20	597.0
266.90	623.0
267.90	316.0
270.90	685.0
271.20	546.0
271.90	745.0
273.10	12437.0
274.00	34720.0
275.00	141504.0
276.00	19872.0
277.00	17992.0
278.10	3460.0
281.90	480.0
283.10	2137.0
284.10	1341.0
285.00	2572.0
286.10	358.0
287.80	636.0
289.10	566.0
289.70	761.0
291.90	457.0
293.00	2853.0
293.80	642.0
294.30	985.0
294.70	838.0
295.10	894.0
296.00	39328.0
297.00	7877.0
300.90	907.0
302.00	1245.0
303.00	5818.0
304.00	1911.0
307.80	729.0
309.50	342.0
310.00	571.0
311.70	527.0
313.30	670.0
314.20	3171.0
315.00	5174.0
316.10	3620.0
317.10	962.0
320.90	835.0
321.90	855.0
323.00	13451.0
323.90	1503.0
327.10	2300.0
328.10	922.0
332.00	1094.0
333.00	1175.0
334.00	8072.0
335.00	1182.0
335.80	766.0
340.80	1463.0
342.00	334.0
346.10	2069.0
347.00	638.0
347.90	830.0
351.70	2728.0
352.10	2400.0
353.00	3342.0
353.90	5011.0
355.00	1058.0
363.90	434.0
365.00	15491.0
365.90	3108.0
369.60	546.0
370.80	1347.0
372.00	5948.0
373.00	1599.0
383.00	2521.0
383.90	542.0
389.90	1564.0
390.90	974.0
392.70	332.0
400.80	435.0

Instrument :
BNA_N
ClientSampleId :
DFTPP099

401.70	2850.0
402.10	2156.0
403.10	4052.0
404.00	1104.0
420.80	3030.0
422.00	3275.0
423.00	24440.0
424.20	5365.0
424.90	660.0
438.30	959.0
439.00	545.0
439.40	608.0
441.00	44576.0
442.00	266368.0
443.00	55296.0
444.00	6159.0

Instrument :
BNA_N
ClientSampleId :
DFTPP099

m/z	Abundance
36.90	623.0
38.20	1677.0
39.00	6282.0
40.10	457.0
43.80	931.0
44.10	769.0
47.10	322.0
49.10	1514.0
50.00	38136.0
51.00	106768.0
52.00	3541.0
53.00	877.0
54.90	801.0
55.90	3004.0
57.00	8462.0
58.00	487.0
58.80	303.0
59.90	529.0
60.60	438.0
61.10	1858.0
62.10	2456.0
63.10	4274.0
64.00	795.0
65.10	2126.0
67.10	761.0
69.00	150848.0
69.90	1400.0
73.10	2153.0
74.00	20624.0
75.00	26480.0
76.10	9805.0
77.00	157376.0
78.00	10857.0
78.90	18976.0
80.00	9382.0
81.00	15137.0
81.90	2843.0
83.00	2685.0
84.00	583.0
85.10	4676.0
85.90	5601.0
86.90	1933.0
88.00	726.0
90.00	743.0
91.10	4360.0
92.00	3916.0
93.00	26128.0
94.00	3257.0
94.90	652.0
95.90	1002.0
97.10	1203.0
98.00	25552.0
99.00	15627.0
99.80	1284.0
100.20	949.0
100.90	8951.0
103.10	2660.0
103.80	5578.0
104.90	7051.0
106.00	2612.0
106.90	69208.0
108.00	8214.0
108.80	2433.0
109.10	2225.0
110.00	131840.0
111.00	18976.0
112.00	1841.0
112.80	488.0
114.70	539.0
116.00	4155.0
117.00	66288.0
118.00	5789.0
119.20	366.0
120.00	532.0
121.00	655.0
122.00	3044.0
122.90	4841.0
123.90	2316.0
124.80	2851.0
125.10	2848.0
127.00	194880.0
128.00	11829.0
129.00	76808.0
129.90	6738.0
130.90	2503.0
132.10	729.0
133.00	1206.0
134.00	2967.0
135.00	6361.0
136.00	2980.0
137.00	3668.0
137.90	1590.0

Instrument :
BNA_N
ClientSampleId :
DFTPP099

138.90	1578.0
139.70	1197.0
140.90	13016.0
141.90	4701.0
143.00	1849.0
143.90	680.0
144.70	427.0
146.10	1847.0
146.90	5751.0
148.00	16218.0
149.00	3073.0
149.80	394.0
150.20	464.0
150.80	1361.0
151.50	688.0
152.10	756.0
153.00	4176.0
153.90	3001.0
155.00	6100.0
156.00	13035.0
157.00	1397.0
158.00	2424.0
158.80	1583.0
160.00	2998.0
160.80	5224.0
161.20	4780.0
162.20	1276.0
163.20	425.0
164.90	6234.0
166.00	5176.0
167.00	25472.0
168.00	15448.0
169.00	1496.0
169.90	545.0
170.80	1528.0
172.10	1983.0
173.00	3942.0
173.90	6624.0
175.00	8962.0
175.90	2443.0
176.90	4884.0
177.90	2567.0
178.90	16383.0
180.10	12018.0
180.90	4849.0
182.10	1454.0
183.30	770.0
184.00	2555.0
185.00	9848.0
186.00	75608.0
187.00	23696.0
188.00	3021.0
188.50	1766.0
189.10	4792.0
190.00	1194.0
191.00	3295.0
192.00	4665.0
193.00	6014.0
194.20	2135.0
195.00	1774.0
195.90	13517.0
197.90	655360.0
199.00	45032.0
200.00	3519.0
200.90	1608.0
201.30	1894.0
201.70	1784.0
202.30	1043.0
202.90	4947.0
204.00	18904.0
205.10	37048.0
206.00	147072.0
207.00	17792.0
208.00	4277.0
209.00	1357.0
210.20	2401.0
211.00	6601.0
212.10	720.0
214.70	1600.0
215.20	2299.0
216.10	3843.0
216.90	35008.0
217.90	4612.0
218.90	414.0
220.10	1717.0
220.80	13093.0
221.90	4088.0
223.10	10513.0
224.10	77512.0
225.10	21576.0
226.10	3156.0
227.00	41392.0
227.90	5293.0
229.00	7201.0
229.80	743.0
230.10	849.0

Instrument :
BNA_N
ClientSampleId :
DFTPP099

230.90	2113.0
232.80	747.0
233.90	2135.0
235.10	2780.0
236.00	1225.0
236.90	1499.0
238.20	811.0
238.80	1648.0
240.20	1361.0
241.00	1681.0
242.00	3553.0
242.90	4235.0
244.10	62624.0
245.00	9421.0
246.00	16154.0
247.00	2240.0
248.20	775.0
249.00	2641.0
251.10	906.0
251.60	1005.0
252.10	1190.0
253.20	2963.0
255.00	361280.0
256.00	48952.0
257.00	3053.0
258.00	24640.0
258.90	2856.0
259.90	848.0
261.10	777.0
261.90	788.0
264.10	1252.0
264.90	8774.0
269.90	871.0
271.20	565.0
271.70	1067.0
273.00	9958.0
274.00	34168.0
275.00	157056.0
276.00	20928.0
276.90	16712.0
277.90	3241.0
278.90	327.0
283.00	988.0
284.20	855.0
285.00	2606.0
286.10	1143.0
288.80	608.0
290.10	368.0
291.10	411.0
291.90	520.0
293.00	3689.0
293.80	2432.0
296.00	51048.0
297.00	6368.0
298.20	331.0
299.30	321.0
300.60	337.0
301.10	366.0
302.20	532.0
303.00	4925.0
303.90	2207.0
308.00	1228.0
310.00	941.0
314.00	2327.0
315.10	6321.0
316.00	3143.0
317.10	813.0
320.90	1889.0
323.10	17456.0
324.00	3867.0
327.00	3375.0
327.90	1684.0
332.00	1518.0
332.70	890.0
333.20	1483.0
334.00	9379.0
335.00	3269.0
338.90	760.0
339.60	503.0
340.90	2766.0
341.70	410.0
346.00	3777.0
346.70	503.0
349.80	436.0
350.80	669.0
352.10	5207.0
353.00	5479.0
354.20	4019.0
355.10	1205.0
362.20	345.0
363.30	384.0
365.00	18000.0
366.00	2909.0
366.90	333.0
370.10	1920.0
370.90	1697.0

Instrument :
BNA_N
ClientSampleId :
DFTPP099

372.00	7642.0
373.20	2210.0
382.90	2395.0
383.80	953.0
389.80	935.0
391.10	2040.0
391.90	428.0
402.00	5654.0
402.90	4185.0
403.90	1287.0
405.10	826.0
410.00	464.0
420.90	4779.0
422.00	5131.0
423.00	27624.0
424.10	4596.0
424.90	1860.0
437.30	315.0
438.30	1073.0
438.80	568.0
439.40	1698.0
441.00	69968.0
442.10	379136.0
443.00	72776.0
444.00	6285.0

Instrument :
BNA_N
ClientSampleId :
DFTPP099

m/z	Abundance
36.80	577.0
38.00	1190.0
38.90	1968.0
43.20	349.0
44.20	573.0
50.00	18728.0
51.00	62928.0
52.10	3119.0
55.00	723.0
55.90	2542.0
57.00	3982.0
57.90	419.0
61.00	783.0
61.90	1297.0
62.90	2669.0
65.00	2585.0
68.90	84624.0
70.00	957.0
73.20	1552.0
74.00	14911.0
75.00	16145.0
76.00	5824.0
77.00	103720.0
78.00	7546.0
79.00	9883.0
80.00	7110.0
81.00	9595.0
81.80	1637.0
82.90	2738.0
84.90	2023.0
86.00	4371.0
87.00	1402.0
88.00	575.0
91.00	2644.0
92.20	2528.0
93.00	17200.0
94.00	1235.0
94.90	699.0
95.80	1578.0
97.20	1243.0
98.00	17064.0
99.00	9544.0
99.80	823.0
101.00	6583.0
101.90	641.0
102.90	2973.0
104.10	3917.0
105.00	2815.0
105.80	1397.0
107.00	39920.0
107.90	5866.0
109.90	93824.0
110.90	13112.0
111.90	2173.0
114.10	704.0
114.80	427.0
117.00	43816.0
118.00	3733.0
120.00	1134.0
121.20	441.0
121.80	2081.0
122.20	1654.0
123.00	4760.0
124.00	2888.0
124.90	1649.0
125.80	1330.0
126.20	1685.0
127.00	108608.0
128.00	10110.0
129.00	57448.0
130.00	4872.0
130.80	1533.0
132.10	945.0
133.90	2325.0
135.00	5878.0
136.20	1628.0
136.80	2123.0
139.20	679.0
140.90	7090.0
141.90	2847.0
143.00	1419.0
146.10	1665.0
147.00	6086.0
148.00	9994.0
149.10	2565.0
150.30	738.0
150.70	420.0
151.30	699.0
152.10	948.0
152.90	2331.0
153.70	1219.0
154.10	1155.0

Instrument :
BNA_N
ClientSampleId :
DFTPP099

154.90	6540.0
156.00	6429.0
157.00	1009.0
157.90	1412.0
158.80	948.0
159.80	2978.0
161.00	4620.0
162.00	1302.0
164.10	322.0
165.00	2486.0
165.90	2349.0
166.90	15903.0
167.90	10114.0
169.00	2357.0
170.00	509.0
170.70	883.0
172.00	2719.0
172.90	2207.0
173.90	3175.0
174.90	5260.0
175.90	984.0
176.40	832.0
177.00	4169.0
177.80	1849.0
178.90	11793.0
179.90	9348.0
180.90	3560.0
185.00	8737.0
186.00	59176.0
187.00	21464.0
188.10	1839.0
188.90	3352.0
190.10	763.0
191.00	1958.0
191.90	3840.0
192.90	4930.0
194.10	1411.0
195.00	1069.0
195.90	7125.0
197.90	484992.0
199.00	33176.0
200.10	2539.0
201.50	1384.0
202.90	2526.0
203.90	13472.0
205.00	22768.0
206.00	108536.0
207.00	15419.0
207.90	3102.0
209.00	2342.0
210.00	956.0
211.00	3302.0
212.10	826.0
212.90	848.0
215.10	2183.0
216.00	4070.0
217.00	28304.0
218.00	2999.0
219.70	378.0
220.20	1801.0
220.90	10205.0
221.80	2525.0
223.10	6168.0
224.00	61384.0
225.10	11342.0
226.90	34576.0
227.90	5087.0
228.90	5859.0
230.10	406.0
231.20	933.0
233.20	1234.0
233.90	1509.0
234.90	1734.0
235.90	1048.0
236.90	2001.0
237.80	675.0
239.10	1066.0
239.90	778.0
241.00	2438.0
242.00	2614.0
242.90	3624.0
244.00	47408.0
244.90	5900.0
245.90	13171.0
247.10	1411.0
248.80	1591.0
250.10	537.0
251.20	812.0
252.00	1052.0
253.10	2362.0
255.00	268224.0
256.00	37896.0
257.10	3015.0
258.00	19816.0
259.00	2465.0
259.80	581.0

Instrument :
BNA_N
ClientSampleId :
DFTPP099

260.60	1102.0
264.10	1264.0
265.00	6677.0
266.10	1627.0
269.90	446.0
270.90	1458.0
271.80	794.0
272.90	10931.0
274.00	29720.0
275.00	143232.0
276.00	19320.0
276.90	16456.0
277.90	2034.0
281.10	633.0
282.00	543.0
282.80	490.0
283.20	838.0
284.10	1153.0
285.10	1871.0
286.00	977.0
291.20	412.0
292.10	500.0
293.00	3837.0
294.10	493.0
295.90	38080.0
297.10	4273.0
299.10	392.0
300.80	957.0
301.90	936.0
303.00	4883.0
307.80	695.0
308.80	390.0
309.40	314.0
310.00	409.0
312.90	467.0
313.70	1106.0
314.00	862.0
315.00	5485.0
316.00	3469.0
320.90	2268.0
321.90	1084.0
323.00	18592.0
324.00	3680.0
325.90	519.0
326.90	2268.0
328.00	1110.0
332.20	925.0
333.10	1159.0
334.00	9679.0
335.00	1598.0
341.00	3304.0
341.90	540.0
346.00	1861.0
347.10	586.0
352.00	3527.0
353.00	3090.0
353.90	4889.0
355.20	766.0
359.10	398.0
362.80	442.0
364.90	16880.0
365.80	2653.0
366.90	479.0
369.70	590.0
371.00	1755.0
372.10	7421.0
373.10	2432.0
382.80	1432.0
383.80	1344.0
385.00	342.0
389.90	1236.0
391.10	408.0
394.90	375.0
402.00	4691.0
403.10	6745.0
404.00	1584.0
415.00	471.0
421.00	4742.0
422.00	4526.0
423.00	24328.0
424.10	5971.0
425.00	429.0
438.10	410.0
438.40	531.0
439.10	852.0
439.70	1733.0
441.00	46536.0
442.00	380928.0
443.00	69968.0
444.10	4615.0
444.90	518.0

Instrument :
BNA_N
ClientSampleId :
DFTPP099