

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061323\
 Data File : BN025819.D
 Acq On : 13 Jun 2023 23:24
 Operator : MA/JU
 Sample : 02950-13MS
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 EW990MS

Quant Time: Jun 14 00:03:30 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN060923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jun 10 02:01:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.005	152	6686	0.400	ng/ul	0.00
4) Naphthalene-d8	10.820	136	20689	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.640	164	11780	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.385	188	21797	0.400	ng/ul	0.00
17) Chrysene-d12	21.564	240	11114	0.400	ng/ul	0.00
23) Perylene-d12	24.049	264	11301	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.356	96	1698	0.258	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.403	152	4137	0.158	ng/ul	0.00
18) Fluoranthene-d10	19.405	212	8794	0.251	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.390	88	5511	0.783	ng/ul	97
5) Naphthalene	10.869	128	9607	0.183	ng/ul	98
7) 2-Methylnaphthalene	12.475	142	6221	0.196	ng/ul	100
8) 1-Methylnaphthalene	12.695	142	6583	0.189	ng/ul	100
10) Acenaphthylene	14.362	152	9060	0.195	ng/ul	96
11) Acenaphthene	14.700	153	7493	0.193	ng/ul	98
12) Fluorene	15.685	166	8424	0.195	ng/ul	98
14) Pentachlorophenol	17.039	266	116	0.049	ng/ul#	80
15) Phenanthrene	17.423	178	15702	0.247	ng/ul	98
16) Anthracene	17.516	178	11476	0.214	ng/ul	96
19) Fluoranthene	19.438	202	17441	0.343	ng/ul	99
20) Pyrene	19.800	202	17829	0.339	ng/ul	98
21) Benzo(a)anthracene	21.546	228	11036	0.311	ng/ul	96
22) Chrysene	21.602	228	12493	0.308	ng/ul	97
24) Benzo(b)fluoranthene	23.286	252	11860	0.257	ng/ul	86
25) Benzo(k)fluoranthene	23.335	252	11199	0.242	ng/ul#	88
26) Benzo(a)pyrene	23.940	252	9526	0.243	ng/ul#	86
27) Indeno(1,2,3-cd)pyrene	26.651	276	15268	0.342	ng/ul#	91
28) Dibenzo(a,h)anthracene	26.668	278	10090	0.298	ng/ul	96
29) Benzo(g,h,i)perylene	27.450	276	10148	0.253	ng/ul	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061323\
Data File : BN025819.D
Acq On : 13 Jun 2023 23:24
Operator : MA/JU
Sample : 02950-13MS
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
EW990MS

Quant Time: Jun 14 00:03:30 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN060923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jun 10 02:01:49 2023
Response via : Initial Calibration

