

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061323\
 Data File : BN025820.D
 Acq On : 14 Jun 2023 00:01
 Operator : MA/JU
 Sample : 02950-14MSD
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 EW990MSD

Quant Time: Jun 14 00:30:58 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN060923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jun 10 02:01:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.001	152	8528	0.400	ng/ul	-0.01
4) Naphthalene-d8	10.820	136	26033	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.640	164	15289	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.385	188	29610	0.400	ng/ul	0.00
17) Chrysene-d12	21.561	240	15117	0.400	ng/ul	0.00
23) Perylene-d12	24.040	264	14657	0.400	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.352	96	1853	0.221	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.403	152	4662	0.141	ng/ul	0.00
18) Fluoranthene-d10	19.405	212	9420	0.198	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.386	88	6886	0.767	ng/ul	94
5) Naphthalene	10.869	128	12195	0.184	ng/ul	98
7) 2-Methylnaphthalene	12.475	142	8025	0.201	ng/ul	99
8) 1-Methylnaphthalene	12.695	142	8350	0.190	ng/ul	100
10) Acenaphthylene	14.362	152	11576	0.192	ng/ul	98
11) Acenaphthene	14.700	153	9544	0.189	ng/ul	99
12) Fluorene	15.685	166	10998	0.196	ng/ul	98
15) Phenanthrene	17.423	178	20129	0.233	ng/ul	99
16) Anthracene	17.516	178	15213	0.209	ng/ul	98
19) Fluoranthene	19.438	202	20739	0.300	ng/ul	99
20) Pyrene	19.796	202	21309	0.298	ng/ul	97
21) Benzo(a)anthracene	21.544	228	12834	0.266	ng/ul	96
22) Chrysene	21.599	228	14392	0.261	ng/ul	97
24) Benzo(b)fluoranthene	23.277	252	13318	0.222	ng/ul	90
25) Benzo(k)fluoranthene	23.324	252	12393	0.206	ng/ul#	89
26) Benzo(a)pyrene	23.932	252	10714	0.211	ng/ul#	88
27) Indeno(1,2,3-cd)pyrene	26.645	276	15701	0.271	ng/ul#	96
28) Dibenzo(a,h)anthracene	26.662	278	11108	0.253	ng/ul	99
29) Benzo(g,h,i)perylene	27.436	276	11150	0.215	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061323\
Data File : BN025820.D
Acq On : 14 Jun 2023 00:01
Operator : MA/JU
Sample : 02950-14MSD
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
EW990MSD

Quant Time: Jun 14 00:30:58 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN060923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jun 10 02:01:49 2023
Response via : Initial Calibration

