


Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN061419\
 Data File : BN006325.D
 Acq On : 14 Jun 2019 08:46
 Operator : JU/SJ
 Sample : DFTPP92
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP92

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN053019.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Jun 13 11:20:30 2019



AutoFind: Scans 2466, 2467, 2468; Background Corrected with Scan 2459

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.0	187840	PASS
68	69	0.00	2	1.9	3460	PASS
69	198	0.00	100	37.6	185941	PASS
70	69	0.00	2	0.6	1048	PASS
127	198	10	80	45.1	223274	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	494805	PASS
199	198	5	9	6.9	34154	PASS
275	198	10	60	25.8	127504	PASS
365	198	1	100	3.2	15652	PASS
441	198	0.01	100	13.5	66717	PASS
442	198	50	100	85.6	423722	PASS
443	442	15	24	19.8	83893	PASS

m/z	Abundance
39.10	319.0
39.90	2766.0
41.00	413.0
43.00	409.0
44.00	3813.0
55.10	405.0
57.10	327.0
73.00	462.0
193.00	302.0
207.00	1177.0

Instrument :
BNA_N
ClientSampleId :
DFTPP92

m/z	Abundance
37.00	782.0
38.00	2045.0
39.00	12903.0
40.00	2941.0
41.00	716.0
43.00	607.0
44.00	3678.0
45.00	374.0
49.00	1585.0
50.00	44864.0
51.00	175744.0
52.00	9051.0
53.00	504.0
55.00	1110.0
56.00	5492.0
57.00	11718.0
58.00	616.0
61.00	2438.0
62.00	2672.0
63.00	6761.0
64.00	1038.0
65.00	3752.0
66.00	416.0
67.00	375.0
68.00	3187.0
69.00	171712.0
70.00	1099.0
73.00	1516.0
74.00	17584.0
75.00	28640.0
76.10	10195.0
77.00	204224.0
78.00	15292.0
79.00	12413.0
80.00	9036.0
81.00	14116.0
82.00	3589.0
83.00	3266.0
84.00	460.0
85.00	2727.0
86.00	3920.0
87.00	1683.0
88.00	647.0
89.00	418.0
91.00	3156.0
92.00	3240.0
93.00	21632.0
94.00	1338.0
95.00	471.0
96.00	1070.0
97.00	717.0
98.00	16704.0
99.00	14505.0
100.00	1251.0
101.00	7714.0
101.90	478.0
103.00	2611.0
104.00	4199.0
105.00	4583.0
106.10	1512.0
107.00	57976.0
108.00	8694.0
109.00	1701.0
110.00	107480.0
111.00	16624.0
112.10	2165.0
112.90	626.0
114.90	325.0
116.00	3309.0
117.00	46928.0
118.00	3983.0
118.80	578.0
120.00	736.0
121.10	340.0
122.00	3481.0
123.00	5654.0
124.00	2729.0
125.00	2220.0
127.00	198720.0
128.00	16219.0
129.00	72304.0
130.00	6300.0
131.00	1375.0
131.90	641.0
133.00	425.0
134.00	2133.0
135.00	6060.0
136.00	2126.0
137.00	3355.0
138.00	883.0
139.00	485.0

Instrument :
BNA_N
ClientSampleId :
DFTPP92

140.00	1077.0
141.00	9779.0
142.00	2843.0
143.00	2155.0
144.00	555.0
144.90	605.0
145.90	1591.0
147.00	4972.0
148.00	11629.0
149.00	2589.0
149.90	680.0
151.00	1274.0
152.00	774.0
153.00	3344.0
154.00	2495.0
155.00	5422.0
156.00	8005.0
157.00	1361.0
158.00	1797.0
158.90	1549.0
160.00	3047.0
161.00	4442.0
162.00	1460.0
163.90	553.0
165.00	3690.0
166.00	3051.0
167.00	19128.0
168.00	12465.0
169.00	1968.0
170.00	512.0
171.00	697.0
171.90	1902.0
173.00	2217.0
174.00	4247.0
175.00	7446.0
176.00	2275.0
177.00	3174.0
178.10	1108.0
179.00	13536.0
180.00	10568.0
181.00	4603.0
181.90	965.0
183.20	337.0
184.00	1019.0
185.00	6569.0
186.00	53328.0
187.00	15428.0
188.00	1833.0
189.00	2640.0
190.10	512.0
190.90	1922.0
192.00	3932.0
193.00	4909.0
194.00	1124.0
195.00	773.0
196.00	11377.0
198.00	400640.0
199.00	28592.0
200.00	1985.0
201.50	1902.0
203.00	2574.0
204.00	14142.0
205.00	24568.0
206.00	100632.0
207.00	14342.0
208.00	3626.0
209.00	1312.0
210.10	1631.0
211.00	3519.0
211.70	1112.0
213.00	420.0
215.00	1197.0
216.00	1933.0
217.00	24544.0
218.00	3389.0
218.90	310.0
221.00	16584.0
221.90	3574.0
223.00	5327.0
224.00	53944.0
225.00	14657.0
226.10	1788.0
227.00	21640.0
228.00	3427.0
229.00	4837.0
230.00	679.0
231.10	1910.0
232.00	434.0
233.00	453.0
234.10	1273.0
235.00	1457.0
236.00	1185.0
237.00	2744.0
239.00	761.0
239.90	659.0

Instrument :
BNA_N
ClientSampleId :
DFTPP92

241.00	1045.0
242.00	2831.0
243.00	3254.0
244.00	40816.0
245.00	5475.0
246.00	7290.0
247.00	1536.0
248.00	432.0
249.00	1498.0
250.90	476.0
252.10	443.0
253.00	945.0
255.00	203968.0
256.00	30088.0
257.00	2729.0
258.00	11065.0
259.00	1755.0
260.00	409.0
261.00	423.0
265.00	4300.0
265.90	645.0
271.00	457.0
272.00	786.0
273.00	5516.0
274.00	17312.0
275.00	93680.0
276.00	13818.0
277.00	6900.0
278.00	1523.0
279.10	307.0
282.00	415.0
283.00	888.0
283.90	554.0
285.00	1330.0
288.90	376.0
290.00	334.0
292.10	368.0
293.00	1659.0
294.00	586.0
295.00	752.0
296.00	23048.0
297.00	3565.0
301.90	393.0
303.00	2643.0
304.00	699.0
309.00	323.0
314.00	1117.0
315.00	2269.0
316.00	1674.0
320.90	751.0
322.00	455.0
323.00	8073.0
324.00	1458.0
327.00	1329.0
328.00	720.0
331.90	628.0
333.00	655.0
334.00	4674.0
335.00	1179.0
341.00	934.0
346.00	1845.0
352.00	2132.0
353.00	1553.0
354.00	2347.0
355.00	621.0
365.00	10781.0
366.00	1654.0
370.90	560.0
372.00	4148.0
372.90	1077.0
383.00	1049.0
390.00	637.0
391.00	354.0
392.00	307.0
400.90	503.0
402.00	1712.0
403.00	2629.0
404.00	1103.0
421.00	2083.0
422.00	1967.0
423.00	14251.0
424.00	4638.0
425.00	670.0
441.00	38440.0
442.00	240960.0
443.00	48912.0
444.00	4632.0
445.00	321.0

Instrument :
BNA_N
ClientSampleId :
DFTPP92

m/z	Abundance
37.00	977.0
38.00	2252.0
39.00	16058.0
40.00	3091.0
41.00	818.0
43.00	577.0
44.00	3039.0
45.00	550.0
49.10	1730.0
50.00	55304.0
51.00	221696.0
52.00	10548.0
53.00	553.0
55.00	1822.0
56.00	6580.0
57.00	15045.0
58.00	675.0
61.00	3036.0
62.00	3055.0
63.00	8323.0
64.10	1205.0
65.00	4349.0
67.00	414.0
68.00	4245.0
69.00	220352.0
70.00	962.0
73.00	2040.0
74.00	22464.0
75.00	35720.0
76.00	12828.0
77.00	264960.0
78.00	18144.0
79.00	16600.0
80.00	12902.0
81.00	18168.0
82.00	4567.0
83.00	4248.0
84.00	641.0
85.00	4558.0
86.00	4994.0
87.00	2483.0
88.00	824.0
89.00	398.0
91.00	4110.0
92.00	4423.0
93.00	27312.0
94.00	2097.0
95.00	495.0
96.00	1260.0
97.00	765.0
98.00	21920.0
99.00	17976.0
99.90	1550.0
101.00	9743.0
101.90	668.0
103.00	3385.0
104.00	6039.0
105.00	5825.0
106.00	1988.0
107.00	76504.0
108.00	12391.0
109.10	2724.0
110.00	144384.0
111.00	21816.0
112.00	2613.0
113.00	880.0
115.00	494.0
116.00	4169.0
117.00	59696.0
118.00	4566.0
119.00	821.0
120.00	840.0
121.00	459.0
122.00	4808.0
123.00	7613.0
124.00	3487.0
125.00	3341.0
127.00	262080.0
128.00	19528.0
129.00	96128.0
130.00	8700.0
131.00	1338.0
132.00	890.0
132.90	484.0
134.00	2392.0
135.00	7606.0
136.00	3462.0
137.00	4157.0
137.90	1053.0
139.10	514.0
140.00	1337.0

Instrument :
BNA_N
ClientSampleId :
DFTPP92

141.00	13031.0
142.00	4143.0
143.00	2986.0
144.00	815.0
145.00	619.0
146.00	2349.0
147.00	6642.0
148.00	15234.0
149.00	3471.0
149.90	892.0
151.10	1786.0
151.90	1070.0
153.00	4506.0
154.00	3027.0
155.00	7515.0
156.00	11504.0
157.00	2223.0
157.90	2236.0
159.00	1982.0
160.00	4824.0
161.00	5719.0
162.00	1728.0
163.00	632.0
163.90	795.0
165.00	5123.0
166.00	4015.0
167.00	27056.0
168.00	15851.0
169.00	2178.0
170.00	843.0
171.00	1223.0
172.00	2418.0
173.00	3081.0
174.00	6138.0
175.00	10460.0
176.00	3562.0
177.00	4630.0
178.00	1604.0
179.00	20368.0
180.00	13821.0
181.00	6943.0
182.00	1064.0
182.90	650.0
184.00	1383.0
185.00	9296.0
186.00	78784.0
187.00	21392.0
188.00	2635.0
189.00	4885.0
189.90	855.0
191.00	2043.0
192.00	6479.0
193.00	6912.0
194.00	1797.0
195.10	944.0
196.00	17240.0
198.00	590848.0
199.00	40712.0
200.00	3334.0
201.50	2517.0
203.00	3628.0
204.00	20536.0
205.00	34464.0
206.00	149184.0
207.00	20320.0
208.00	4774.0
209.10	1565.0
210.20	2408.0
211.00	5380.0
213.00	391.0
215.00	1483.0
216.00	3000.0
217.00	36360.0
218.00	5161.0
219.00	431.0
221.00	25256.0
221.70	5292.0
223.00	8698.0
224.00	81096.0
225.00	21544.0
226.00	2197.0
227.00	31440.0
228.00	4884.0
229.00	6435.0
230.00	780.0
231.00	3154.0
232.10	527.0
233.00	473.0
234.00	2188.0
235.00	2475.0
235.90	1809.0
237.00	3053.0
237.90	499.0
239.00	1215.0
239.90	1060.0

Instrument :
BNA_N
ClientSampleId :
DFTPP92

241.00	2141.0
242.00	4496.0
243.00	4654.0
244.00	63704.0
245.10	8380.0
246.00	11852.0
247.00	2512.0
248.00	652.0
249.00	2299.0
250.00	544.0
251.00	541.0
251.90	557.0
253.00	1305.0
255.00	327744.0
256.00	48176.0
257.10	3971.0
258.00	16792.0
259.00	2894.0
260.00	442.0
260.80	504.0
264.00	417.0
265.00	7143.0
266.00	1089.0
270.00	348.0
271.10	460.0
272.00	798.0
273.00	9706.0
274.00	25664.0
275.00	151616.0
276.00	20936.0
277.00	12068.0
278.00	2066.0
278.90	400.0
280.90	395.0
283.00	1077.0
284.00	667.0
285.00	2198.0
285.90	430.0
288.90	458.0
290.00	404.0
291.10	339.0
292.00	625.0
293.00	3270.0
294.00	646.0
296.00	39176.0
297.00	5271.0
298.00	459.0
300.90	625.0
301.90	735.0
303.00	4342.0
304.00	1418.0
308.00	546.0
309.00	391.0
309.80	332.0
310.10	395.0
314.00	1912.0
315.00	4275.0
316.00	2750.0
316.90	455.0
321.00	970.0
322.00	566.0
323.00	12655.0
324.00	2482.0
326.90	2285.0
328.00	1343.0
332.10	1012.0
333.10	1154.0
334.00	8386.0
335.00	2332.0
336.00	434.0
341.00	1365.0
342.00	433.0
345.90	3088.0
347.00	488.0
351.10	360.0
352.00	3834.0
353.00	2975.0
354.00	4413.0
355.00	1043.0
359.10	470.0
365.00	18232.0
366.00	2869.0
370.00	395.0
371.00	1325.0
372.00	7732.0
373.00	1994.0
383.00	1877.0
383.90	642.0
390.00	993.0
391.00	631.0
392.00	642.0
401.00	443.0
402.00	2966.0
403.00	4303.0
404.00	1604.0

Instrument :
BNA_N
ClientSampleId :
DFTPP92

421.00	3871.0
422.10	3668.0
423.00	28840.0
424.00	7840.0
425.00	1215.0
441.10	77200.0
442.00	490880.0
443.00	97840.0
444.00	9025.0
444.90	559.0

Instrument :
BNA_N
ClientSampleId :
DFTPP92

m/z	Abundance
37.00	682.0
38.00	1723.0
39.00	11733.0
40.00	2719.0
41.10	845.0
43.00	520.0
44.00	3451.0
45.00	538.0
49.10	1271.0
50.00	41800.0
51.00	166080.0
52.00	8686.0
53.00	564.0
55.00	1256.0
56.00	4831.0
57.00	11925.0
58.00	643.0
61.00	1905.0
62.00	2125.0
63.00	6628.0
64.10	840.0
65.00	3316.0
66.00	314.0
67.00	332.0
68.00	2950.0
69.00	165760.0
70.00	1085.0
73.00	1702.0
74.00	16584.0
75.00	27424.0
76.00	9057.0
77.00	201728.0
78.10	13862.0
79.00	12539.0
80.00	9614.0
81.00	13441.0
82.00	3405.0
83.00	3438.0
84.00	654.0
84.90	3744.0
86.00	3982.0
87.00	1664.0
88.00	640.0
89.00	433.0
91.00	3314.0
92.00	3367.0
93.00	21472.0
94.00	1448.0
95.00	503.0
96.00	1020.0
97.00	748.0
98.00	16888.0
99.00	13529.0
100.00	1259.0
101.00	7080.0
101.90	537.0
103.00	2423.0
104.00	4754.0
105.00	4745.0
106.00	1502.0
107.00	58144.0
108.00	8837.0
109.10	1719.0
110.00	115344.0
111.00	16888.0
112.00	2053.0
112.90	622.0
114.90	506.0
116.00	3297.0
117.00	47232.0
117.90	3933.0
119.00	843.0
120.00	969.0
121.10	329.0
122.00	4383.0
123.00	5878.0
124.00	3258.0
125.00	2493.0
126.10	1258.0
127.00	209024.0
128.00	15551.0
129.00	76264.0
130.00	6714.0
131.00	1155.0
132.10	837.0
133.00	530.0
134.00	2428.0
135.00	7121.0
136.00	2285.0
137.00	3706.0
137.90	714.0

Instrument :
BNA_N
ClientSampleId :
DFTPP92

139.00	438.0
140.00	1055.0
141.00	10524.0
142.00	3343.0
143.00	2298.0
144.00	693.0
145.00	711.0
146.00	2049.0
147.00	5832.0
148.00	11843.0
149.00	2688.0
149.90	906.0
151.10	1301.0
151.90	1002.0
153.00	3599.0
154.00	2696.0
155.00	6046.0
156.00	8803.0
157.00	1917.0
157.90	1866.0
159.00	1612.0
160.00	2919.0
161.00	5118.0
161.90	1584.0
162.80	406.0
163.90	704.0
165.00	4311.0
166.00	3423.0
167.00	23280.0
168.00	12454.0
169.00	2457.0
169.90	648.0
171.00	969.0
172.00	1998.0
173.00	2734.0
174.00	4873.0
175.00	9488.0
176.00	2363.0
177.00	4554.0
178.00	1478.0
179.00	16856.0
180.00	11590.0
181.00	5030.0
182.00	806.0
182.90	573.0
184.00	1507.0
185.00	8367.0
186.00	65528.0
187.00	18040.0
188.00	2003.0
189.00	4017.0
190.00	771.0
191.00	2000.0
192.00	4923.0
193.00	5527.0
194.00	1289.0
194.90	821.0
196.00	14211.0
198.00	492928.0
199.00	33160.0
200.00	2490.0
201.40	2233.0
203.00	3350.0
204.00	16888.0
205.00	31128.0
206.00	125336.0
207.00	16848.0
208.00	4230.0
209.00	1655.0
209.90	1779.0
210.90	4409.0
215.00	1422.0
216.10	2974.0
217.00	32680.0
218.00	4388.0
218.90	599.0
221.00	21928.0
223.00	7678.0
224.00	70384.0
225.00	17512.0
226.10	2165.0
227.00	27880.0
228.00	4016.0
229.00	6583.0
230.00	924.0
231.00	2745.0
232.10	467.0
233.00	581.0
234.00	2134.0
235.00	2245.0
236.00	1214.0
237.00	2155.0
238.00	449.0
238.90	1253.0
240.00	918.0

Instrument :
BNA_N
ClientSampleId :
DFTPP92

241.00	1483.0
242.00	3930.0
243.00	4016.0
244.00	58688.0
245.10	7161.0
246.00	10338.0
247.00	2436.0
247.90	411.0
249.00	2037.0
250.00	382.0
250.90	395.0
251.90	541.0
253.00	1587.0
255.00	290880.0
256.00	41936.0
257.00	3182.0
258.00	14994.0
259.00	2499.0
259.90	480.0
261.00	407.0
263.90	525.0
265.00	6391.0
266.00	1101.0
271.00	545.0
272.00	700.0
273.00	8737.0
274.00	24736.0
275.00	137216.0
276.00	18072.0
277.00	11439.0
278.00	1695.0
279.10	393.0
283.00	1274.0
284.00	954.0
285.00	1926.0
286.00	368.0
289.00	551.0
290.00	302.0
291.00	308.0
292.10	602.0
293.00	2643.0
294.10	648.0
296.00	36312.0
297.00	5029.0
298.00	350.0
301.10	494.0
301.90	672.0
303.00	4472.0
304.10	1129.0
308.00	534.0
308.90	381.0
310.00	539.0
313.10	394.0
314.00	1767.0
315.00	4021.0
316.00	2003.0
317.10	357.0
321.00	1085.0
322.10	648.0
323.00	13206.0
324.00	2251.0
326.90	2264.0
328.00	1125.0
329.00	321.0
332.00	1089.0
333.10	1014.0
334.00	7945.0
335.00	2036.0
341.00	1676.0
342.10	541.0
346.00	3211.0
346.90	462.0
352.00	4341.0
353.00	2905.0
354.10	4541.0
355.00	853.0
365.00	17944.0
366.00	2281.0
370.00	428.0
371.00	1228.0
372.00	7541.0
373.00	1924.0
383.00	2062.0
384.00	590.0
390.00	954.0
391.00	638.0
392.00	572.0
400.90	532.0
402.00	3502.0
403.00	4461.0
404.00	1455.0
405.00	395.0
421.00	4258.0
422.00	3949.0
423.00	30408.0

Instrument :
BNA_N
ClientSampleId :
DFTPP92

424.00	6636.0
425.10	744.0
441.00	84512.0
442.10	539328.0
443.10	104928.0
444.00	9274.0
445.00	527.0

Instrument :
BNA_N
ClientSampleId :
DFTPP92