




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061523\
Data File : BN025849.D
Acq On : 15 Jun 2023 09:29
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 16 04:35:45 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 16 04:35:40 2023
Response via : Initial Calibration



TIC: BN025849.D\data.ms

(70) Pentachlorophenol (C)




17.022min (0.000) 18603.67 ng

response 283380

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	61.65
264.00	61.60	60.29
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061523\
 Data File : BN025849.D
 Acq On : 15 Jun 2023 09:29
 Operator : MA/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 16 04:35:45 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 16 04:35:40 2023
 Response via : Initial Calibration



TIC: BN025849.D\data.ms

(77) Benzidine**19.616min (0.000) 745098.91 ng****response 1401045**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.75
183.00	13.20	11.96
0.00	0.00	0.00