

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061620\
 Data File : BN010903.D
 Acq On : 16 Jun 2020 12:13
 Operator : CG/JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDC00.4

Quant Time: Jun 16 12:49:55 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 16 08:49:35 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	112	0.00
2	1,4-Dioxane	0.392	0.419	-6.9	114	0.00
3	n-Nitrosodimethylamine	0.228	0.192	15.8	107	0.00
4 S	2-Fluorophenol	0.772	0.894	-15.8	126	0.00
5 S	Phenol-d6	0.910	1.028	-13.0	126	0.00
6	bis(2-Chloroethyl)ether	0.887	0.948	-6.9	119	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	117	-0.01
8 S	Nitrobenzene-d5	0.296	0.299	-1.0	114	0.00
9	Naphthalene	1.005	1.063	-5.8	118	0.00
10	Hexachlorobutadiene	0.248	0.246	0.8	108	-0.01
11 SURR	2-Methylnaphthalene-d10	0.620	0.644	-3.9	116	0.00
12	2-Methylnaphthalene	0.673	0.702	-4.3	119	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	118	0.00
14 S	2,4,6-Tribromophenol	0.143	0.142	0.7	121	0.00
15 S	2-Fluorobiphenyl	1.578	1.619	-2.6	113	0.00
16	Acenaphthylene	1.639	1.704	-4.0	122	0.00
17	Acenaphthene	1.117	1.174	-5.1	119	0.00
18	Fluorene	1.441	1.492	-3.5	119	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	120	0.00
20	4-Bromophenyl-phenylether	0.240	0.230	4.2	112	0.00
21	Hexachlorobenzene	0.251	0.254	-1.2	116	0.00
22	Pentachlorophenol	0.084	0.085	-1.2	125	0.00
23	Phenanthrene	1.093	1.105	-1.1	118	0.00
24	Anthracene	0.912	0.921	-1.0	122	0.00
25 SURR	Fluoranthene-d10	1.104	1.134	-2.7	122	0.00
26	Fluoranthene	1.240	1.259	-1.5	122	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	124	0.00
28	Pvrene	1.323	1.367	-3.3	123	0.00
29 S	Terphenyl-d14	0.936	0.951	-1.6	120	0.00
30	Benzo(a)anthracene	1.176	1.158	1.5	123	0.00
31	Chrysene	1.370	1.470	-7.3	123	0.00
32	Bis(2-ethylhexyl)phthalate	0.392	0.429	-9.4	139	0.00
33	Indeno(1,2,3-cd)pyrene	1.531	1.549	-1.2	120	0.00
34 I	Perylene-d12	1.000	1.000	0.0	127	0.00
35	Benzo(b)fluoranthene	1.360	1.396	-2.6	126	0.00
36	Benzo(k)fluoranthene	1.449	1.432	1.2	119	0.00
37 C	Benzo(a)pyrene	1.167	1.207	-3.4	131	0.00
38	Dibenzo(a,h)anthracene	1.330	1.281	3.7	120	0.00
39	Benzo(a,h,i)perylene	1.391	1.378	0.9	118	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061620\
Data File : BN010903.D
Acq On : 16 Jun 2020 12:13
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 16 12:49:55 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 16 08:49:35 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			