

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061620\
 Data File : BN010919.D
 Acq On : 16 Jun 2020 22:27
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 17 03:37:45 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 16 08:49:35 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	118	0.00
2	1,4-Dioxane	0.392	0.500	-27.6#	144	0.00
3	n-Nitrosodimethylamine	0.228	0.225	1.3	132	0.02
4 S	2-Fluorophenol	0.772	1.026	-32.9#	153#	0.00
5 S	Phenol-d6	0.910	1.114	-22.4	145	0.00
6	bis(2-Chloroethyl)ether	0.887	1.065	-20.1	142	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	126	-0.01
8 S	Nitrobenzene-d5	0.296	0.315	-6.4	129	0.00
9	Naphthalene	1.005	1.070	-6.5	128	0.00
10	Hexachlorobutadiene	0.248	0.251	-1.2	118	-0.01
11 SURR	2-Methylnaphthalene-d10	0.620	0.707	-14.0	137	0.00
12	2-Methylnaphthalene	0.673	0.809	-20.2	148	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	128	0.00
14 S	2,4,6-Tribromophenol	0.143	0.141	1.4	130	0.00
15 S	2-Fluorobiphenyl	1.578	1.698	-7.6	129	0.00
16	Acenaphthylene	1.639	1.707	-4.1	132	0.00
17	Acenaphthene	1.117	1.175	-5.2	129	-0.01
18	Fluorene	1.441	1.510	-4.8	130	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	128	-0.01
20	4-Bromophenyl-phenylether	0.240	0.237	1.3	123	0.00
21	Hexachlorobenzene	0.251	0.258	-2.8	126	0.00
22	Pentachlorophenol	0.084	0.081	3.6	128	0.00
23	Phenanthrene	1.093	1.142	-4.5	130	-0.01
24	Anthracene	0.912	0.931	-2.1	132	0.00
25 SURR	Fluoranthene-d10	1.104	1.273	-15.3	146	0.00
26	Fluoranthene	1.240	1.445	-16.5	149	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	132	-0.01
28	Pvrene	1.323	1.518	-14.7	146	0.00
29 S	Terphenyl-d14	0.936	1.039	-11.0	140	0.00
30	Benzo(a)anthracene	1.176	1.223	-4.0	139	-0.01
31	Chrysene	1.370	1.481	-8.1	132	-0.01
32	Bis(2-ethylhexyl)phthalate	0.392	0.435	-11.0	150	-0.01
33	Indeno(1,2,3-cd)pyrene	1.531	1.849	-20.8	152#	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	141	-0.01
35	Benzo(b)fluoranthene	1.360	1.372	-0.9	137	-0.01
36	Benzo(k)fluoranthene	1.449	1.484	-2.4	137	0.00
37 C	Benzo(a)pyrene	1.167	1.217	-4.3	146	-0.01
38	Dibenzo(a,h)anthracene	1.330	1.460	-9.8	152#	-0.02
39	Benzo(a,h,i)perylene	1.391	1.382	0.6	132	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061620\
Data File : BN010919.D
Acq On : 16 Jun 2020 22:27
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 17 03:37:45 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 16 08:49:35 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			