

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061620\
 Data File : BN010921.D
 Acq On : 17 Jun 2020 00:55
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 17 03:52:27 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 16 08:49:35 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	126	0.00
2	1,4-Dioxane	0.392	0.430	-9.7	133	0.00
3	n-Nitrosodimethylamine	0.228	0.224	1.8	141	0.00
4 S	2-Fluorophenol	0.772	0.854	-10.6	137	0.00
5 S	Phenol-d6	0.910	0.984	-8.1	137	0.00
6	bis(2-Chloroethyl)ether	0.887	0.900	-1.5	128	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	123	-0.01
8 S	Nitrobenzene-d5	0.296	0.313	-5.7	126	0.00
9	Naphthalene	1.005	1.075	-7.0	126	0.00
10	Hexachlorobutadiene	0.248	0.252	-1.6	116	-0.01
11 SURR	2-Methylnaphthalene-d10	0.620	0.638	-2.9	121	0.00
12	2-Methylnaphthalene	0.673	0.706	-4.9	126	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	124	0.00
14 S	2,4,6-Tribromophenol	0.143	0.141	1.4	126	0.00
15 S	2-Fluorobiphenyl	1.578	1.698	-7.6	125	0.00
16	Acenaphthylene	1.639	1.699	-3.7	127	0.00
17	Acenaphthene	1.117	1.180	-5.6	126	-0.02
18	Fluorene	1.441	1.485	-3.1	124	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	120	0.00
20	4-Bromophenyl-phenylether	0.240	0.244	-1.7	119	0.00
21	Hexachlorobenzene	0.251	0.265	-5.6	121	0.00
22	Pentachlorophenol	0.084	0.089	-6.0	131	0.00
23	Phenanthrene	1.093	1.157	-5.9	124	0.00
24	Anthracene	0.912	0.951	-4.3	126	0.00
25 SURR	Fluoranthene-d10	1.104	1.130	-2.4	122	0.00
26	Fluoranthene	1.240	1.291	-4.1	125	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	127	0.00
28	Pvrene	1.323	1.357	-2.6	126	0.00
29 S	Terphenyl-d14	0.936	0.992	-6.0	129	0.00
30	Benzo(a)anthracene	1.176	1.213	-3.1	133	0.00
31	Chrysene	1.370	1.590	-16.1	137	0.00
32	Bis(2-ethylhexyl)phthalate	0.392	0.419	-6.9	139	0.00
33	Indeno(1,2,3-cd)pyrene	1.531	1.650	-7.8	131	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	134	-0.01
35	Benzo(b)fluoranthene	1.360	1.428	-5.0	136	-0.01
36	Benzo(k)fluoranthene	1.449	1.506	-3.9	133	-0.01
37 C	Benzo(a)pyrene	1.167	1.208	-3.5	138	-0.02
38	Dibenzo(a,h)anthracene	1.330	1.317	1.0	131	-0.02
39	Benzo(a,h,i)perylene	1.391	1.380	0.8	125	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN061620\
Data File : BN010921.D
Acq On : 17 Jun 2020 00:55
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 17 03:52:27 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 16 08:49:35 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			