

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN061620\
 Data File : BN010937.D
 Acq On : 17 Jun 2020 11:05
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 17 11:35:24 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 16 08:49:35 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	138	0.00
2	1,4-Dioxane	0.392	0.432	-10.2	146	0.00
3	n-Nitrosodimethylamine	0.228	0.191	16.2	132	0.00
4 S	2-Fluorophenol	0.772	0.888	-15.0	156#	0.00
5 S	Phenol-d6	0.910	0.955	-4.9	145	0.00
6	bis(2-Chloroethyl)ether	0.887	0.866	2.4	135	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	132	0.01
8 S	Nitrobenzene-d5	0.296	0.318	-7.4	136	0.01
9	Naphthalene	1.005	1.078	-7.3	135	0.01
10	Hexachlorobutadiene	0.248	0.261	-5.2	129	0.01
11 SURR	2-Methylnaphthalene-d10	0.620	0.637	-2.7	129	0.00
12	2-Methylnaphthalene	0.673	0.715	-6.2	137	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	131	0.02
14 S	2,4,6-Tribromophenol	0.143	0.142	0.7	135	0.00
15 S	2-Fluorobiphenyl	1.578	1.713	-8.6	134	0.00
16	Acenaphthylene	1.639	1.714	-4.6	136	0.00
17	Acenaphthene	1.117	1.181	-5.7	133	0.00
18	Fluorene	1.441	1.498	-4.0	133	0.02
19 I	Phenanthrene-d10	1.000	1.000	0.0	137	0.00
20	4-Bromophenyl-phenylether	0.240	0.231	3.7	129	0.00
21	Hexachlorobenzene	0.251	0.252	-0.4	131	0.00
22	Pentachlorophenol	0.084	0.080	4.8	134	0.00
23	Phenanthrene	1.093	1.099	-0.5	134	0.00
24	Anthracene	0.912	0.900	1.3	136	0.00
25 SURR	Fluoranthene-d10	1.104	1.061	3.9	130	0.00
26	Fluoranthene	1.240	1.211	2.3	134	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	137	0.00
28	Pvrene	1.323	1.357	-2.6	135	0.00
29 S	Terphenyl-d14	0.936	0.982	-4.9	137	0.00
30	Benzo(a)anthracene	1.176	1.197	-1.8	140	0.00
31	Chrysene	1.370	1.423	-3.9	131	0.00
32	Bis(2-ethylhexyl)phthalate	0.392	0.425	-8.4	151#	0.00
33	Indeno(1,2,3-cd)pyrene	1.531	1.637	-6.9	140	0.00
34 I	Perylene-d12	1.000	1.000	0.0	149	0.00
35	Benzo(b)fluoranthene	1.360	1.330	2.2	141	0.00
36	Benzo(k)fluoranthene	1.449	1.468	-1.3	143	0.00
37 C	Benzo(a)pyrene	1.167	1.167	0.0	148	0.00
38	Dibenzo(a,h)anthracene	1.330	1.262	5.1	139	0.00
39	Benzo(a,h,i)perylene	1.391	1.374	1.2	138	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN061620\
Data File : BN010937.D
Acq On : 17 Jun 2020 11:05
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 17 11:35:24 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 16 08:49:35 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			