

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061623\
 Data File : BN025882.D
 Acq On : 16 Jun 2023 14:23
 Operator : MA/JU
 Sample : 03173-12DL 10X
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 OW-07B-62.5-67.5-061223DL

Quant Time: Jun 16 15:10:18 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN053123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 16 14:10:48 2023
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/19/2023
 Supervised By :mohammad ahmed 06/20/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.005	152	4796	0.400	ng	0.00
7) Naphthalene-d8	10.825	136	11546	0.400	ng	0.01
13) Acenaphthene-d10	14.641	164	6516	0.400	ng	0.01
19) Phenanthrene-d10	17.393	188	12998	0.400	ng	0.01
29) Chrysene-d12	21.569	240	9755	0.400	ng	0.00
35) Perylene-d12	24.031	264	10054	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.571	112	163	0.012	ng	0.03
5) Phenol-d6	7.239	99	73	0.004	ng	0.09
8) Nitrobenzene-d5	9.244	82	343m	0.032	ng	0.07
11) 2-Methylnaphthalene-d10	12.453	152	558m	0.034	ng	0.06
14) 2,4,6-Tribromophenol	16.152	330	79	0.039	ng	0.02
15) 2-Fluorobiphenyl	13.304	172	908m	0.033	ng	0.04
27) Fluoranthene-d10	19.421	212	1436	0.054	ng	0.01
31) Terphenyl-d14	20.006	244	1132	0.048	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.383	88	17027	3.026	ng	Qvalue # 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061623\
Data File : BN025882.D
Acq On : 16 Jun 2023 14:23
Operator : MA/JU
Sample : 03173-12DL 10X
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
OW-07B-62.5-67.5-061223DL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/19/2023
Supervised By :mohammad ahmed 06/20/2023

Quant Time: Jun 16 15:10:18 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN053123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 16 14:10:48 2023
Response via : Initial Calibration

