

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061623\
 Data File : BN025885.D
 Acq On : 16 Jun 2023 18:30
 Operator : MA/JU
 Sample : 03299-01
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 GWOW-BR-05-190-210-061523

Quant Time: Jun 17 02:35:12 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN053123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 16 14:10:48 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.005	152	4630	0.400	ng	0.00
7) Naphthalene-d8	10.814	136	12607	0.400	ng	0.00
13) Acenaphthene-d10	14.630	164	6754	0.400	ng	0.00
19) Phenanthrene-d10	17.381	188	14794	0.400	ng	# 0.00
29) Chrysene-d12	21.560	240	10706	0.400	ng	0.00
35) Perylene-d12	24.040	264	11036	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.557	112	1424	0.109	ng	0.01
5) Phenol-d6	7.167	99	1008	0.063	ng	0.01
8) Nitrobenzene-d5	9.180	82	3432	0.293	ng	0.01
11) 2-Methylnaphthalene-d10	12.400	152	6498	0.358	ng	0.00
14) 2,4,6-Tribromophenol	16.127	330	869	0.417	ng	0.00
15) 2-Fluorobiphenyl	13.261	172	10469	0.367	ng	0.00
27) Fluoranthene-d10	19.402	212	14926	0.490	ng	0.00
31) Terphenyl-d14	19.992	244	11344	0.437	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.383	88	187769	34.566	ng	# 93
25) Phenanthrene	17.418	178	1807	0.040	ng	98
28) Fluoranthene	19.435	202	2318	0.052	ng	# 73
30) Pyrene	19.797	202	1302	0.021	ng	96
34) Bis(2-ethylhexyl)phtha...	21.453	149	10077	0.409	ng	99
36) Indeno(1,2,3-cd)pyrene	26.656	276	966	0.020	ng	# 61

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061623\
Data File : BN025885.D
Acq On : 16 Jun 2023 18:30
Operator : MA/JU
Sample : 03299-01
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
GWOW-BR-05-190-210-061523

Quant Time: Jun 17 02:35:12 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN053123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 16 14:10:48 2023
Response via : Initial Calibration

