

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN061720\
 Data File : BN010939.D
 Acq On : 17 Jun 2020 12:19
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 17 13:15:59 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 17 13:14:12 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	110	0.00
2	1,4-Dioxane	0.392	0.425	-8.4	114	0.00
3	n-Nitrosodimethylamine	0.228	0.222	2.6	122	0.00
4 S	2-Fluorophenol	0.772	0.990	-28.2#	138	0.00
5 S	Phenol-d6	0.910	1.025	-12.6	124	0.00
6	bis(2-Chloroethyl)ether	0.887	0.974	-9.8	120	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	114	0.00
8 S	Nitrobenzene-d5	0.296	0.308	-4.1	114	0.00
9	Naphthalene	1.005	1.059	-5.4	114	0.00
10	Hexachlorobutadiene	0.248	0.255	-2.8	109	0.00
11 SURR	2-Methylnaphthalene-d10	0.620	0.641	-3.4	112	0.00
12	2-Methylnaphthalene	0.673	0.710	-5.5	117	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	114	0.00
14 S	2,4,6-Tribromophenol	0.143	0.140	2.1	115	0.00
15 S	2-Fluorobiphenyl	1.578	1.780	-12.8	121	0.00
16	Acenaphthylene	1.639	1.718	-4.8	119	0.00
17	Acenaphthene	1.117	1.181	-5.7	116	0.00
18	Fluorene	1.441	1.512	-4.9	117	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	116	0.00
20	4-Bromophenyl-phenylether	0.240	0.235	2.1	111	0.00
21	Hexachlorobenzene	0.251	0.257	-2.4	113	0.00
22	Pentachlorophenol	0.084	0.086	-2.4	123	0.00
23	Phenanthrene	1.093	1.127	-3.1	117	0.00
24	Anthracene	0.912	0.926	-1.5	119	0.00
25 SURR	Fluoranthene-d10	1.104	1.150	-4.2	120	0.00
26	Fluoranthene	1.240	1.282	-3.4	120	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	123	0.00
28	Pvrene	1.323	1.372	-3.7	122	0.00
29 S	Terphenyl-d14	0.936	0.971	-3.7	122	0.00
30	Benzo(a)anthracene	1.176	1.207	-2.6	127	0.00
31	Chrysene	1.370	1.458	-6.4	121	0.00
32	Bis(2-ethylhexyl)phthalate	0.392	0.413	-5.4	132	0.00
33	Indeno(1,2,3-cd)pyrene	1.531	1.579	-3.1	121	0.00
34 I	Perylene-d12	1.000	1.000	0.0	146	0.00
35	Benzo(b)fluoranthene	1.360	1.189	12.6	123	0.00
36	Benzo(k)fluoranthene	1.449	1.311	9.5	125	0.00
37 C	Benzo(a)pyrene	1.167	1.020	12.6	127	0.00
38	Dibenzo(a,h)anthracene	1.330	1.108	16.7	119	0.00
39	Benzo(a,h,i)perylene	1.391	1.200	13.7	118	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN061720\
Data File : BN010939.D
Acq On : 17 Jun 2020 12:19
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 17 13:15:59 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 17 13:14:12 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			