

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN061720\
 Data File : BN010952.D
 Acq On : 17 Jun 2020 21:58
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 17 22:32:25 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 17 13:14:12 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	138	0.00
2	1,4-Dioxane	0.400	0.441	-10.2	145	0.00
3	n-Nitrosodimethylamine	0.400	0.420	-5.0	165	-0.02
4 S	2-Fluorophenol	0.400	0.462	-15.5	156	0.00
5 S	Phenol-d6	0.400	0.493	-23.2	170	0.00
6	bis(2-Chloroethyl)ether	0.400	0.470	-17.5	162	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	139	0.00
8 S	Nitrobenzene-d5	0.400	0.429	-7.2	143	-0.01
9	Naphthalene	0.400	0.430	-7.5	142	-0.01
10	Hexachlorobutadiene	0.400	0.396	1.0	128	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.455	-13.7	151	0.00
12	2-Methylnaphthalene	0.400	0.459	-14.8	155	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	143	0.00
14 S	2,4,6-Tribromophenol	0.400	0.397	0.8	147	0.00
15 S	2-Fluorobiphenyl	0.400	0.421	-5.2	141	0.00
16	Acenaphthylene	0.400	0.423	-5.7	150	0.00
17	Acenaphthene	0.400	0.419	-4.7	144	0.00
18	Fluorene	0.400	0.417	-4.2	145	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	143	0.00
20	4-Bromophenyl-phenylether	0.400	0.408	-2.0	143	0.00
21	Hexachlorobenzene	0.400	0.413	-3.2	140	0.00
22	Pentachlorophenol	0.400	0.405	-1.3	149	0.00
23	Phenanthrene	0.400	0.421	-5.2	147	0.00
24	Anthracene	0.400	0.417	-4.2	150	0.00
25 SURR	Fluoranthene-d10	0.400	0.461	-15.3	163	0.00
26	Fluoranthene	0.400	0.460	-15.0	164	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	153	0.00
28	Pvrene	0.400	0.412	-3.0	151	0.00
29 S	Terphenyl-d14	0.400	0.410	-2.5	150	0.00
30	Benzo(a)anthracene	0.400	0.424	-6.0	164	0.00
31	Chrysene	0.400	0.407	-1.7	144	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.461	-15.3	180	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.434	-8.5	158	0.00
34 I	Perylene-d12	0.400	0.400	0.0	165	0.00
35	Benzo(b)fluoranthene	0.400	0.382	4.5	152	0.00
36	Benzo(k)fluoranthene	0.400	0.414	-3.5	162	0.00
37 C	Benzo(a)pyrene	0.400	0.402	-0.5	165	0.00
38	Dibenzo(a,h)anthracene	0.400	0.383	4.3	155	0.00
39	Benzo(a,h,i)perylene	0.400	0.382	4.5	148	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN061720\
Data File : BN010952.D
Acq On : 17 Jun 2020 21:58
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 17 22:32:25 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 17 13:14:12 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				