


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061819\
Data File : BN006368.D
Acq On : 18 Jun 2019 08:37
Operator : JU/SJ
Sample : DFTPP94
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP94

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN053019.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Mon Jun 17 11:13:07 2019



AutoFind: Scans 2466, 2467, 2468; Background Corrected with Scan 2459

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.6	189757	PASS
68	69	0.00	2	1.9	3502	PASS
69	198	0.00	100	36.3	182920	PASS
70	69	0.00	2	0.6	1146	PASS
127	198	10	80	44.9	226730	PASS
197	198	0.00	2	0.1	610	PASS
198	198	100	100	100.0	504490	PASS
199	198	5	9	6.8	34061	PASS
275	198	10	60	26.6	134416	PASS
365	198	1	100	3.3	16627	PASS
441	198	0.01	100	14.5	73360	PASS
442	198	50	100	96.7	487616	PASS
443	442	15	24	18.9	92216	PASS

Scan 2459 (17.451 min): BN006368.D
DFTPP94

m/z	Abundance
39.90	2858.0
41.00	397.0
44.00	3946.0
72.90	321.0
207.00	1314.0

Instrument :
BNA_N
ClientSampleId :
DFTPP94

m/z	Abundance
37.10	924.0
38.00	2751.0
39.00	16504.0
40.00	2898.0
41.00	617.0
44.00	3132.0
45.00	455.0
49.00	1836.0
50.00	59880.0
51.00	239872.0
52.00	12595.0
53.00	445.0
55.00	1416.0
56.00	6471.0
57.00	15887.0
58.00	687.0
61.00	2921.0
62.00	3052.0
63.00	9429.0
64.00	1528.0
65.00	4788.0
65.90	358.0
67.20	422.0
68.00	4398.0
69.00	226880.0
70.00	1412.0
73.00	2334.0
74.00	23976.0
75.00	37832.0
76.10	13175.0
77.00	276736.0
78.00	19560.0
79.00	16616.0
80.00	13121.0
81.00	18248.0
82.00	5153.0
83.00	4387.0
83.90	611.0
85.00	4680.0
86.00	5567.0
87.00	2475.0
88.00	984.0
88.90	637.0
91.00	4201.0
92.00	4733.0
93.00	29712.0
94.00	2385.0
94.90	539.0
96.00	1370.0
97.00	729.0
98.00	23312.0
99.00	19408.0
100.00	1824.0
101.00	10301.0
102.00	642.0
103.00	3152.0
104.00	6585.0
105.00	6042.0
106.00	2081.0
107.00	78048.0
108.00	12764.0
109.00	2443.0
110.00	146624.0
111.00	22264.0
112.00	2654.0
112.90	745.0
115.00	410.0
116.00	4401.0
117.00	62176.0
118.00	5120.0
119.10	563.0
119.90	966.0
122.00	4884.0
123.00	7879.0
124.00	3548.0
125.00	3369.0
127.00	273728.0
128.00	20496.0
129.00	99528.0
130.00	8307.0
131.00	1753.0
132.00	866.0
133.00	403.0
134.00	2960.0
135.00	8184.0
136.00	3349.0
137.00	4652.0
138.00	1116.0
139.00	556.0
140.00	1359.0
141.00	12822.0

Instrument :
BNA_N
ClientSampleId :
DFTPP94

142.00	3991.0
143.00	2999.0
144.00	799.0
145.00	805.0
145.90	2534.0
147.00	7030.0
148.00	16377.0
149.00	3364.0
150.00	1042.0
151.10	1335.0
151.50	1050.0
153.00	3994.0
154.00	2996.0
155.00	7437.0
156.00	11281.0
157.10	2452.0
157.90	2416.0
159.00	1664.0
160.00	4400.0
161.00	6288.0
161.90	1697.0
163.00	547.0
164.00	600.0
165.00	5060.0
166.00	3791.0
167.00	26544.0
168.00	17064.0
169.00	2396.0
170.00	597.0
170.90	1140.0
172.00	2431.0
173.00	3103.0
174.00	5525.0
175.00	10453.0
176.00	3099.0
177.00	4668.0
178.00	1727.0
179.00	19240.0
180.00	15166.0
181.00	6423.0
182.00	965.0
183.00	603.0
184.00	1510.0
185.00	9409.0
186.00	77800.0
187.00	22312.0
188.00	2592.0
189.00	4787.0
189.90	741.0
191.00	2162.0
192.00	6806.0
193.00	7236.0
194.00	1521.0
195.10	711.0
196.00	17344.0
198.00	573376.0
199.00	40080.0
200.00	3020.0
201.50	2857.0
203.00	3717.0
204.00	19624.0
205.00	34784.0
206.00	145024.0
207.00	20296.0
208.00	4702.0
209.00	1654.0
210.20	2457.0
211.00	5184.0
211.70	1375.0
212.90	486.0
215.00	1537.0
216.00	2817.0
217.00	35992.0
218.00	5266.0
219.00	536.0
221.00	24656.0
223.00	8440.0
224.00	78952.0
225.00	20312.0
226.10	2516.0
227.00	31768.0
228.00	4576.0
229.00	6503.0
230.00	860.0
231.00	2689.0
232.00	590.0
233.00	498.0
234.00	2041.0
235.00	2451.0
236.00	1475.0
237.00	3392.0
238.00	553.0
239.00	1219.0
240.00	1044.0
240.90	1829.0

Instrument :
BNA_N
ClientSampleId :
DFTPP94

242.00	4331.0
243.10	4219.0
244.00	63112.0
245.00	8683.0
246.00	11547.0
247.00	2280.0
247.90	565.0
249.00	2277.0
250.00	401.0
250.90	395.0
252.00	521.0
253.00	1150.0
255.00	306432.0
256.00	43984.0
257.00	4237.0
258.00	15604.0
259.00	2732.0
260.00	554.0
261.00	658.0
263.90	463.0
265.00	6933.0
265.90	881.0
270.90	762.0
272.00	785.0
273.00	9504.0
274.00	24512.0
275.00	141440.0
276.00	19144.0
277.00	11179.0
278.00	1995.0
279.00	477.0
283.00	1232.0
283.90	736.0
285.00	1865.0
289.00	352.0
290.00	396.0
292.00	494.0
293.00	2967.0
294.00	761.0
296.00	35792.0
297.00	5141.0
298.00	444.0
301.00	493.0
302.10	504.0
303.00	4126.0
304.00	1447.0
305.00	313.0
308.00	627.0
308.90	327.0
310.00	528.0
312.90	419.0
314.00	1952.0
315.00	3808.0
316.00	2258.0
317.00	395.0
321.00	1129.0
322.10	535.0
323.00	12526.0
324.00	2518.0
327.00	2163.0
328.00	1128.0
331.90	947.0
333.00	1365.0
334.00	7593.0
335.00	2002.0
341.00	1017.0
341.90	375.0
346.00	2641.0
346.90	712.0
352.00	3910.0
352.90	2810.0
354.00	3736.0
355.00	870.0
365.00	16076.0
366.00	2600.0
371.00	883.0
372.00	7092.0
373.00	1773.0
383.00	1711.0
383.90	581.0
390.00	808.0
390.90	691.0
392.00	561.0
401.00	349.0
402.00	2392.0
403.00	4089.0
404.00	1529.0
405.00	460.0
421.00	3275.0
422.00	3311.0
423.00	25520.0
424.00	6837.0
425.00	943.0
441.00	65264.0
442.00	427648.0

Instrument :
BNA_N
ClientSampleId :
DFTPP94

443.00	85912.0
444.00	7754.0
445.00	523.0

Instrument :
BNA_N
ClientSampleId :
DFTPP94

m/z	Abundance
35.90	312.0
37.00	820.0
38.10	2721.0
39.00	15667.0
40.00	2800.0
41.10	797.0
44.00	3034.0
44.90	449.0
49.00	1834.0
50.00	53832.0
51.00	216576.0
52.00	10425.0
53.00	462.0
55.00	1517.0
56.00	6365.0
57.00	14147.0
57.90	536.0
61.00	2388.0
62.00	2940.0
63.00	8702.0
64.00	1135.0
65.00	4136.0
66.90	445.0
68.00	3909.0
69.00	210688.0
70.00	1286.0
73.10	2080.0
74.00	22864.0
75.00	34480.0
76.00	11537.0
77.00	255040.0
78.00	17304.0
79.00	16332.0
80.00	12837.0
81.00	16608.0
82.00	4205.0
83.00	4309.0
83.90	582.0
85.00	4712.0
86.00	5147.0
87.00	2439.0
88.00	988.0
89.00	518.0
91.00	3829.0
92.00	4312.0
93.00	26464.0
94.00	1806.0
95.00	449.0
96.00	1336.0
97.10	637.0
98.00	21000.0
99.00	17896.0
100.00	1519.0
101.00	9643.0
102.00	626.0
102.90	3427.0
104.00	5904.0
105.00	6248.0
106.10	2087.0
107.00	73736.0
108.00	11471.0
109.00	2405.0
110.00	138304.0
111.00	20992.0
112.00	2651.0
112.90	822.0
114.90	390.0
116.00	3814.0
117.00	60072.0
118.00	4652.0
119.00	621.0
120.00	928.0
120.80	480.0
122.00	4509.0
123.00	7277.0
124.00	3435.0
124.90	3177.0
127.00	259712.0
128.00	19064.0
129.00	94624.0
130.00	8515.0
131.00	1864.0
131.90	1113.0
133.10	440.0
134.00	2907.0
135.00	7624.0
136.00	2988.0
137.00	4425.0
137.90	1115.0
139.00	583.0
140.00	1312.0

Instrument :
BNA_N
ClientSampleId :
DFTPP94

141.00	11886.0
142.00	4319.0
142.90	3179.0
144.00	726.0
145.00	902.0
146.00	2378.0
146.90	7005.0
148.00	15302.0
149.00	3090.0
150.00	897.0
151.00	1815.0
151.90	1235.0
153.00	4076.0
154.00	3428.0
155.00	6917.0
156.00	10229.0
157.10	1890.0
158.00	2608.0
159.00	1439.0
160.00	4150.0
161.00	6317.0
162.00	1978.0
162.90	576.0
163.90	760.0
165.00	4883.0
166.00	3944.0
167.00	28064.0
168.00	15169.0
169.00	2527.0
170.00	872.0
170.90	1266.0
171.90	2259.0
173.00	3087.0
174.00	5342.0
175.00	10956.0
176.00	3381.0
177.00	4995.0
177.90	1710.0
179.00	19000.0
180.00	14710.0
181.00	6841.0
182.10	980.0
183.00	647.0
184.00	1751.0
185.00	9121.0
186.00	78208.0
187.00	21512.0
188.00	2426.0
189.00	4570.0
190.00	792.0
191.00	1991.0
192.00	6514.0
193.00	7422.0
194.00	1572.0
195.00	873.0
196.00	16314.0
198.00	585984.0
199.00	39552.0
200.00	3132.0
201.50	2472.0
203.00	3798.0
204.00	19688.0
205.00	33064.0
206.00	147264.0
207.10	19944.0
208.00	5441.0
209.00	1687.0
211.00	5370.0
213.10	386.0
215.00	1441.0
216.00	3188.0
217.00	37912.0
218.00	4632.0
219.00	462.0
221.00	26648.0
221.80	5493.0
223.00	8248.0
224.00	82216.0
225.00	21480.0
226.00	2377.0
227.00	32112.0
228.00	5116.0
229.00	7452.0
229.90	859.0
231.00	3477.0
231.90	570.0
233.00	586.0
234.00	2248.0
235.00	2379.0
235.90	1505.0
237.00	3122.0
238.00	344.0
239.00	1525.0
239.90	890.0
241.00	2108.0

Instrument :

BNA_N

ClientSampleId :

DFTPP94

242.00	4272.0
243.00	4881.0
244.00	66136.0
245.00	8558.0
246.00	11547.0
247.00	2529.0
247.90	517.0
249.00	2321.0
250.10	366.0
251.00	510.0
252.00	528.0
253.00	1569.0
255.00	326848.0
256.00	49544.0
257.00	3726.0
258.00	16896.0
259.00	2899.0
260.00	543.0
260.90	611.0
265.00	7472.0
266.00	1000.0
270.00	378.0
271.00	714.0
272.00	805.0
273.00	10505.0
274.00	27656.0
275.00	159488.0
276.00	20840.0
277.00	11827.0
278.00	2148.0
279.00	442.0
282.00	419.0
283.10	1003.0
284.00	905.0
285.10	2199.0
286.00	505.0
288.90	389.0
289.90	375.0
291.00	392.0
292.20	581.0
293.00	3194.0
294.00	754.0
296.00	40512.0
297.00	5231.0
298.10	380.0
300.80	437.0
302.00	771.0
303.00	4799.0
304.00	1415.0
307.90	620.0
308.90	364.0
310.00	543.0
312.90	387.0
314.00	1921.0
315.00	4190.0
316.00	2777.0
316.90	485.0
321.10	1484.0
322.00	749.0
323.10	13967.0
324.00	2704.0
326.00	412.0
327.00	2466.0
327.90	1549.0
332.00	889.0
333.00	1583.0
334.00	9288.0
335.00	2406.0
335.90	433.0
341.00	1124.0
342.10	506.0
345.90	3322.0
347.00	556.0
351.00	328.0
352.00	4350.0
353.00	3157.0
354.00	5033.0
355.00	986.0
359.00	329.0
365.00	19944.0
365.90	2586.0
370.00	387.0
371.10	1010.0
372.00	7901.0
373.00	2466.0
383.00	2431.0
384.00	661.0
385.10	331.0
390.00	1144.0
391.00	866.0
392.00	666.0
401.00	569.0
402.00	3298.0
403.00	5083.0
404.00	1940.0

Instrument :
BNA_N
ClientSampleId :
DFTPP94

405.00	415.0
421.00	4635.0
422.00	4730.0
423.00	32872.0
424.00	8680.0
425.00	1006.0
441.10	88064.0
442.00	589888.0
443.00	112432.0
444.10	10233.0
445.00	638.0

Instrument :
BNA_N
ClientSampleId :
DFTPP94

m/z	Abundance
37.00	425.0
38.00	1325.0
39.00	7856.0
40.00	3066.0
41.00	836.0
43.00	455.0
44.00	3753.0
48.90	921.0
50.00	27272.0
51.00	112824.0
52.00	6150.0
55.00	1024.0
56.00	3294.0
57.00	7608.0
58.00	312.0
61.00	1363.0
62.00	1499.0
63.00	4383.0
64.00	673.0
65.00	2426.0
67.00	322.0
68.00	2201.0
69.00	111192.0
70.00	741.0
73.00	1178.0
74.00	11749.0
75.00	17656.0
76.00	5934.0
77.00	135488.0
78.00	9732.0
79.00	8153.0
80.00	6359.0
81.00	8860.0
82.00	2191.0
83.00	2467.0
84.00	546.0
85.00	2346.0
86.00	2716.0
87.00	1175.0
87.80	477.0
91.00	2405.0
92.00	2353.0
93.00	14458.0
94.00	884.0
95.00	342.0
96.10	907.0
97.10	583.0
98.00	11361.0
99.00	10696.0
100.00	760.0
101.00	4924.0
103.00	1755.0
104.00	3210.0
105.00	3300.0
106.00	1110.0
107.00	42296.0
108.00	5888.0
108.90	1190.0
110.00	75720.0
111.00	11892.0
111.90	1445.0
112.90	361.0
116.00	2271.0
117.00	33344.0
118.00	2871.0
119.00	548.0
120.00	514.0
122.00	2569.0
123.00	4277.0
124.00	2047.0
124.90	1789.0
126.10	759.0
127.00	146752.0
128.00	10382.0
129.00	53096.0
130.00	4147.0
131.00	847.0
131.90	524.0
132.90	428.0
134.00	1683.0
135.00	4631.0
136.00	1851.0
137.00	3042.0
137.90	567.0
139.90	744.0
141.00	7368.0
141.90	2224.0
143.00	1530.0
144.00	451.0
145.00	348.0
145.90	1156.0

Instrument :
BNA_N
ClientSampleId :
DFTPP94

147.00	4229.0
148.00	8617.0
149.00	1989.0
150.00	508.0
151.10	831.0
151.70	799.0
153.00	2556.0
154.00	2289.0
155.00	4687.0
156.00	5928.0
157.00	1254.0
157.90	1221.0
159.00	1017.0
160.00	2398.0
161.00	3837.0
162.10	1083.0
162.80	390.0
163.80	311.0
165.00	3012.0
166.00	2292.0
167.00	15382.0
168.00	9166.0
169.00	1327.0
169.90	484.0
170.90	713.0
172.00	1336.0
173.00	1945.0
174.00	3345.0
175.00	6486.0
176.00	1446.0
176.90	2792.0
178.00	1136.0
179.00	11961.0
180.00	8213.0
181.00	3843.0
181.90	482.0
183.00	356.0
184.00	934.0
185.00	6077.0
186.00	45608.0
187.00	13396.0
188.00	1314.0
188.90	2898.0
189.90	485.0
191.00	1470.0
192.00	4182.0
193.00	4704.0
194.00	1282.0
195.00	468.0
196.00	10474.0
196.80	1831.0
198.00	354112.0
199.00	22552.0
200.00	1967.0
200.90	554.0
201.40	1296.0
203.00	2403.0
204.00	11872.0
205.00	19800.0
206.00	90504.0
207.00	11905.0
208.00	3281.0
209.00	1102.0
210.00	1283.0
211.10	3800.0
215.00	1182.0
215.90	1751.0
217.00	22376.0
218.00	3558.0
219.00	402.0
221.00	15936.0
221.80	3407.0
223.00	4905.0
224.00	48808.0
225.00	12433.0
226.00	1451.0
227.00	21032.0
228.00	2976.0
229.00	4491.0
230.00	651.0
231.00	1858.0
232.80	367.0
233.90	1381.0
235.00	1818.0
236.00	969.0
237.00	1934.0
239.00	912.0
240.00	758.0
241.00	1272.0
242.00	3067.0
243.00	3163.0
244.00	41776.0
245.00	5199.0
246.00	7493.0
247.00	1685.0

Instrument :
BNA_N
ClientSampleId :
DFTPP94

247.90	351.0
249.00	1505.0
250.90	410.0
252.00	366.0
252.90	828.0
255.00	204608.0
256.00	29392.0
257.00	2471.0
258.00	11589.0
259.00	1682.0
264.10	422.0
265.00	4580.0
265.90	637.0
270.90	421.0
271.90	619.0
273.00	6365.0
274.00	18304.0
275.00	102320.0
276.00	13451.0
277.00	8194.0
277.90	1655.0
281.90	387.0
283.00	882.0
283.90	595.0
285.10	1344.0
288.90	316.0
292.00	441.0
293.00	1989.0
294.10	395.0
296.00	26960.0
297.00	3396.0
300.90	378.0
302.00	578.0
303.00	3039.0
304.00	963.0
308.00	499.0
308.90	355.0
309.90	358.0
314.00	1214.0
315.00	3253.0
316.00	1738.0
321.00	838.0
322.00	512.0
323.00	9442.0
324.00	1789.0
326.90	1713.0
327.90	875.0
332.00	786.0
333.10	749.0
334.00	6656.0
335.00	1461.0
341.00	1060.0
342.10	411.0
346.00	2207.0
346.90	471.0
352.00	2883.0
353.00	2182.0
354.00	3297.0
355.00	782.0
365.00	13862.0
366.00	2174.0
371.00	1046.0
372.00	5895.0
373.00	1643.0
383.00	1532.0
383.90	429.0
390.10	732.0
391.00	633.0
391.90	536.0
401.00	522.0
402.00	2499.0
403.00	4097.0
404.00	1469.0
421.00	2949.0
422.00	3255.0
423.00	24736.0
424.00	5921.0
424.90	679.0
439.00	311.0
439.70	338.0
441.00	66752.0
442.00	445312.0
443.00	78304.0
444.00	6840.0
445.00	467.0

Instrument :
BNA_N
ClientSampleId :
DFTPP94