

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061923\
 Data File : BN025905.D
 Acq On : 19 Jun 2023 17:53
 Operator : MA/JU
 Sample : 03259-02DL 10X
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 GWOW-BR-05-129-149-061423DL

Quant Time: Jun 19 18:40:48 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN053123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 16 14:10:48 2023
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/20/2023
 Supervised By :mohammad ahmed 06/20/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.012	152	4000	0.400	ng	0.01
7) Naphthalene-d8	10.835	136	9353	0.400	ng	# 0.02
13) Acenaphthene-d10	14.640	164	5765	0.400	ng	0.01
19) Phenanthrene-d10	17.393	188	11496	0.400	ng	0.01
29) Chrysene-d12	21.569	240	8382	0.400	ng	# 0.00
35) Perylene-d12	24.048	264	8766	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.585	112	157	0.014	ng	0.04
5) Phenol-d6	7.138	99	6	0.000	ng	-0.01
8) Nitrobenzene-d5	9.170	82	10	0.001	ng	0.00
11) 2-Methylnaphthalene-d10	12.472	152	273m	0.020	ng	0.08
14) 2,4,6-Tribromophenol	16.152	330	23	0.013	ng	0.02
15) 2-Fluorobiphenyl	13.315	172	291	0.012	ng	0.05
27) Fluoranthene-d10	19.430	212	561	0.024	ng	0.02
31) Terphenyl-d14	20.011	244	436	0.021	ng	0.01
Target Compounds						Qvalue
2) 1,4-Dioxane	3.383	88	14647	3.121	ng	# 94
34) Bis(2-ethylhexyl)phtha...	21.453	149	1244	0.065	ng	# 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061923\
Data File : BN025905.D
Acq On : 19 Jun 2023 17:53
Operator : MA/JU
Sample : 03259-02DL 10X
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
GWOW-BR-05-129-149-061423DL

Quant Time: Jun 19 18:40:48 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN053123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 16 14:10:48 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/20/2023
Supervised By :mohammad ahmed 06/20/2023

