

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061923\
 Data File : BN025907.D
 Acq On : 19 Jun 2023 19:11
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 20 04:17:17 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN053123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 16 14:10:48 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	65	0.00
2	1,4-Dioxane	0.400	0.420	-5.0	62	0.00
3	n-Nitrosodimethylamine	0.400	0.179	55.3#	27	0.02
4 S	2-Fluorophenol	0.400	0.326	18.5	48	0.00
5 S	Phenol-d6	0.400	0.302	24.5	45	0.00
6	bis(2-Chloroethyl)ether	0.400	0.347	13.3	51	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	56	0.00
8 S	Nitrobenzene-d5	0.400	0.336	16.0	46	0.01
9	Naphthalene	0.400	0.389	2.8	52	0.01
10	Hexachlorobutadiene	0.400	0.470	-17.5	62	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.384	4.0	50	0.01
12	2-Methylnaphthalene	0.400	0.383	4.3	50	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	52	0.00
14 S	2,4,6-Tribromophenol	0.400	0.327	18.3	43	0.01
15 S	2-Fluorobiphenyl	0.400	0.378	5.5	47	0.01
16	Acenaphthylene	0.400	0.387	3.3	49	0.00
17	Acenaphthene	0.400	0.401	-0.3	50	0.00
18	Fluorene	0.400	0.403	-0.8	48	0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	51	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.273	31.8#	35	0.01
21	4-Bromophenyl-phenylether	0.400	0.400	0.0	49	0.01
22	Hexachlorobenzene	0.400	0.457	-14.2	56	0.00
23	Atrazine	0.400	0.390	2.5	46	0.00
24	Pentachlorophenol	0.400	0.275	31.3#	35	0.00
25	Phenanthrene	0.400	0.398	0.5	48	0.00
26	Anthracene	0.400	0.379	5.3	46	0.01
27 SURR	Fluoranthene-d10	0.400	0.450	-12.5	52	0.00
28	Fluoranthene	0.400	0.458	-14.5	53	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	64	0.00
30	Pyrene	0.400	0.357	10.8	54	0.00
31 S	Terphenyl-d14	0.400	0.337	15.8	52	0.00
32	Benzo(a)anthracene	0.400	0.377	5.8	59	0.00
33	Chrysene	0.400	0.417	-4.2	62	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.463	-15.8	63	0.00
35 I	Perylene-d12	0.400	0.400	0.0	82	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.423	-5.7	88	0.00
37	Benzo(b)fluoranthene	0.400	0.377	5.8	71	0.00
38	Benzo(k)fluoranthene	0.400	0.404	-1.0	78	0.00
39 C	Benzo(a)pyrene	0.400	0.389	2.8	76	0.00
40	Dibenzo(a,h)anthracene	0.400	0.397	0.8	83	0.02
41	Benzo(g,h,i)perylene	0.400	0.464	-16.0	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4