

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062018\
 Data File : BN001886.D
 Acq On : 20 Jun 2018 14:33
 Operator : JU/SJ
 Sample : J3568-02
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DAXN8

Quant Time: Jun 20 16:48:40 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN061918.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jun 19 14:50:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.72	152	518544	20.00	ng/ul	0.00
18) Naphthalene-d8	10.49	136	2362484	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.35	164	1447958	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.09	188	3322921	20.00	ng/ul	0.00
75) Chrysene-d12	21.30	240	3367693	20.00	ng/ul	0.00
83) Perylene-d12	23.54	264	3081353	20.00	ng/ul	0.01

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.27	96	46976	3.93	ng/uL	0.00
5) Phenol-d5	7.06	99	766133	16.33	ng/ul	0.17
7) Bis-(2-Chloroethyl)ether-d	7.06	67	733632	25.34	ng/ul	0.00
9) 2-Chlorophenol-d4	7.25	132	931	0.03	ng/ul	0.00
13) 4-Methylphenol-d8	8.42	113	2903	0.08	ng/ul	0.00
19) Nitrobenzene-d5	8.86	128	153110	8.73	ng/ul	0.00
22) 2-Nitrophenol-d4	9.58	143	113256	6.24	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.10	165	293	0.01	ng/ul	-0.01
29) 4-Chloroaniline-d4	10.63	131	6896	0.17	ng/ul	0.00
43) Dimethylphthalate-d6	13.76	166	3217831	26.32	ng/ul	0.00
46) Acenaphthylene-d8	14.03	160	2006	0.01	ng/ul	0.00
51) 4-Nitrophenol-d4	14.55	143	2475	0.11	ng/ul	0.00
57) Fluorene-d10	15.34	176	46209	0.43	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.45	200	329062	18.18	ng/ul	0.00
70) Anthracene-d10	17.19	188	9118	0.06	ng/ul	0.00
76) Pyrene-d10	19.49	212	1376	0.01	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.54	264	3067266	18.30	ng/ul	0.15


Target Compounds					Qvalue	
44) Dimethylphthalate	13.80	163	1456330	12.118	ng/ul	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062018\
Data File : BN001886.D
Acq On : 20 Jun 2018 14:33
Operator : JU/SJ
Sample : J3568-02
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DAXN8

Quant Time: Jun 20 16:48:40 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN061918.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Jun 19 14:50:02 2018
Response via : Initial Calibration





#44

Dimethylphthalate

Concen: 12.118 ng/ul

RT: 13.80 min Scan# 1839

Delta R.T. -0.01 min

Lab File: BN001886.D


Acq: 20 Jun 2018 14:33

Instrument :

BNA_N

ClientSampleId :

DAXN8




Tgt Ion:163 Resp: 1456330

Ion Ratio Lower Upper

163 100

194 4.0 4.0 6.0

164 9.9 8.4 12.6



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E

