

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062023\
 Data File : BN025910.D
 Acq On : 20 Jun 2023 09:58
 Operator : MA/JU
 Sample : PB153492BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB153492BL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/22/2023
 Supervised By :mohammad ahmed 06/23/2023

Quant Time: Jun 20 13:30:01 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN053123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 20 13:20:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.005	152	5150	0.400	ng	0.00
7) Naphthalene-d8	10.825	136	12446	0.400	ng	# 0.01
13) Acenaphthene-d10	14.641	164	6588	0.400	ng	0.01
19) Phenanthrene-d10	17.393	188	11688	0.400	ng	0.01
29) Chrysene-d12	21.569	240	9256	0.400	ng	# 0.00
35) Perylene-d12	24.054	264	10112	0.400	ng	# 0.03
System Monitoring Compounds						
4) 2-Fluorophenol	5.557	112	4137	0.286	ng	0.00
5) Phenol-d6	7.196	99	3638m	0.205	ng	0.04
8) Nitrobenzene-d5	9.223	82	3337m	0.289	ng	0.05
11) 2-Methylnaphthalene-d10	12.423	152	6060	0.338	ng	0.02
14) 2,4,6-Tribromophenol	16.164	330	390	0.192	ng	0.02
15) 2-Fluorobiphenyl	13.283	172	8770m	0.315	ng	0.02
27) Fluoranthene-d10	19.416	212	10671	0.444	ng	0.00
31) Terphenyl-d14	19.997	244	7326	0.326	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062023\
Data File : BN025910.D
Acq On : 20 Jun 2023 09:58
Operator : MA/JU
Sample : PB153492BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB153492BL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/22/2023
Supervised By :mohammad ahmed 06/23/2023

Quant Time: Jun 20 13:30:01 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN053123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 20 13:20:07 2023
Response via : Initial Calibration

