




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062023\
Data File : BN025908.D
Acq On : 20 Jun 2023 07:59
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 20 08:56:52 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 20 08:56:48 2023
Response via : Initial Calibration



TIC: BN025908.D\data.ms

(70) Pentachlorophenol (C)




17.022min (0.000) 16362.82 ng

response 388059

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	63.31
264.00	61.60	61.21
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062023\
 Data File : BN025908.D
 Acq On : 20 Jun 2023 07:59
 Operator : MA/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 20 08:56:52 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 20 08:56:48 2023
 Response via : Initial Calibration



TIC: BN025908.D\data.ms

(77) Benzidine

19.610min (0.000) 0.00 ng

response 3104031

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.06
183.00	13.20	11.28
0.00	0.00	0.00