

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN062121\
 Data File : BN014952.D
 Acq On : 21 Jun 2021 12:49
 Operator : CG/JU
 Sample : SST00.801
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST00.8001

Quant Time: Jun 21 13:21:55 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN062121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 21 13:17:31 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.75	152	3115	0.40	ng/ul	0.00
4) Naphthalene-d8	10.52	136	10589	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.37	164	5410	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.11	188	10678	0.40	ng/ul	0.00
17) Chrysene-d12	21.32	240	8580	0.40	ng/ul	0.00
23) Perylene-d12	23.57	264	6496	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.30	96	2778	0.63	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.11	152	13363	0.75	ng/ul	0.00
18) Fluoranthene-d10	19.15	212	23704	0.88	ng/ul	0.00
Target Compounds				Ovalue		
2) 1,4-Dioxane	3.33	88	3053	0.743	ng/ul	91
5) Naphthalene	10.57	128	26699	0.795	ng/ul	99
7) 2-Methylnaphthalene	12.18	142	17580	0.796	ng/ul	99
8) 1-Methylnaphthalene	12.40	142	17264	0.809	ng/ul	98
10) Acenaphthylene	14.09	152	22060	0.774	ng/ul	100
11) Acenaphthene	14.43	153	16991	0.841	ng/ul	99
12) Fluorene	15.42	166	18960	0.819	ng/ul	99
14) Pentachlorophenol	16.76	266	1725	0.645	ng/ul	99
15) Phenanthrene	17.16	178	30798	0.865	ng/ul	100
16) Anthracene	17.25	178	26664	0.769	ng/ul	99
19) Fluoranthene	19.18	202	32745	0.953	ng/ul	99
20) Pyrene	19.54	202	32221	0.915	ng/ul	99
21) Benzo(a)anthracene	21.30	228	25401	0.818	ng/ul	100
22) Chrysene	21.35	228	30152	0.892	ng/ul	100
24) Benzo(b)fluoranthene	22.90	252	24578	0.965	ng/ul	99
25) Benzo(k)fluoranthene	22.95	252	27830	0.942	ng/ul	99
26) Benzo(a)pyrene	23.48	252	20428	0.876	ng/ul	99
27) Indeno(1,2,3-cd)pyrene	25.85	276	24969	0.928	ng/ul#	99
28) Dibenzo(a,h)anthracene	25.87	278	19665	0.919	ng/ul	99
29) Benzo(g,h,i)perylene	26.54	276	23245	0.942	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN062121\
Data File : BN014952.D
Acq On : 21 Jun 2021 12:49
Operator : CG/JU
Sample : SST0.801
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.8001

Quant Time: Jun 21 13:21:55 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN062121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 21 13:17:31 2021
Response via : Initial Calibration

