

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN062121\
 Data File : BN014953.D
 Acq On : 21 Jun 2021 13:25
 Operator : CG/JU
 Sample : SSTD1.602
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD1.6002

Quant Time: Jun 21 13:53:57 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN062121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 21 13:17:31 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.74	152	3271	0.40	ng/ul	0.00
4) Naphthalene-d8	10.52	136	11215	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.37	164	5667	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.11	188	10766	0.40	ng/ul	0.00
17) Chrysene-d12	21.32	240	8977	0.40	ng/ul	0.00
23) Perylene-d12	23.58	264	6502	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.30	96	5886	1.27	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.11	152	28446	1.51	ng/ul	0.00
18) Fluoranthene-d10	19.15	212	49678	1.76	ng/ul	0.00
Target Compounds				Ovalue		
2) 1,4-Dioxane	3.33	88	6430	1.489	ng/ul#	87
5) Naphthalene	10.57	128	56976	1.602	ng/ul	99
7) 2-Methylnaphthalene	12.18	142	37595	1.608	ng/ul	100
8) 1-Methylnaphthalene	12.40	142	36704	1.625	ng/ul	99
10) Acenaphthylene	14.09	152	47621	1.594	ng/ul	99
11) Acenaphthene	14.43	153	36202	1.710	ng/ul	100
12) Fluorene	15.42	166	40413	1.667	ng/ul	100
14) Pentachlorophenol	16.76	266	3739	1.386	ng/ul	99
15) Phenanthrene	17.16	178	63592	1.771	ng/ul	100
16) Anthracene	17.25	178	56939	1.630	ng/ul	99
19) Fluoranthene	19.18	202	69706	1.938	ng/ul	100
20) Pyrene	19.54	202	67510	1.831	ng/ul	100
21) Benzo(a)anthracene	21.30	228	55101	1.696	ng/ul	99
22) Chrysene	21.36	228	63496	1.795	ng/ul	99
24) Benzo(b)fluoranthene	22.90	252	51381	2.015	ng/ul	98
25) Benzo(k)fluoranthene	22.95	252	59046	1.996	ng/ul	98
26) Benzo(a)pyrene	23.48	252	43108	1.847	ng/ul	96
27) Indeno(1,2,3-cd)pyrene	25.86	276	54093	2.008	ng/ul#	99
28) Dibenzo(a,h)anthracene	25.88	278	43017	2.009	ng/ul	96
29) Benzo(g,h,i)perylene	26.54	276	48613	1.969	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN062121\
Data File : BN014953.D
Acq On : 21 Jun 2021 13:25
Operator : CG/JU
Sample : SSTD1.602
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD1.6002

Quant Time: Jun 21 13:53:57 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN062121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 21 13:17:31 2021
Response via : Initial Calibration

