


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062121\
 Data File : BN014948.D
 Acq On : 21 Jun 2021 10:25
 Operator : CG/JU
 Sample : DFTPP151
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP151

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062121.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Jun 22 11:53:52 2021



AutoFind: Scans 2463, 2464, 2465; Background Corrected with Scan 2456

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.4	80781	PASS
68	69	0.00	2	1.9	1595	PASS
69	198	0.00	100	34.6	83747	PASS
70	69	0.00	2	0.5	423	PASS
127	198	10	80	44.6	108048	PASS
197	198	0.00	2	0.4	1025	PASS
198	198	100	100	100.0	242027	PASS
199	198	5	9	6.9	16726	PASS
275	198	10	60	23.2	56061	PASS
365	198	1	100	2.8	6790	PASS
441	198	0.01	100	14.7	35549	PASS
442	198	50	100	93.6	226603	PASS
443	442	15	24	18.3	41384	PASS

m/z	Abundance
39.90	914.0
41.00	336.0
44.00	3839.0
48.90	917.0
83.80	627.0
85.90	320.0
207.00	1529.0
208.00	351.0

Instrument :
BNA_N
ClientSampleId :
DFTPP151

m/z	Abundance
37.10	417.0
38.00	1071.0
39.10	7495.0
40.00	1264.0
41.00	572.0
44.00	3487.0
44.90	327.0
49.00	1351.0
50.00	24552.0
51.00	99192.0
52.00	5612.0
55.00	590.0
56.00	2611.0
57.00	6921.0
57.90	421.0
60.90	1536.0
62.00	1538.0
63.00	3974.0
64.00	499.0
65.00	2185.0
68.00	1980.0
69.00	99520.0
70.00	519.0
73.00	1188.0
74.00	10120.0
75.00	16241.0
76.00	5703.0
77.00	118656.0
78.00	8459.0
79.00	7914.0
80.00	5354.0
81.00	8144.0
82.00	1875.0
83.00	1865.0
83.90	748.0
85.00	1528.0
86.00	2943.0
86.90	1155.0
88.00	538.0
90.90	1721.0
92.00	1920.0
93.00	12259.0
94.00	1020.0
96.10	880.0
98.00	9760.0
99.00	8809.0
100.00	705.0
101.00	4570.0
102.90	1482.0
103.90	2875.0
105.00	2622.0
105.90	821.0
107.00	35368.0
108.00	5475.0
109.10	1045.0
110.00	66056.0
111.00	10199.0
111.90	1113.0
112.90	455.0
116.00	2067.0
117.00	28816.0
118.00	2200.0
120.00	529.0
122.00	2361.0
123.00	3304.0
123.90	1452.0
125.00	1462.0
126.20	677.0
127.00	128104.0
128.00	9971.0
129.00	45704.0
129.90	3847.0
131.00	660.0
132.80	344.0
134.00	1324.0
135.00	4027.0
136.00	1520.0
137.00	2294.0
138.00	394.0
140.00	513.0
141.00	5666.0
142.00	1961.0
142.90	1301.0
144.90	392.0
145.90	864.0
147.00	3268.0
148.00	6878.0
149.00	1735.0
149.80	445.0
151.10	857.0
151.90	456.0
153.00	1923.0

Instrument :
BNA_N
ClientSampleId :
DFTPP151

153.90	1689.0
155.00	3367.0
156.00	4972.0
157.10	948.0
157.80	1030.0
158.90	894.0
159.90	2048.0
161.00	3066.0
161.90	812.0
163.00	334.0
164.00	409.0
164.90	2271.0
166.00	1647.0
167.00	12018.0
168.00	8762.0
169.00	987.0
170.00	461.0
170.90	550.0
171.90	902.0
173.00	1308.0
174.00	2258.0
175.00	5023.0
176.00	1849.0
177.00	2374.0
178.00	785.0
179.00	8251.0
180.00	6158.0
181.00	3210.0
181.90	546.0
183.80	778.0
185.00	4146.0
186.00	36416.0
187.00	10093.0
188.00	1032.0
189.00	1879.0
191.00	1257.0
191.90	3055.0
193.00	3094.0
194.00	794.0
195.00	473.0
196.00	7507.0
198.00	264128.0
199.00	18992.0
200.00	1196.0
201.40	998.0
203.00	1597.0
204.00	8795.0
205.00	14906.0
206.00	66736.0
207.00	10795.0
208.00	2562.0
208.90	865.0
210.00	825.0
210.90	2340.0
214.90	861.0
216.00	1387.0
217.00	15872.0
218.00	2093.0
220.10	353.0
221.00	10722.0
223.00	3933.0
224.00	32592.0
225.00	8936.0
226.00	1114.0
227.00	13004.0
228.00	2390.0
228.90	3093.0
230.00	454.0
231.00	1381.0
234.00	896.0
235.00	911.0
236.00	695.0
237.00	1441.0
238.90	665.0
239.90	504.0
241.00	526.0
242.00	1871.0
243.00	1757.0
244.00	27152.0
245.00	3547.0
246.00	5370.0
247.00	1051.0
249.00	1134.0
252.90	613.0
255.00	131968.0
256.00	20416.0
257.00	1553.0
258.00	7091.0
259.00	1286.0
265.00	2770.0
266.00	338.0
271.00	337.0
272.00	368.0
273.00	4114.0
274.00	10898.0
275.00	56328.0

Instrument :
BNA_N
ClientSampleId :
DFTPP151

276.00	9703.0
277.00	4816.0
277.90	813.0
283.10	550.0
284.00	451.0
285.10	860.0
293.00	1150.0
293.90	408.0
295.10	484.0
295.90	15164.0
297.00	2382.0
302.00	444.0
303.00	2119.0
304.00	497.0
314.00	649.0
315.00	1960.0
315.90	932.0
321.00	467.0
323.10	4650.0
324.00	738.0
326.90	988.0
328.00	554.0
332.00	413.0
333.00	502.0
334.00	4021.0
335.00	908.0
341.00	654.0
346.00	1020.0
352.00	1720.0
353.00	1154.0
354.00	1790.0
365.00	5649.0
366.00	930.0
371.00	506.0
372.00	2711.0
373.00	763.0
382.90	697.0
389.90	323.0
390.90	369.0
401.90	1050.0
402.90	1505.0
404.10	553.0
421.00	1933.0
422.00	1388.0
423.00	9573.0
424.00	3571.0
425.10	336.0
441.00	32000.0
442.00	194624.0
443.00	31688.0
443.90	4194.0

Instrument :
BNA_N
ClientSampleId :
DFTPP151

m/z	Abundance
36.00	326.0
36.90	403.0
38.00	1144.0
39.10	6379.0
39.90	935.0
41.00	427.0
43.00	351.0
44.00	2962.0
49.00	1377.0
50.00	22768.0
51.00	93408.0
52.00	4619.0
55.10	540.0
56.00	3141.0
57.00	6245.0
61.00	1312.0
62.00	1237.0
63.00	4044.0
64.00	630.0
65.00	1946.0
68.00	1748.0
69.00	98952.0
69.90	448.0
73.00	1093.0
74.00	9683.0
75.00	15622.0
76.00	5034.0
77.00	112384.0
78.00	8077.0
79.00	7234.0
80.00	5211.0
81.00	7718.0
81.90	1976.0
83.00	1977.0
84.00	712.0
84.90	1729.0
86.00	2762.0
87.00	1017.0
88.00	577.0
91.00	1896.0
92.00	1989.0
93.00	12759.0
94.00	882.0
95.10	318.0
96.00	809.0
98.00	9224.0
99.00	8524.0
100.00	728.0
101.00	5013.0
103.00	1722.0
104.00	2870.0
105.00	2747.0
106.00	862.0
107.00	35784.0
108.00	5574.0
109.00	1109.0
110.00	66432.0
111.00	9721.0
111.90	1209.0
112.80	339.0
116.00	1995.0
117.00	26176.0
118.00	2125.0
120.00	426.0
122.00	2205.0
123.00	3484.0
124.00	1583.0
125.00	1267.0
126.10	804.0
127.00	123416.0
128.00	9606.0
129.00	46984.0
130.00	4549.0
131.00	762.0
131.90	402.0
134.00	1261.0
135.00	3771.0
136.00	1467.0
137.00	2265.0
137.90	507.0
140.00	589.0
141.00	6272.0
142.00	2039.0
143.00	1333.0
143.90	359.0
145.00	561.0
146.00	1224.0
147.00	3245.0
148.00	7345.0
149.00	1531.0
149.90	498.0
151.10	771.0

Instrument :
BNA_N
ClientSampleId :
DFTPP151

151.80	492.0
152.90	2107.0
154.00	1591.0
155.00	3580.0
156.00	5566.0
156.90	1073.0
158.00	1279.0
159.00	814.0
160.00	1972.0
161.00	2997.0
161.90	967.0
163.00	304.0
165.00	2373.0
166.10	1645.0
167.00	11900.0
168.00	7496.0
169.00	1239.0
169.80	415.0
170.90	524.0
172.00	1031.0
173.00	1399.0
174.00	2631.0
175.00	5072.0
176.00	1611.0
177.00	2326.0
177.90	959.0
179.00	8686.0
180.00	7048.0
181.00	3106.0
181.90	448.0
182.70	303.0
183.90	658.0
185.00	4496.0
186.00	36912.0
187.00	10407.0
188.00	1002.0
188.90	2379.0
191.00	1197.0
192.00	3009.0
193.00	3301.0
193.90	763.0
195.00	420.0
196.00	7758.0
197.00	1844.0
198.00	278592.0
199.00	19736.0
200.00	1431.0
201.40	1531.0
203.00	1833.0
204.00	9300.0
205.00	16472.0
206.00	71632.0
207.00	11097.0
208.00	2648.0
209.00	975.0
210.30	1218.0
210.90	2593.0
215.00	878.0
216.00	1520.0
217.00	17616.0
218.00	2614.0
221.00	11528.0
223.00	4148.0
224.00	39904.0
225.00	10193.0
226.10	1040.0
227.00	14460.0
228.00	2236.0
229.00	2858.0
230.00	383.0
231.00	1680.0
233.90	1056.0
235.00	1152.0
236.00	905.0
237.00	1371.0
239.00	570.0
239.90	303.0
240.90	1081.0
242.00	1967.0
243.00	2239.0
244.00	29016.0
245.00	4752.0
246.00	5533.0
247.00	1151.0
249.00	996.0
252.90	571.0
255.00	164160.0
256.00	22152.0
257.00	1873.0
258.00	7939.0
258.90	1585.0
265.00	3949.0
265.90	487.0
271.90	447.0
273.00	5339.0
274.00	12380.0

Instrument :
BNA_N
ClientSampleId :
DFTPP151

275.00	65656.0
276.00	9595.0
277.00	5935.0
277.90	1097.0
281.10	316.0
283.00	723.0
284.00	573.0
285.00	1119.0
291.90	309.0
292.90	978.0
294.10	399.0
295.00	365.0
296.00	16648.0
297.00	2476.0
301.80	327.0
303.10	1995.0
304.10	628.0
308.00	310.0
314.00	1079.0
314.90	1725.0
316.00	1046.0
320.90	581.0
321.90	326.0
323.00	6333.0
324.00	1086.0
326.90	1310.0
328.00	580.0
332.00	394.0
333.00	708.0
334.00	3270.0
335.00	1027.0
341.10	851.0
346.00	1348.0
352.00	1826.0
352.90	1261.0
354.00	2077.0
355.10	483.0
364.90	7634.0
365.90	1442.0
371.00	487.0
372.00	3225.0
373.00	789.0
383.00	903.0
383.90	316.0
389.90	450.0
391.10	358.0
392.00	349.0
402.00	1531.0
403.00	2288.0
404.00	846.0
421.00	2264.0
422.00	1745.0
423.00	13726.0
424.00	3195.0
424.90	422.0
441.00	40360.0
442.10	265088.0
443.00	53312.0
444.00	4587.0

Instrument :
BNA_N
ClientSampleId :
DFTPP151

m/z	Abundance
38.00	668.0
39.00	3862.0
40.00	1101.0
41.10	373.0
44.00	3691.0
48.90	961.0
50.00	13037.0
51.00	49744.0
52.00	2460.0
55.00	499.0
55.90	1656.0
57.00	3311.0
61.00	671.0
62.00	704.0
63.00	2195.0
63.90	323.0
65.00	1319.0
68.00	1058.0
69.00	52768.0
70.00	302.0
73.00	739.0
74.00	5220.0
75.00	8564.0
76.00	2646.0
77.10	64136.0
78.00	4929.0
79.00	4240.0
80.00	3256.0
81.00	4310.0
82.00	1020.0
83.00	1409.0
83.90	636.0
85.00	1148.0
86.00	1467.0
86.90	802.0
88.00	337.0
91.00	1139.0
92.00	1087.0
93.00	6710.0
94.00	453.0
96.00	561.0
98.00	5051.0
99.00	4777.0
100.00	414.0
100.90	2789.0
103.00	849.0
104.00	1601.0
105.00	1597.0
106.00	425.0
107.00	19768.0
108.00	3038.0
109.00	590.0
110.00	39392.0
111.00	5588.0
111.90	842.0
112.90	351.0
116.00	1326.0
117.00	15718.0
118.00	1205.0
121.90	1381.0
122.90	1763.0
123.90	921.0
124.90	857.0
126.10	440.0
127.00	72624.0
128.00	5693.0
129.00	26112.0
129.90	2290.0
130.90	469.0
133.90	694.0
135.00	2282.0
135.90	898.0
137.00	1431.0
140.00	411.0
141.00	3570.0
141.90	1226.0
142.90	866.0
146.00	697.0
147.00	1893.0
147.90	3842.0
149.00	861.0
151.00	607.0
151.90	322.0
152.90	1230.0
154.00	957.0
155.00	2150.0
156.00	3013.0
157.00	566.0
157.90	711.0
159.00	630.0
159.90	1086.0
161.00	1988.0

Instrument :
BNA_N
ClientSampleId :
DFTPP151

161.90	446.0
165.00	1430.0
165.90	1180.0
167.00	7674.0
168.00	4594.0
168.90	776.0
171.90	812.0
173.00	892.0
174.00	1816.0
175.00	3265.0
176.00	929.0
176.90	1770.0
178.00	588.0
179.00	6192.0
180.00	3972.0
181.00	1996.0
183.90	400.0
185.00	3029.0
186.00	22728.0
187.00	6653.0
188.00	511.0
188.90	1568.0
190.90	820.0
192.00	1888.0
193.00	2478.0
194.00	437.0
195.00	366.0
196.00	5398.0
197.10	1231.0
198.00	183360.0
199.00	11450.0
200.00	833.0
201.30	679.0
203.00	965.0
204.00	5971.0
205.00	10612.0
206.00	43880.0
207.00	7369.0
208.00	1505.0
208.90	687.0
211.00	1797.0
211.90	321.0
214.90	427.0
216.00	848.0
216.90	11301.0
218.00	1415.0
221.00	6876.0
221.70	1561.0
222.90	2529.0
224.00	23504.0
225.00	6023.0
226.00	752.0
227.00	9435.0
228.00	1474.0
229.00	1962.0
230.10	405.0
231.00	810.0
233.90	649.0
234.90	820.0
235.90	418.0
237.00	974.0
239.00	427.0
240.00	334.0
241.00	590.0
241.90	1284.0
243.00	1265.0
244.00	20576.0
245.00	2261.0
246.00	3676.0
246.90	963.0
248.90	746.0
253.00	420.0
254.00	919.0
255.00	102664.0
256.00	14549.0
257.00	854.0
258.00	5598.0
258.90	830.0
265.00	2369.0
266.00	348.0
270.90	361.0
273.00	2910.0
274.00	8774.0
275.00	46200.0
276.00	6512.0
277.00	3593.0
278.00	713.0
283.00	517.0
285.00	660.0
292.90	961.0
294.00	318.0
295.90	11459.0
297.00	1887.0
303.00	1465.0
304.00	362.0
313.90	590.0

Instrument :
BNA_N
ClientSampleId :
DFTPP151

315.00	1528.0
316.10	839.0
321.00	681.0
323.00	4460.0
323.90	766.0
327.00	589.0
328.00	562.0
331.90	330.0
332.90	445.0
334.00	2657.0
335.00	930.0
341.00	465.0
346.00	987.0
352.00	1296.0
353.00	1146.0
354.00	1690.0
365.00	7086.0
365.90	1175.0
370.90	375.0
372.00	2601.0
373.00	793.0
383.00	745.0
384.00	313.0
389.90	387.0
402.00	1034.0
403.00	1800.0
403.90	684.0
421.00	1602.0
422.00	1483.0
423.00	10575.0
423.90	2717.0
425.00	345.0
441.00	34288.0
442.10	220096.0
443.00	39152.0
444.00	3198.0

Instrument :
BNA_N
ClientSampleId :
DFTPP151