

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN062219\
Data File : BN006519.D
Acq On : 23 Jun 2019 10:11
Operator : HP/JU
Sample : K3360-18
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
A41X3

Manual Integrations
APPROVED

mohammad
6/25/2019 8:45:43 AM

Quant Time: Jun 24 08:38:03 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN053019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 17 11:13:07 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.63	152	5100	0.40	ng/ul	-0.01
2) Naphthalene-d8	10.41	136	19594	0.40	ng/ul	-0.02
6) Acenaphthene-d10	14.28	164	10688	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.04	188	20033m	0.40	ng/ul	-0.02
16) Chrysene-d12	21.26	240	11222m	0.40	ng/ul	-0.01
20) Perylene-d12	23.48	264	9532m	0.40	ng/ul	-0.03

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Methylnaphthalene-d10	12.01	152	10941	0.36	ng/ul	-0.02
14) Fluoranthene-d10	19.08	212	19725m	0.37	ng/ul	-0.01

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN062219\
Data File : BN006519.D
Acq On : 23 Jun 2019 10:11
Operator : HP/JU
Sample : K3360-18
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
A41X3

Manual Integrations
APPROVED

mohammad
6/25/2019 8:45:43 AM

Quant Time: Jun 24 08:38:03 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN053019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 17 11:13:07 2019
Response via : Initial Calibration

