

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN062219\
 Data File : BN006495.D
 Acq On : 22 Jun 2019 18:41
 Operator : HP/JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.427

Quant Time: Jun 24 08:29:08 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN053019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 17 11:13:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.63	152	3960	0.40	ng/ul	0.00
2) Naphthalene-d8	10.41	136	16055	0.40	ng/ul	-0.02
6) Acenaphthene-d10	14.28	164	8828	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.04	188	18479	0.40	ng/ul	-0.01
16) Chrysene-d12	21.26	240	12261	0.40	ng/ul	0.00
20) Perylene-d12	23.49	264	11124	0.40	ng/ul	-0.02

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Methylnaphthalene-d10	12.01	152	10287	0.42	ng/ul	-0.01
14) Fluoranthene-d10	19.08	212	19905	0.40	ng/ul	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Naphthalene	10.46	128	18996	0.429	ng/ul	100
5) 2-Methylnaphthalene	12.08	142	12845	0.416	ng/ul	98
7) Acenaphthylene	14.00	152	19485	0.397	ng/ul	99
8) Acenaphthene	14.34	153	14583	0.428	ng/ul	99
9) Fluorene	15.34	166	15648	0.393	ng/ul	96
11) Pentachlorophenol	16.71	266	1103	0.235	ng/ul	95
12) Phenanthrene	17.08	178	24468	0.412	ng/ul	99
13) Anthracene	17.17	178	21600	0.381	ng/ul	99
15) Fluoranthene	19.11	202	26608	0.406	ng/ul	97
17) Pyrene	19.48	202	26982	0.418	ng/ul	100
18) Benzo(a)anthracene	21.24	228	20331	0.375	ng/ul	99
19) Chrysene	21.29	228	22555	0.443	ng/ul	99
21) Benzo(b)fluoranthene	22.82	252	18997	0.410	ng/ul	93
22) Benzo(k)fluoranthene	22.87	252	19990	0.442	ng/ul	96
23) Benzo(a)pyrene	23.39	252	18257	0.423	ng/ul#	90
24) Indeno(1,2,3-cd)pyrene	25.72	276	19753	0.422	ng/ul#	97
25) Dibenzo(a,h)anthracene	25.73	278	15512	0.415	ng/ul	95
26) Benzo(g,h,i)perylene	26.41	276	16650	0.425	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN062219\
Data File : BN006495.D
Acq On : 22 Jun 2019 18:41
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.427

Quant Time: Jun 24 08:29:08 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN053019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 17 11:13:07 2019
Response via : Initial Calibration

