

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062222\
 Data File : BN020344.D
 Acq On : 22 Jun 2022 15:36
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICBN062222

Quant Time: Jun 23 02:31:42 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 22 15:57:15 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	106	0.00
2	1,4-Dioxane	0.458	0.447	2.4	102	0.00
3	n-Nitrosodimethylamine	0.417	0.347	16.8	131	0.00
4 S	2-Fluorophenol	1.172	1.123	4.2	99	0.00
5 S	Phenol-d6	1.511	1.359	10.1	98	0.00
6	bis(2-Chloroethyl)ether	1.173	1.088	7.2	103	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	100	0.00
8 S	Nitrobenzene-d5	0.279	0.260	6.8	103	0.00
9	Naphthalene	1.192	1.174	1.5	102	0.00
10	Hexachlorobutadiene	0.201	0.194	3.5	102	0.00
11 SURR	2-Methylnaphthalene-d10	0.744	0.723	2.8	101	0.00
12	2-Methylnaphthalene	0.875	0.840	4.0	100	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	98	0.00
14 S	2,4,6-Tribromophenol	0.190	0.178	6.3	96	0.00
15 S	2-Fluorobiphenyl	1.431	1.340	6.4	98	0.00
16	Acenaphthylene	1.875	1.736	7.4	99	0.00
17	Acenaphthene	1.306	1.284	1.7	98	0.00
18	Fluorene	1.879	1.818	3.2	98	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	97	0.00
20	4,6-Dinitro-2-methylphenol	0.062	0.058	6.5	105	0.01
21	4-Bromophenyl-phenylether	0.221	0.203	8.1	98	0.00
22	Hexachlorobenzene	0.249	0.237	4.8	100	0.00
23	Atrazine	0.164	0.150	8.5	96	0.00
24	Pentachlorophenol	0.110	0.095	13.6	95	0.00
25	Phenanthrene	1.377	1.279	7.1	97	0.01
26	Anthracene	1.185	1.077	9.1	96	0.00
27 SURR	Fluoranthene-d10	1.403	1.364	2.8	99	0.00
28	Fluoranthene	1.819	1.717	5.6	98	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	98	0.00
30	Pyrene	1.476	1.360	7.9	97	0.00
31 S	Terphenyl-d14	0.797	0.748	6.1	98	0.00
32	Benzo(a)anthracene	1.435	1.271	11.4	97	0.00
33	Chrysene	1.635	1.611	1.5	99	0.00
34	Bis(2-ethylhexyl)phthalate	0.659	0.575	12.7	95	0.00
35 I	Perylene-d12	1.000	1.000	0.0	98	0.00
36	Indeno(1,2,3-cd)pyrene	1.964	1.885	4.0	101	0.00
37	Benzo(b)fluoranthene	1.580	1.462	7.5	98	0.00
38	Benzo(k)fluoranthene	1.627	1.511	7.1	97	0.00
39 C	Benzo(a)pyrene	1.378	1.302	5.5	98	0.00
40	Dibenzo(a,h)anthracene	1.560	1.516	2.8	102	0.00
41	Benzo(g,h,i)perylene	1.743	1.717	1.5	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
ClientSampleId :
ICVBN062222