

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062222\
 Data File : BN020344.D
 Acq On : 22 Jun 2022 15:36
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICBVN062222

Quant Time: Jun 23 02:31:42 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 22 15:57:15 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	106	0.00
2	1,4-Dioxane	0.400	0.364	9.0	102	0.00
3	n-Nitrosodimethylamine	0.400	0.332	17.0	131	0.00
4 S	2-Fluorophenol	0.400	0.383	4.3	99	0.00
5 S	Phenol-d6	0.400	0.360	10.0	98	0.00
6	bis(2-Chloroethyl)ether	0.400	0.371	7.3	103	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	100	0.00
8 S	Nitrobenzene-d5	0.400	0.373	6.8	103	0.00
9	Naphthalene	0.400	0.394	1.5	102	0.00
10	Hexachlorobutadiene	0.400	0.387	3.3	102	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.389	2.8	101	0.00
12	2-Methylnaphthalene	0.400	0.384	4.0	100	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	98	0.00
14 S	2,4,6-Tribromophenol	0.400	0.374	6.5	96	0.00
15 S	2-Fluorobiphenyl	0.400	0.375	6.3	98	0.00
16	Acenaphthylene	0.400	0.370	7.5	99	0.00
17	Acenaphthene	0.400	0.393	1.8	98	0.00
18	Fluorene	0.400	0.387	3.3	98	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	97	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.393	1.8	105	0.01
21	4-Bromophenyl-phenylether	0.400	0.368	8.0	98	0.00
22	Hexachlorobenzene	0.400	0.381	4.8	100	0.00
23	Atrazine	0.400	0.365	8.8	96	0.00
24	Pentachlorophenol	0.400	0.346	13.5	95	0.00
25	Phenanthrene	0.400	0.372	7.0	97	0.01
26	Anthracene	0.400	0.363	9.3	96	0.00
27 SURR	Fluoranthene-d10	0.400	0.389	2.8	99	0.00
28	Fluoranthene	0.400	0.378	5.5	98	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	98	0.00
30	Pyrene	0.400	0.368	8.0	97	0.00
31 S	Terphenyl-d14	0.400	0.376	6.0	98	0.00
32	Benzo(a)anthracene	0.400	0.354	11.5	97	0.00
33	Chrysene	0.400	0.394	1.5	99	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.349	12.8	95	0.00
35 I	Perylene-d12	0.400	0.400	0.0	98	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.384	4.0	101	0.00
37	Benzo(b)fluoranthene	0.400	0.370	7.5	98	0.00
38	Benzo(k)fluoranthene	0.400	0.371	7.3	97	0.00
39 C	Benzo(a)pyrene	0.400	0.378	5.5	98	0.00
40	Dibenzo(a,h)anthracene	0.400	0.388	3.0	102	0.00
41	Benzo(g,h,i)perylene	0.400	0.394	1.5	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN062222