

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062323\
 Data File : BN025965.D
 Acq On : 23 Jun 2023 13:18
 Operator : MA/JU
 Sample : 03359-02
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 OWBR-03-128-148-062023-FD

Quant Time: Jun 23 16:01:04 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN053123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 23 15:28:11 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.976	152	6554	0.400	ng	0.00
7) Naphthalene-d8	10.792	136	17832	0.400	ng	0.01
13) Acenaphthene-d10	14.608	164	9872	0.400	ng	0.00
19) Phenanthrene-d10	17.356	188	19033	0.400	ng	0.00
29) Chrysene-d12	21.542	240	13433	0.400	ng	# 0.00
35) Perylene-d12	24.013	264	14483	0.400	ng	#-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.528	112	2244	0.122	ng	0.01
5) Phenol-d6	7.138	99	1685	0.074	ng	0.00
8) Nitrobenzene-d5	9.148	82	5053	0.305	ng	0.01
11) 2-Methylnaphthalene-d10	12.373	152	7884	0.307	ng	0.00
14) 2,4,6-Tribromophenol	16.102	330	1471	0.483	ng	0.00
15) 2-Fluorobiphenyl	13.240	172	15944	0.382	ng	0.00
27) Fluoranthene-d10	19.388	212	17466	0.446	ng	0.00
31) Terphenyl-d14	19.978	244	13693	0.420	ng	0.00
Target Compounds						
9) Naphthalene	10.835	128	1655	0.031	ng	# 76
34) Bis(2-ethylhexyl)phtha...	21.444	149	4993	0.162	ng	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062323\
Data File : BN025965.D
Acq On : 23 Jun 2023 13:18
Operator : MA/JU
Sample : 03359-02
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
OWBR-03-128-148-062023-FD

Quant Time: Jun 23 16:01:04 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN053123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 23 15:28:11 2023
Response via : Initial Calibration

