

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN062419\
 Data File : BN006541.D
 Acq On : 24 Jun 2019 21:04
 Operator : HP/JU
 Sample : K3362-02
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 A41Z5

Manual Integrations
 APPROVED

mohammad
 6/25/2019 10:20:27 PM

Quant Time: Jun 25 03:27:07 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN053019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 17 11:13:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.63	152	5055	0.40	ng/ul	-0.01
2) Naphthalene-d8	10.41	136	18184	0.40	ng/ul	-0.02
6) Acenaphthene-d10	14.28	164	8820	0.40	ng/ul	-0.01
10) Phenanthrene-d10	17.04	188	15440m	0.40	ng/ul	-0.02
16) Chrysene-d12	21.27	240	10825m	0.40	ng/ul	0.00
20) Perylene-d12	23.50	264	10537m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.01	152	14173	0.51	ng/ul	-0.02
14) Fluoranthene-d10	19.08	212	24253m	0.58	ng/ul	0.00
Target Compounds					Ovalue	
12) Phenanthrene	17.08	178	2288m	0.046	ng/ul	
15) Fluoranthene	19.11	202	7272m	0.133	ng/ul	
17) Pyrene	19.48	202	6007	0.105	ng/ul#	93
18) Benzo(a)anthracene	21.25	228	2960	0.062	ng/ul#	92
19) Chrysene	21.30	228	2818	0.063	ng/ul#	91
21) Benzo(b)fluoranthene	22.84	252	4052m	0.092	ng/ul	
22) Benzo(k)fluoranthene	22.88	252	1559m	0.036	ng/ul	
23) Benzo(a)pyrene	23.40	252	2673m	0.065	ng/ul	
24) Indeno(1,2,3-cd)pyrene	25.74	276	1857m	0.042	ng/ul	
26) Benzo(g,h,i)perylene	26.42	276	1504m	0.041	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN062419\
Data File : BN006541.D
Acq On : 24 Jun 2019 21:04
Operator : HP/JU
Sample : K3362-02
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
A41Z5

Manual Integrations
APPROVED

mohammad
6/25/2019 10:20:27 PM

Quant Time: Jun 25 03:27:07 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN053019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 17 11:13:07 2019
Response via : Initial Calibration

