

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN062419\
 Data File : BN006543.D
 Acq On : 24 Jun 2019 22:13
 Operator : HP/JU
 Sample : K3362-04
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 A41T8

Manual Integrations
 APPROVED

mohammad
 6/25/2019 10:20:38 PM

Quant Time: Jun 25 03:43:55 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN053019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 17 11:13:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.63	152	4766	0.40	ng/ul	-0.01
2) Naphthalene-d8	10.41	136	18097	0.40	ng/ul	-0.02
6) Acenaphthene-d10	14.28	164	9346	0.40	ng/ul	-0.01
10) Phenanthrene-d10	17.04	188	17447m	0.40	ng/ul	-0.02
16) Chrysene-d12	21.26	240	9523m	0.40	ng/ul	-0.01
20) Perylene-d12	23.49	264	8407m	0.40	ng/ul	-0.02
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.01	152	10492	0.38	ng/ul	-0.02
14) Fluoranthene-d10	19.08	212	18086m	0.39	ng/ul	0.00
Target Compounds					Ovalue	
12) Phenanthrene	17.08	178	2025m	0.036	ng/ul	
15) Fluoranthene	19.11	202	4026m	0.065	ng/ul	
17) Pyrene	19.48	202	3135	0.062	ng/ul#	92
18) Benzo(a)anthracene	21.24	228	1609m	0.038	ng/ul	
19) Chrysene	21.30	228	2249	0.057	ng/ul	94
21) Benzo(b)fluoranthene	22.82	252	2623	0.075	ng/ul	67
22) Benzo(k)fluoranthene	22.87	252	1495m	0.044	ng/ul	
23) Benzo(a)pyrene	23.39	252	1602m	0.049	ng/ul	
24) Indeno(1,2,3-cd)pyrene	25.73	276	1668m	0.047	ng/ul	
25) Dibenzo(a,h)anthracene	25.73	278	837m	0.030	ng/ul	
26) Benzo(g,h,i)perylene	26.42	276	1367m	0.046	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN062419\
Data File : BN006543.D
Acq On : 24 Jun 2019 22:13
Operator : HP/JU
Sample : K3362-04
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
A41T8

Manual Integrations
APPROVED

mohammad
6/25/2019 10:20:38 PM

Quant Time: Jun 25 03:43:55 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN053019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 17 11:13:07 2019
Response via : Initial Calibration

