

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN062419\
 Data File : BN006555.D
 Acq On : 25 Jun 2019 05:16
 Operator : HP/JU
 Sample : K3388-02
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 A4215

Manual Integrations
 APPROVED

mohammad
 6/25/2019 10:21:20 PM

Quant Time: Jun 25 07:40:55 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN053019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 25 01:51:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.63	152	5754	0.40	ng/ul	0.00
2) Naphthalene-d8	10.41	136	20254	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.28	164	9493	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.04	188	17336m	0.40	ng/ul	0.00
16) Chrysene-d12	21.25	240	10956m	0.40	ng/ul	-0.01
20) Perylene-d12	23.48	264	12461m	0.40	ng/ul	-0.02
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.01	152	4622	0.15	ng/ul	0.00
14) Fluoranthene-d10	19.08	212	9299m	0.20	ng/ul	0.00
Target Compounds						Ovalue
3) Naphthalene	10.46	128	15976	0.286	ng/ul	99
5) 2-Methylnaphthalene	12.08	142	8023	0.206	ng/ul	99
7) Acenaphthylene	14.00	152	63444	1.202	ng/ul	99
8) Acenaphthene	14.34	153	8834	0.241	ng/ul	99
9) Fluorene	15.34	166	12199	0.285	ng/ul	96
12) Phenanthrene	17.08	178	121286m	2.179	ng/ul	
13) Anthracene	17.17	178	48105m	0.905	ng/ul	
15) Fluoranthene	19.10	202	202063m	3.288	ng/ul	
17) Pyrene	19.47	202	180497	3.128	ng/ul	99
18) Benzo(a)anthracene	21.24	228	95022m	1.961	ng/ul	
19) Chrysene	21.29	228	96698	2.126	ng/ul	94
21) Benzo(b)fluoranthene	22.81	252	221629m	4.265	ng/ul	
22) Benzo(k)fluoranthene	22.86	252	69149m	1.366	ng/ul	
23) Benzo(a)pyrene	23.38	252	176640m	3.650	ng/ul	
24) Indeno(1,2,3-cd)pyrene	25.72	276	196015	3.742	ng/ul#	88
25) Dibenzo(a,h)anthracene	25.71	278	45958m	1.098	ng/ul	
26) Benzo(a,h,i)perylene	26.40	276	211431	4.819	ng/ul	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN062419\
Data File : BN006555.D
Acq On : 25 Jun 2019 05:16
Operator : HP/JU
Sample : K3388-02
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
A4215

Manual Integrations
APPROVED

mohammad
6/25/2019 10:21:20 PM

Quant Time: Jun 25 07:40:55 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN053019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 25 01:51:13 2019
Response via : Initial Calibration

